

# SIEMENS

## ***ARCADIS Varic, Orbic (3D), Avantic VC10A***

**SP**

### **DICOM Conformance Statement**

Rev. 00 25-September-2007

© Siemens AG 2007  
All rights reserved

Siemens AG, Medical Solutions,  
Henkestr. 127, D-91050 Erlangen, Germany  
Siemensstr. 1, D-91301 Forchheim, Germany

Headquarters: Berlin and Munich  
Siemens AG, Wittelsbacher Platz 2, D-80333 Munich, Germany

Printed in the Federal Republic of Germany

*syngo* is a registered trademark of Siemens AG.

# Table of Contents

<b>1</b>	<b><i>Introduction</i></b> .....	<b>7</b>
1.1	Overview .....	7
1.2	Scope and Field.....	7
1.3	Audience .....	7
1.4	Remarks .....	7
1.5	Definitions, Terms and Abbreviations.....	8
1.6	References .....	8
1.7	Structure .....	8
<b>2</b>	<b><i>Implementation Model Verification</i></b> .....	<b>9</b>
2.1	Application Data Flow Diagram .....	9
2.2	Functional Definitions of Applications.....	9
2.3	Sequencing of Real-World Activities.....	9
<b>3</b>	<b><i>Application Entity Specification Verification</i></b> .....	<b>10</b>
3.1	Verification AE Specification .....	10
3.1.1	Association Establishment Policies .....	10
3.1.1.1	General .....	10
3.1.1.2	Number of Associations.....	10
3.1.1.3	Asynchronous Nature .....	10
3.1.1.4	Implementation Identifying Information .....	10
3.1.2	Association Initiation Policy .....	10
3.1.2.1	Associated Real-World Activity - Verification .....	10
3.1.2.1.1	Associated Real-World Activity – Request Verification “verification” .....	10
3.1.2.1.2	Proposed Presentation Contexts.....	11
3.1.2.1.3	SOP Specific Conformance – Verification SCU .....	11
3.1.3	Association Acceptance Policy .....	11
<b>4</b>	<b><i>Implementation Model Storage</i></b> .....	<b>12</b>
4.1	Application Data Flow Diagram .....	12
4.2	Functional Definitions of Application Entities.....	12
4.3	Sequencing of Real-World Activities.....	12
<b>5</b>	<b><i>Application Entity Specification Storage</i></b> .....	<b>13</b>
5.1	Storage AEs Specification.....	13
5.1.1	Association Establishment Policies .....	14
5.1.1.1	General .....	14
5.1.1.2	Number of Associations.....	14
5.1.1.3	Asynchronous Nature .....	14
5.1.1.4	Implementation Identifying Information .....	14
5.1.2	Association Initiation Policy .....	14
5.1.2.1	Associated Real-World Activity - Send.....	14
5.1.2.1.1	Associated Real-World Activity – Send Image Objects to a Network Destination	14
5.1.2.1.2	Proposed Presentation Context – Send Objects.....	15

5.1.2.1.3	SOP specific Conformance to Storage SOP classes .....	17
5.1.3	Association Acceptance Policy .....	19
5.1.3.1	Associated Real-World Activity - Receive .....	19
5.1.3.1.1	Associated Real-World Activity – Receiving Images from a Remote Node .....	19
5.1.3.1.2	Accepted Presentation Context – Receiving Images .....	19
5.1.3.1.3	SOP-specific Conformance Statement – Receiving Images .....	21
5.1.3.1.4	Presentation Context Acceptance Criterion .....	24
5.1.3.1.5	Transfer Syntax Selection Policies .....	24
<b>6</b>	<b><i>Implementation Model Storage Commitment.....</i></b>	<b>26</b>
6.1	Application Data Flow Diagram .....	26
6.2	Functional Definitions of Application Entities.....	26
6.3	Sequencing of real World Activities .....	27
<b>7</b>	<b><i>AE Specification Storage Commitment .....</i></b>	<b>28</b>
7.1	Storage Commitment AE Specification.....	28
7.1.1	Association Establishment Policies .....	28
7.1.1.1	General.....	28
7.1.1.2	Number of Associations.....	28
7.1.1.3	Asynchronous Nature .....	28
7.1.1.4	Implementation Identifying Information .....	28
7.1.2	Association Initiation Policy .....	29
7.1.2.1	Real World Activity – Storage Commitment .....	29
7.1.2.1.1	Associated Real-World Activity - Job Completed.....	29
7.1.2.1.2	Proposed Presentation Contexts - Job Completed .....	29
7.1.2.1.3	SOP Specific Conformance Statement- Job Completed.....	29
7.1.2.1.4	Associated Real-World Activity - Send Commit Response .....	30
7.1.2.1.5	Proposed Presentation Contexts - Send Commitment Response .....	30
7.1.2.1.6	SOP Specific Conformance Statement - Send Commitment Response.....	30
7.1.3	Association Acceptance Policy .....	30
7.1.3.1	Associated Real-World Activity - Commit SCP .....	31
7.1.3.1.1	Associated Real-World Activity - Receive Commit Request .....	31
7.1.3.1.2	Accepted Presentation Contexts - Receive Commit Request.....	31
7.1.3.1.3	SOP-specific Conformance Statement - Receive Commit Request .....	31
7.1.3.2	Associated Real-World Activity - Commit SCU .....	31
7.1.3.2.1	Associated Real-World Activity - Update Flags.....	31
7.1.3.2.2	Accepted Presentation Contexts - Update Flags .....	32
7.1.3.2.3	SOP-specific Conformance Statement - Update Flags .....	32
<b>8</b>	<b><i>Implementation Model Query / Retrieve .....</i></b>	<b>33</b>
8.1	Application Data Flow Diagram .....	33
8.2	Functional Definitions of Application Entities.....	33
8.3	Sequencing of Real-World Activities.....	34
<b>9</b>	<b><i>Application Entity Specification Query/Retrieve.....</i></b>	<b>35</b>
9.1	Query/Retrieve Service AEs Specification.....	35
9.1.1	Association Establishment Policies .....	35
9.1.1.1	General.....	35
9.1.1.2	Number of Associations.....	35
9.1.1.3	Asynchronous Nature .....	36
9.1.1.4	Implementation Identifying Information .....	36
9.1.2	Association Initiation Policy .....	36
9.1.2.1	Real World Activity - Find SCU.....	36
9.1.2.1.1	Associated Real-World Activity - Find SCU “Search” .....	36
9.1.2.1.2	Proposed Presentation Contexts - Find SCU.....	36

9.1.2.1.3	Conformance Statement - Find SCU.....	37
9.1.2.2	Real-World Activity – Move SCU.....	40
9.1.2.2.1	Associated Real-World Activity – Move SCU “Import” .....	40
9.1.2.2.2	Proposed Presentation Contexts - Move SCU “Import” .....	41
9.1.2.2.3	SOP Specific Conformance Statement - Move SCU “Import” .....	41
9.1.3	Association Acceptance Policy.....	42
9.1.3.1	Real-World Activity - Find SCP.....	42
9.1.3.1.1	Associated Real-World Activity - Find SCP.....	42
9.1.3.1.2	Accepted Presentation Contexts - Find SCP .....	42
9.1.3.1.3	SOP Specific Conformance Statement - Find SCP.....	42
9.1.3.2	Real-World Activity - Move SCP.....	47
9.1.3.2.1	Associated Real-World Activity - Move SCP .....	47
9.1.3.2.2	Accepted Presentation Contexts - Move SCP.....	47
9.1.3.2.3	SOP Specific Conformance Statement - Move SCP.....	47
<b>10</b>	<b><i>Implementation Model Print.....</i></b>	<b>49</b>
<b>10.1</b>	<b>Application Data Flow Diagram .....</b>	<b>49</b>
<b>10.2</b>	<b>Functional Definition of Application Entities .....</b>	<b>49</b>
<b>10.3</b>	<b>Sequencing of Real-World Activities .....</b>	<b>49</b>
<b>11</b>	<b><i>Application Entity Specification Print.....</i></b>	<b>50</b>
<b>11.1</b>	<b>Print Management AE Specification .....</b>	<b>50</b>
11.1.1	Association Establishment Policies .....	50
11.1.1.1	General.....	50
11.1.1.2	Number of Associations.....	51
11.1.1.3	Asynchronous Nature.....	51
11.1.1.4	Implementation Identifying Information .....	51
11.1.2	Association Initiation Policy .....	51
11.1.2.1	Associated Real-World Activity.....	51
11.1.2.1.1	Associated Real-World Activity – Printing a Printer Job Queue Entry .....	51
11.1.2.1.2	Proposed Presentation Context (Presentation Context Table) .....	52
11.1.2.1.3	SOP specific Conformance Statement – Meta SOP Classes .....	52
11.1.3	Association Acceptance Policy.....	58
<b>12</b>	<b><i>Implementation Model Worklist.....</i></b>	<b>59</b>
<b>12.1</b>	<b>Application Data Flow Diagram .....</b>	<b>59</b>
<b>12.2</b>	<b>Functional Definitions of Application Entities .....</b>	<b>59</b>
<b>12.3</b>	<b>Sequencing of Real-World Activities .....</b>	<b>60</b>
<b>13</b>	<b><i>Application Entity Specification Worklist.....</i></b>	<b>61</b>
<b>13.1</b>	<b>Modality Worklist Service AE Specification .....</b>	<b>61</b>
13.1.1	Association Establishment Policies .....	61
13.1.1.1	General.....	61
13.1.1.2	Number of Associations.....	61
13.1.1.3	Asynchronous Nature.....	61
13.1.1.4	Implementation Identifying Information .....	61
13.1.2	Association Initiation Policy .....	62
13.1.2.1	Real-World Activity .....	62
13.1.2.1.1	Associated Real-World Activity - Query (Update) Worklist .....	62
13.1.2.1.2	Proposed Presentation Contexts.....	62
13.1.2.1.3	SOP Specific Conformance Statement .....	63
13.1.2.1.4	Associated Real-World Activity – Get Worklist.....	66
13.1.2.1.5	Proposed Presentation Contexts – Get Worklist.....	67
13.1.2.1.6	SOP Specific Conformance – Get Worklist.....	67
<b>14</b>	<b><i>Implementation Model MPPS.....</i></b>	<b>68</b>

14.1	Application Data Flow Diagram .....	68
14.2	Functional Definitions of Application Entities .....	68
<b>15</b>	<b><i>AE Specification MPPS</i></b> .....	<b>69</b>
15.1	Modality Performed Procedure Step AE Specification.....	69
15.1.1.1	General.....	69
15.1.1.2	Number of Associations.....	69
15.1.1.3	Asynchronous Nature.....	69
15.1.1.4	Implementation Identifying Information .....	69
15.1.2	Association Initiation Policy .....	69
15.1.2.1	Real World Activity .....	70
15.1.2.1.1	Associated Real-World Activity - Patient registered.....	70
15.1.2.1.2	Proposed Presentation Contexts - Patient registered.....	70
15.1.2.1.3	SOP Specific Conformance Statement- Patient registered.....	70
15.1.2.1.4	Associated Real-World Activity – MPPS UI-Update.....	72
15.1.2.1.5	Proposed Presentation Contexts – MPPS UI-Update.....	72
15.1.2.1.6	SOP Specific Conformance Statement – MPPS UI-Update .....	72
<b>16</b>	<b><i>Communication Profiles</i></b> .....	<b>74</b>
16.1	Supported Communication Stacks .....	74
16.1.1	TCP/IP Stack .....	74
16.1.1.1	API.....	74
16.1.1.2	Physical Media Support.....	74
<b>17</b>	<b><i>Extensions / Specializations / Privatizations</i></b> .....	<b>75</b>
17.1.1	Standard Extended / Specialized / Private SOPs.....	75
17.1.2	Private Transfer Syntaxes .....	75
<b>18</b>	<b><i>Configuration</i></b> .....	<b>75</b>
18.1	AE Title/Presentation Address Mapping.....	75
18.1.1	DICOM Verification .....	75
18.1.2	DICOM AE Title .....	75
18.2	Configurable Parameters .....	75
18.2.1	Storage, Storage Commitment and Query/Retrieve.....	75
18.2.2	Print.....	76
18.2.3	Modality Worklist.....	76
18.3	Default Parameters .....	77
<b>19</b>	<b><i>Support of Extended Character Sets</i></b> .....	<b>77</b>
<b><i>Media Storage Conformance Statement</i></b> .....		<b>78</b>
<b>1</b>	<b><i>Introduction</i></b> .....	<b>79</b>
1.1	Purpose.....	79
1.2	Scope .....	79
1.3	Definitions, Abbreviations.....	79
1.3.1	Definitions .....	79
1.3.2	Abbreviations .....	79
1.4	References .....	79
1.5	Remarks .....	80
<b>2</b>	<b><i>Implementation Model</i></b> .....	<b>81</b>
2.1	Application Data Flow Diagram .....	81

**2.2 Functional Definitions of AEs .....81**

**2.3 Sequencing of Real-World Activities.....81**

**2.4 File Meta Information Options.....82**

**3 AE Specifications .....82**

**3.1 DICOM Archive Specification.....82**

3.1.1 File Meta Information for the Application Entity ..... 82

3.1.2 Real-World Activities for this Application Entity ..... 82

3.1.2.1 Real-World Activity: Browse Directory Information..... 82

3.1.2.1.1 Application Profiles for the RWA: Browse Directory Information..... 83

3.1.2.2 Real-World Activity: Import into local Storage ..... 83

3.1.2.2.1 Application Profiles for the RWA: Import into local Storage ..... 83

3.1.2.3 Real-World Activity: Export to local Archive Media ..... 84

3.1.2.3.1 Application Profiles for the RWA: Export to local Archive Media ..... 84

**4 Augmented and Private Profiles.....85**

**4.1 Augmented Application Profiles.....85**

**4.2 AUG-STD-GEN-CD and STD-XA1K-CD.....85**

4.2.1 Transfer Syntax Augmentation ..... 85

4.2.2 Directory Augmentation ..... 85

4.2.2.1 Basic Directory IOD Specialization..... 85

4.2.2.2 Additional Keys..... 86

4.2.2.3 Icon Images ..... 87

**5 Configuration .....88**

**5.1 AE Title Mapping .....88**

5.1.1 DICOM Media Storage AE Title..... 88

**6 Support of Extended Character Sets .....88**

# 1 Introduction

## 1.1 Overview

The Conformance Statement describes the DICOM interface for the Siemens ARCADIS family based on software version VC10A in terms of part 2 of [DICOM].

This introduction describes the application's implemented DICOM functionality in general terms.

## 1.2 Scope and Field

The ARCADIS family is a "syngo<sup>®</sup>-speaking<sup>™</sup>" mobile C-arm modality for a broad bandwidth of applications.

This DICOM Conformance Statement refers to SIEMENS ARCADIS family using the software version VC10A. Following products are affected: ARCADIS Varic, Orbic, Orbic 3D and Avantic.

The ARCADIS family is designed to be integrated into an environment of medical DICOM-based devices. The ARCADIS family DICOM network implementation acts as SCU and SCP for the DICOM Storage, Storage Commitment and Query/Retrieve services and as SCU for the DICOM Print, DICOM Basic Worklist and Modality Performed Procedure Step Services. Verification is supported in SCU (only via Service environment) and SCP role. Furthermore the handling of CD offline media is supported as a FSC, FSU and FSR.

## 1.3 Audience

This document is intended for hospital staff, health system integrators, software designers or implementers. It is assumed that the reader has a working understanding of DICOM.

## 1.4 Remarks

DICOM, by itself, does not guarantee interoperability. However, the Conformance Statement facilitates a first-level validation for interoperability between different applications supporting the same DICOM functionality as SCU and SCP, respectively.

This Conformance Statement is not intended to replace validation with other DICOM equipment to ensure proper exchange of information intended.

The scope of this Conformance Statement is to facilitate communication with Siemens and other vendors' Medical equipment. The Conformance Statement should be read and understood in conjunction with the DICOM 3.0 Standard [DICOM]. However, by itself it is not guaranteed to ensure the desired interoperability and a successful interconnectivity.

The user should be aware of the following important issues:

- The comparison of different conformance statements is the first step towards assessing interconnectivity between Siemens and non-Siemens equipment.
- Test procedures should be defined and the user should perform tests to validate the desired connectivity. DICOM itself and the conformance parts do not specify this.
- The standard will evolve to meet the users' future requirements. Siemens is actively involved in developing the standard further and therefore reserves the right to make changes to its products or to discontinue its delivery.
- Siemens reserves the right to modify the design and specifications contained herein without prior notice. Please contact your local Siemens representative for the most recent product information.

---

<sup>a</sup> *syngo* is a registered trademark of Siemens AG

## 1.5 Definitions, Terms and Abbreviations

Definitions, terms and abbreviations used in this document are defined within the different parts of the DICOM standard.

Additional Abbreviations and terms are as follows:

ACR	American College of Radiology
AE	DICOM Application Entity
ASCII	American Standard Code for Information Interchange
CSE	Customer Service Engineer
DB	Database
DCS	DICOM Conformance Statement
DIMSE	DICOM Message Service Element
DSA	Digital Subtraction Angiography
IIDC	Image-Intensifier Distortion Correction
IOD	DICOM Information Object Definition
ISO	International Standard Organization
NEMA	National Electrical Manufacturers Association
PDU	DICOM Protocol Data Unit
R	Required Key Attribute
RIS	Radiology Information System
RWA	Real-World Activity
SCU	DICOM Service Class User (DICOM client)
SCP	DICOM Service Class Provider (DICOM server)
SOP	DICOM Service-Object Pair
U	Unique Key Attribute

## 1.6 References

[DICOM] Digital Imaging and Communications in Medicine (DICOM), NEMA PS 3.1-3.18, 2007

## 1.7 Structure

This Conformance Statement is subdivided into multiple Parts, which relate to individual documents needed to declare Conformance according to the requirements of "Part 2 - Conformance" of the DICOM Standard.

Those parts are:

- - "Network Conformance Statement" for Network related Services
  - • Storage - User/Provider (includes Verification - User/Provider)
  - • Storage Commitment - User/Provider
  - • Query/Retrieve - User/Provider
  - • Basic Grayscale/Color Print - User
  - • Basic Worklist - User
  - • Modality Performed Procedure Step - User
- - "Offline Media Conformance Statement" to support local archive media.
- - A general Appendix.



## 2 Implementation Model Verification

The ARCADIS DICOM Service Tool application requests Verification to verify the ability of a foreign DICOM application on a remote node to respond to DICOM messages.

Responding to Verification requests from remote nodes is handled by the Storage SCP application.

### 2.1 Application Data Flow Diagram

The ARCADIS DICOM network implementation acts as SCU for the verification network service. The product target Operating System is Windows XP.




Figure 1: Application Data Flow Diagram - Verification SCU

### 2.2 Functional Definitions of Applications

The ARCADIS DICOM Service Tool application opens an association when a "verification" of a remote application is requested during a configuration session. This can be done when entering new data for remote application configuration or to verify existing configuration data.

### 2.3 Sequencing of Real-World Activities

Newly entered data have to be saved first, before a "verification" of these data is possible.

### **3 Application Entity Specification Verification**

#### **3.1 Verification AE Specification**

##### **3.1.1 Association Establishment Policies**

###### **3.1.1.1 General**

The ARCADIS DICOM Service Tool application attempts to open an association for verification request whenever the “verification” function is activated during network configuration of a remote DICOM application.

###### **3.1.1.2 Number of Associations**

The ARCADIS DICOM Service Tool application initiates one association at a time to request verification.

###### **3.1.1.3 Asynchronous Nature**

The ARCADIS DICOM software does not support asynchronous communication (multiple outstanding transactions over a single association).

###### **3.1.1.4 Implementation Identifying Information**

Implementation Class UID	1.3.12.2.1107.5.12
Implementation Version Name	SIEMENS_ASPVC10A

#### **3.1.2 Association Initiation Policy**

The ARCADIS DICOM Service Tool application attempts to initiate a new association for

- DIMSE C-ECHO service operations.

##### **3.1.2.1 Associated Real-World Activity - Verification**

###### **3.1.2.1.1 Associated Real-World Activity – Request Verification “verification”**

The associated Real-World activity is a C-ECHO request initiated by Service and Configuration SW environment whenever “verification” is requested. If an association to a remote Application Entity is successfully established, Verification with the configured AET is requested via the open association. If the C-ECHO Response from the remote Application contains a status other than “Success” this will be indicated in the service environment and the association is closed.

**3.1.2.1.2 Proposed Presentation Contexts**

The ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table – Verification SCU					
Abstract Syntax		Transfer Syntax		Role	Extended Negotiation
Name	UID	Name List	UID List		
Verification	1.2.840.10008.1.1	Implicit VR Little Endian Explicit VR Big Endian Explicit VR Little Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.2 1.2.840.10008.1.2.1	SCU	None

**3.1.2.1.3 SOP Specific Conformance – Verification SCU**

The Application conforms to the definitions of the Verification SCU in accordance to the DICOM Standard.

**3.1.3 Association Acceptance Policy**

The Verification SCP is part of the Storage SCP – see section 5.1.3.

## 4 Implementation Model Storage

The ARCADIS DICOM Application Entity both originates associations for Storage of DICOM Composite Information Objects in Remote Application Entities and accepts association requests for Storage from Remote Application Entities.

### 4.1 Application Data Flow Diagram

The ARCADIS DICOM network implementation acts as SCU and SCP for the storage network service and as SCP for the verification network service. The product target Operating System is Windows XP.




Figure 2: Application Data Flow Diagram – Storage SCU/SCP

### 4.2 Functional Definitions of Application Entities

The Storage SCU is invoked by the job control interface that is responsible for processing network archival tasks. The job consists of data describing the composite image objects selected for storage and the destination. An association is negotiated with the destination application entity and the image data is transferred using the C-STORE DIMSE-Service. Status of the transfer is reported to the job control interface.

The Storage SCP component of the ARCADIS DICOM application is operating as background server process. It exists when the machine is powered on and waits for Storage association requests. Upon accepting an association with a negotiated Presentation Context it starts to receive the Composite Image Objects and imports them to local database. Verification requests will be processed and responded by Storage SCP component too.

### 4.3 Sequencing of Real-World Activities

Not applicable

## 5 Application Entity Specification Storage

### 5.1 Storage AEs Specification

The ARCADIS DICOM Storage service class user/service class provider applications use one AE when initiating/receiving associations to/from remote DICOM nodes.

SIEMENS ARCADIS DICOM products are creating X-Ray Angiographic Image objects, CT Images in case of a 3D scan as well as Structured Reports and therefore provide Standard Conformance to the following DICOM V3.0 SOP Class as an SCU and SCP:

SOP Class Name	SOP Class UID
<b>X-Ray Angiographic Image Storage</b>	1.2.840.10008.5.1.4.1.1.12.1
<b>Computed Tomography Image Storage</b>	1.2.840.10008.5.1.4.1.1.2
<b>Enhanced SR Storage</b>	1.2.840.10008.5.1.4.1.1.88.22

Due to *syngo* DICOM network implementation ARCADIS DICOM products also provide at least Standard Conformance to the following DICOM V3.0 SOP Classes as an SCU and SCP:

SOP Class Name	SOP Class UID
<b>Digital X-Ray Image Storage – for Processing</b>	1.2.840.10008.5.1.4.1.1.1.1.1
<b>Digital X-Ray Image Storage – for Presentation</b>	1.2.840.10008.5.1.4.1.1.1.1
<b>Magnetic Resonance Image Storage</b>	1.2.840.10008.5.1.4.1.1.4
<b>Nuclear Medicine Image Storage</b>	1.2.840.10008.5.1.4.1.1.20
<b>Secondary Capture Image Storage</b>	1.2.840.10008.5.1.4.1.1.7
<b>UltraSound Multi-Frame Image Storage</b>	1.2.840.10008.5.1.4.1.1.3.1
<b>UltraSound Image Storage</b>	1.2.840.10008.5.1.4.1.1.6.1
<b>X-Ray RadioFluoroscopic Image Storage</b>	1.2.840.10008.5.1.4.1.1.12.2

**5.1.1 Association Establishment Policies**

**5.1.1.1 General**

The existence of a job queue entry with sending requests to network destinations or an internal trigger from processing a retrieve request will activate the DICOM Storage Application. An association request is sent to the destination AE and upon successful negotiation of a Presentation Context the transfer is started.

The default PDU size used will be 28 KB.

**5.1.1.2 Number of Associations**

The ARCADIS DICOM application initiates several associations at a time, one for each destination to which a transfer request is being processed in the active job queue list.

The ARCADIS DICOM application is able to accept multiple associations at a time. It can handle up to 10 associations in parallel.

The number of Simultaneous DICOM associations can be configured via the Service-UI. The dialog can be found in Configuration / DICOM / General.

**5.1.1.3 Asynchronous Nature**

The ARCADIS DICOM software does not support asynchronous communication (multiple outstanding transactions over a single association).

**5.1.1.4 Implementation Identifying Information**

Implementation Class UID	1.3.12.2.1107.5.12
Implementation Version Name	SIEMENS_ ASPVC10A

**5.1.2 Association Initiation Policy**

If a job with network destination gets active in the job list or a retrieve sub-operation is processed the ARCADIS DICOM application attempts to initiate a new association for

- DIMSE C-STORE Service operations.

**5.1.2.1 Associated Real-World Activity - Send**

**5.1.2.1.1 Associated Real-World Activity – Send Image Objects to a Network Destination**

The associated Real-World activity is a C-STORE request initiated by an internal daemon process triggered by a job with network destination or the processing of an external C-MOVE

retrieve request. If the process successfully establishes an association to a remote Application Entity, it will transfer each image one after another via the open association. If the C-STORE Response from the remote Application contains a status other than "Success" or "Warning", the association is aborted.

**5.1.2.1.2 Proposed Presentation Context – Send Objects**

The ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
<b>X-Ray Angiographic Image</b>	1.2.840.10008.5.1.4.1.1.12.1	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
<b>Computed Tomography Image</b>	1.2.840.10008.5.1.4.1.1.2	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
<b>Digital X-Ray Image for processing</b>	1.2.840.10008.5.1.4.1.1.1.1.1	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
<b>Digital X-Ray Image for presentation</b>	1.2.840.10008.5.1.4.1.1.1.1	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
<b>Magnetic Resonance Image</b>	1.2.840.10008.5.1.4.1.1.4	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Nuclear Medicine Image	1.2.840.10008.5.1.4.1.1.20	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
Secondary Capture Image	1.2.840.10008.5.1.4.1.1.7	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
Ultra-Sound Multi-Frame Image	1.2.840.10008.5.1.4.1.1.3.1	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
Ultra-Sound Image	1.2.840.10008.5.1.4.1.1.6.1	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
X-Ray Radio Fluoroscopic Image	1.2.840.10008.5.1.4.1.1.12.2	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCU	None
Enhanced SR	1.2.840.10008.5.1.4.1.1.88.22	Implicit VR Little Endian Explicit VR Big Endian Explicit VR Little Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.2 1.2.840.10008.1.2.1	SCU	None

\*1: The Transfer Syntax used is strongly influenced by the fact of "how was the accepted Transfer Syntax at the time when the Instance was received". E.g. the Instances received with JPEG Lossy Transfer Syntaxes will not be converted and can only be sent out with the same Transfer Syntax.

**Note:** The proposed Transfer Syntax is highly restricted for images stored internally in lossy compressed format. E.g. instances received with JPEG Loss Transfer Syntaxes will not be converted and can only be sent out with the same Transfer Syntax.

The "MOVE destinations" must be configured as Storage destinations. This would include the configuration of Transfer Syntax capabilities.

Not all the listed transfer syntaxes will be proposed all the time. For some abstract syntax only a list of uncompressed (UC) transfer syntaxes (one or more) will be proposed, for other abstract



syntaxes also JPEG Lossless (LL) syntax will be proposed and/or a list of JPEG Lossy (LY) transfer syntaxes. The contents of this lists is configurable, e.g. UC could be configured to contain only Implicit Little Endian for instance.

Depending on the real world activity initiating the C-STORE, we have the following behaviors:

- If a user initiates the C-STORE, a configuration parameter called Quality Factor (Q) will be used to decide which transfer syntax lists will be proposed. Q can take values between 0 and 100. If Q=0, only UC will be proposed. If Q = 100, UC and LL will be proposed. Else UC and LY will be proposed.
- If the C-MOVE SCP initiates the C-STORE, there is another configuration parameter called Compression Types Supported (CTS), which will be used to decide what transfer syntaxes are proposed. CTS can take integer values. If CTS=0 or CTS > 3, UC will be proposed. If CTS=1, UC and LY will be proposed. If CTS = 2, UC and LL will be proposed. If CTS >= 3, UC, LL and LY will be proposed.

The compression types JPEG lossy and JPEG lossless are parameters, which are part of the Application Entity Properties configuration (storage checked). It can be reached via the Service-UI: Configuration / DICOM / Network nodes.

### 5.1.2.1.3 SOP specific Conformance to Storage SOP classes

The ARCADIS products will create XA IOD type images from image acquisition and postprocessing applications. The ARCADIS Orbic 3D additionally will create CT IOD type images in case of a 3D scan. The XA / CT IODs will be a Standard Extended XA/CT Storage SOP class. Dose and ESWL Structured Reports will be created. For association and DIMSE level time-outs, please refer to Configuration section of this document.

#### 5.1.2.1.3.1 Optional Attributes

- Data Dictionary of DICOM Type 2 and 3 IOD Attributes

Please see the related Image Object definition tables in the Annex for a list of all DICOM IOD attributes of type 2 and 3, which are encoded by the ARCADIS applications.

#### 5.1.2.1.3.2 Subtraction

Since several workstations are not able to display XA images with mask information subtracted, ARCADIS offers the possibility to store images already subtracted. All frames can be stored subtracted, which results in a multiframe XA image object or single frames can be selected and stored in single image objects.

Please consider, that these single objects still contain information valid for the whole scene (e.g. exposure time).

#### 5.1.2.1.3.3 Structured Reports

During each examination a DICOM SR containing summarized dose information will be created. The content is specified in a private template. TID 10001 "Projection X-Ray Radiation Dose Report" is not used, since this template was not yet released during software implementation.

If a lithotripter is connected to the imaging system, the imaging system provides the possibility to create a lithotripsy report (LithoReport) with following content:

- patient data (name, first name, date of birth, gender)
- a checklist for lithotripsy treatment
- shockwave and therapy data from the lithotripter
- summary of the therapy (anesthesia / sedation applied (dose), auxiliary procedures, further comments, localization modality)

The content of such a report is also specified in a template.

Both templates are private definitions and described in Annex 5. We do not really assign template ids Tnx001 (Litho Report) and Tnx002 (Dose Report) are placeholder only for reading the templates.

Since several application are not supporting DICOM Enhanced Structured Reports, these objects are additionally stored as Secondary Capture objects.

#### 5.1.2.1.3.4 Specialized Information Object Definitions

Since the DICOM attribute X-Ray Tube Current (0018, 1151) is of type IS, it might happen that the value is 0 although it is for example 0.3. ARCADIS will additionally enter the tag (0018, 8151) with the exact value.

The DICOM images created by ARCADIS DICOM application conform to the DICOM IOD definitions (Standard extended IODs). But they will contain additional private elements, which have to be discarded by a DICOM system when modifying the image.

The DICOM nodes are responsible for data consistency when modifying images. All unknown private attributes have to be removed upon modification!

- Data Dictionary of applied private IOD Attributes

Please see "A.3 Siemens Standard Extended Modules" in the Annex for a list of possible private IOD attributes.

### 5.1.3 Association Acceptance Policy

The ARCADIS DICOM application attempts to accept a new association for

- DIMSE C-ECHO
- DIMSE C-STORE

service operations. Any Information Objects transmitted on that association will be checked on conformance and stored in database if check was successful.

#### 5.1.3.1 Associated Real-World Activity - Receive

##### 5.1.3.1.1 Associated Real-World Activity – Receiving Images from a Remote Node

The daemon receiving process will accept an association and will receive any images transmitted on that association and will store the images on disk in the own database if the conformance check is performed successfully.

##### 5.1.3.1.2 Accepted Presentation Context – Receiving Images

The ARCADIS DICOM application will accept Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
X-Ray Angiographic Image	1.2.840.10008.5.1.4.1.1.12.1	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Computed Tomography Image	1.2.840.10008.5.1.4.1.1.2	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Digital X-Ray Image for processing	1.2.840.10008.5.1.4.1.1.1.1.1	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Digital X-Ray Image for presentation	1.2.840.10008.5.1.4.1.1.1.1	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Magnetic Resonance Image	1.2.840.10008.5.1.4.1.1.4	JPEG Lossy Extended *1 (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) *1 Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Nuclear Medicine Image	1.2.840.10008.5.1.4.1.1.20	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Secondary Capture Image	1.2.840.10008.5.1.4.1.1.7	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Ultra-Sound Multi-Frame Image (retired) *1	1.2.840.10008.5.1.4.1.1.3	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Ultra-Sound Image (retired) *1	1.2.840.10008.5.1.4.1.1.6	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Ultra-Sound Multi-Frame Image	1.2.840.10008.5.1.4.1.1.3.1	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Ultra-Sound Image	1.2.840.10008.5.1.4.1.1.6.1	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
X-Ray Radio Fluoroscopic Image	1.2.840.10008.5.1.4.1.1.12.2	JPEG Lossy Extended (Process 2 & 4) JPEG Lossless, Process 14 (selection value 1) JPEG Lossy Baseline (Process 1) RLE Lossless Explicit VR Little Endian Explicit VR Big Endian Implicit VR Little Endian	1.2.840.10008.1.2.4.51 1.2.840.10008.1.2.4.70 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.5 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2 1.2.840.10008.1.2	SCP	None
Enhanced SR	1.2.840.10008.5.1.4.1.1.88.22	Implicit VR Little Endian Explicit VR Big Endian Explicit VR Little Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.2 1.2.840.10008.1.2.1	SCP	None
Verification	1.2.840.10008.1.1	Implicit VR Little Endian Explicit VR Big Endian Explicit VR Little Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.2 1.2.840.10008.1.2.1	SCP	None

\*1: US Retired and US Multi-frame Retired images are converted to US Images/US Multi-frame images before storing them into the local database. The conversion creates new images, which implies new UIDs.

**Note:**

With RLE Lossless Transfer Syntax the DICOM application will decompress the image before storing it into the database.

**5.1.3.1.3 SOP-specific Conformance Statement – Receiving Images**

The ARCADIS DICOM application conforms to the Full Storage Class at Level 2. In the event of a successful C-Store operation, the image has successfully been written on disk. For private attributes with VR=SQ only a nesting level of one is supported. This means that private attributes containing another sequence will be removed from the image header during storage.

Upon successful receiving a C-STORE-RQ, the Siemens ARCADIS DICOM receiver performs a quick plausibility test on the received image and available system resources. If this test succeeds, it returns the status SUCCESS, otherwise one of the following status codes is returned and the association is aborted:

- **Refused (A700):**  
This error status indicates a lack of Resources (e.g. not enough disk space) on the modality.
- **Invalid Dataset (0xA900):**  
The dataset is not containing one of the Attributes “Study Instance UID”, “Series Instance UID” or “SOP Instance UID”, or one of them has an invalid value.
- **Processing Error (0110):**  
An error occurred while processing the image, which makes it impossible to proceed

**Attention!** Only after sending the response, the image will be saved into the database. If during this operation an error occurs, the association will be aborted. This implies that a C-STORE-RSP with status SUCCESS does not mean that the image was successfully stored into the database.

In order to confirm that the sent images were successfully stored in the database, the sending application should use Storage Commitment Service.

If an image instance is received that is identified by a SOP Instance UID, which is already used by an Instance stored in database, then the actual received image will be discarded. The existing Instance is not superseded.

The following sections will differentiate the attribute contents required for Image Viewing. The ARCADIS DICOM application supports more formats for Storage of Images than Viewing.

#### 5.1.3.1.3.1 Image Pixel Attribute Acceptance Criterion for Grayscale Images - Viewing

The ARCADIS Multi-Modality Viewing application accepts the MONOCHROME1 and MONOCHROME2 photometric interpretation pixel format and graphic overlay with unsigned integer and 8 or 16 bits allocated. Accepted values for display:

- Pixel plane
  - samples per pixel (attribute 0028, 0002) = 1
  - photometric interpretation (attribute 0028,0004) = "MONOCHROME1"
  - photometric interpretation (attribute 0028,0004) = "MONOCHROME2"
  - pixel representation (attribute 0028, 0103) = 0
  - bits allocated (attribute 0028, 0100) = 8, 16
  - bits stored (attribute 0028,0101) = 8, 10, 12, 14, 16
  - high bit (attribute 0028,0102) = 7, 9, 11, 13, 15
  - only aspect ratio 1:1 is supported
- Overlay plane
  - overlay type (attribute 60xx, 0040) = "G"
  - bits allocated (attribute 60xx, 0100) = 16
  - bit position (attribute 60xx, 0102) = 12, 13, 14, 15 (only bits above high bit permitted)
  - Graphic Overlay will be shifted to fill Overlay Planes from Bit 12 and consecutive.
- Overlay plane
  - overlay type (attribute 60xx, 0040) = "G"
  - bits allocated (attribute 60xx, 0100) = 1
  - bit position (attribute 60xx, 0102) = 0
  - overlay data (attribute 60xx, 3000) = supported

The ARCADIS Multi-Modality Viewing application accepts also the MONOCHROME1 and MONOCHROME2 photometric interpretation pixel format with binary 2's complement integer and 16 bits allocated. Accepted values:

- Pixel plane
  - samples per pixel (attribute 0028, 0002) = 1
  - photometric interpretation (attribute 0028,0004) = "MONOCHROME1"
  - photometric interpretation (attribute 0028,0004) = "MONOCHROME2"
  - pixel representation (attribute 0028, 0103) = 1 (signed)
  - bits allocated (attribute 0028, 0100) = 16

- bits stored (attribute 0028,0101) = 16
- high bit (attribute 0028,0102) = 15
- only aspect ratio 1:1 is supported
- Overlay plane
  - overlay type (attribute 60xx, 0040) = "G"
  - bits allocated (attribute 60xx, 0100) = 1
  - bit position (attribute 60xx, 0102) = 0
  - overlay data (attribute 60xx, 3000) = supported
  - For MOD LUT, both the linear LUT (Rescale Slope/Intercept) and the MOD LUT SQ are supported and considered when pixel data is displayed. However there are two limitations. The MOD LUT SQ will be ignored in the following cases:
- 8-Bit signed pixels
- the pixel format is changed by the MOD LUT (e.g. 8bit -> 16bit)

If the MOD LUT SQ contains multiple LUTs, then only the first one is used.

For VOI LUT, both the linear LUT (Window Center/Width) and the VOI LUT SQ are supported (VOI LUT SQ with 8 or 16 bit LUT data)

But if both, a VOI LUT SQ and a linear MOD LUT, are specified within one image, then the value for Rescale Slope is restricted to 1.

If the VOI LUT SQ contains multiple LUTs, then only the first one is used by default. The other VOI LUTs are selectable.

Only Rectangular and Circular Shutter Shape is supported in this version. Images containing other Shutter Shapes will be displayed w/o shutter.

#### 5.1.3.1.3.2 Image Pixel Attribute Acceptance Criterion for Color Images – Viewing

The ARCADIS Multi-Modality Viewing application supports the RGB color image description with the unsigned integer 24-bit color image plane pixel format. Accepted values:

- samples per pixel (attribute 0028, 0002) = 3
- photometric interpretation (attribute 0028,0004) = "RGB"
- pixel representation (attribute 0028, 0103) = 0
- bits allocated (attribute 0028, 0100) = 8
- bits stored (attribute 0028,0101) = 8
- high bit (attribute 0028,0102) = 7
- planar configuration (attribute 0028,0006) = 0 (pixel interleave) or 1 (plane interleave).

The ARCADIS Multi-modality Viewing application supports the "Palette Color" color image description with the unsigned integer and 2's complement pixel format. Accepted values:

- samples per pixel (attribute 0028, 0002) = 1
- photometric interpretation (attribute 0028,0004) = "PALETTE COLOR"
- pixel representation (attribute 0028, 0103) = 0
- bits allocated (attribute 0028, 0100) = 8 and bits stored (attribute 0028,0101) = 8

- bits allocated (attribute 0028, 0100) = 16 and bits stored (attribute 0028,0101) = 16
- high bit (attribute 0028,0102) = 7, 15

Both 8-bit and 16-bit palettes are supported, but NO Segmented Palette Color LUTs.

The ARCADIS Multi-modality Viewing application supports the YBR\_FULL color image description with the unsigned integer pixel format. Accepted values:

- samples per pixel (attribute 0028, 0002) = 3
- photometric interpretation (attribute 0028,0004) = "YBR\_FULL"
- pixel representation (attribute 0028, 0103) = 0
- bits allocated (attribute 0028, 0100) = 8 and bits stored (attribute 0028,0101) = 8
- high bit (attribute 0028,0102) = 7

If the ARCADIS software is making any persistent changes on an YBR image, the resulting new image will be saved with Photometric Interpretation = "RGB".

#### 5.1.3.1.4 Presentation Context Acceptance Criterion

The ARCADIS DICOM application will accept any number of verification or storage SOP classes that are listed above. The number of presentation contexts accepted is limited to the maximum of 127 (DICOM limit). In the event that the ARCADIS DICOM application runs out of resources, it will reject the association request.

#### 5.1.3.1.5 Transfer Syntax Selection Policies

The ARCADIS DICOM application currently supports

- the Implicit VR Little Endian, the Explicit VR Little Endian and Explicit VR Big Endian Transfer Syntaxes
- the JPEG Lossless Non-hierarchical Transfer Syntax
- the JPEG Baseline and JPEG Extended Transfer Syntaxes (JPEG Lossy).
- the RLE Lossless Transfer Syntax

Any proposed presentation context including one of these Transfer Syntaxes will be accepted. Any proposed presentation context that does not include one of these Transfer Syntaxes will be rejected.

The order of preference in accepting Transfer Syntaxes within Presentation Contexts or Presentation Contexts with single Transfer Syntaxes is:

1. JPEG Lossy Extended
2. JPEG Lossless non-hierarchical
3. JPEG Lossy Baseline
4. RLE Lossless
5. Explicit VR Little Endian
6. Explicit VR Big Endian
7. Implicit VR Little Endian



With RLE Lossless Transfer Syntax the ARCADIS DICOM application will decompress the image before storing it into the database.

With Implicit VR Little Endian Transfer Syntax the ARCADIS DICOM application will remove any Private Attributes not known to the application. Decision on removal of a Private Element is done if there is NO entry in the attribute-dictionary of the ARCADIS DICOM application.

Therefore the Storage SCU's shall preferably use any Explicit VR Transfer Syntax when sending Composite Image Instances to the ARCADIS DICOM application.

## 6 Implementation Model Storage Commitment

The Storage Commitment service class defines an application-level class of service, which facilitates the commitment to storage. It performs an additional task of commitment of composite objects apart from the network-based storage of images as defined by the Storage Service class. The ARCADIS DICOM implementation supports the Storage Commitment Push Model as SCU and SCP.

### 6.1 Application Data Flow Diagram

The ARCADIS DICOM network implementation acts as SCU/SCP for the Storage Commitment Push Model Service SOP Class.

The product target Operating System is Windows XP.




Figure 3: Application Data Flow Diagram – Storage Commitment SCU/SCP

### 6.2 Functional Definitions of Application Entities

With each successfully completed send job, the ARCADIS DICOM Application will trigger a Storage Commit Request, if configured. Depending on configuration, the ARCADIS DICOM application will keep the association open for responses with a configurable time-out, or closes the association and expects responses on a different association that has to be established by the remote Storage Commitment SCP.

The commitment status derived from the related trigger response will be indicated in the related Status Flags of the related entity. It is possible to create triggers (“auto rules”) from this event.

The Transaction UIDs of the pending commitment request are kept “open” for a configurable amount in time (default: 1h). If the “open time” for a pending commitment request has elapsed without receiving a related response from the provider, the Transaction UID is removed and the related entities are indicated as “commit failed”.

In any case, commitment will only be requested for previously and successfully sent images.

The Storage Commitment SCP is running in background and is ready to receive request when the system is started.

### **6.3 Sequencing of real World Activities**

The Storage Commitment trigger is automatically derived from the successful completion of a Send Job.

## 7 AE Specification Storage Commitment

### 7.1 Storage Commitment AE Specification

SIEMENS ARCADIS DICOM application provides Standard Conformance to the following DICOM V3.0 SOP Class as an SCU and SCP:

SOP Class Name	SOP Class UID
Storage Commitment Push Model	1.2.840.10008.1.20.1

#### 7.1.1 Association Establishment Policies

##### 7.1.1.1 General

With a Send Job successfully completed, the DICOM application will open a new association, sent a Storage Commitment Request over that single opened association and wait for Status responses of the Storage Commitment Request. If the Provider accepts the Storage Commitment with Success Status, the generated Transaction UID, together with study identification data and a time-stamp, is kept. Depending on configuration, the association is closed when the configured time-out has elapsed or a response was received before. If the association is closed before a response was received, the response is then expected on a different association. Multiple Storage Commitment Requests can be pending.

The default PDU size used will be 28 KB.

##### 7.1.1.2 Number of Associations

The ARCADIS DICOM application initiates several associations at a time, one for each destination to which a transfer request is being processed in the active job queue list.

The ARCADIS DICOM application is able to accept multiple associations at a time. It can handle up to 10 associations in parallel.

##### 7.1.1.3 Asynchronous Nature

The ARCADIS DICOM software does not support asynchronous communication (multiple outstanding transactions over a single association).

##### 7.1.1.4 Implementation Identifying Information

Implementation Class UID	1.3.12.2.1107.5.12
Implementation Version Name	SIEMENS_ ASPVC10A

**7.1.2 Association Initiation Policy**

The ARCADIS DICOM Application Entity acts as a Service Class User (SCU) for the Storage Commitment Push Model Service Class (to request commitment for storage of instances previously sent).

To do so, ARCADIS will issue a

- N-ACTION DIMSE to request commitment or a
- N-EVENT-REPORT DIMSE to respond to a received storage commitment request and the association was closed by the remote system prior to response.

**7.1.2.1 Real World Activity – Storage Commitment**

**7.1.2.1.1 Associated Real-World Activity - Job Completed**

The ARCADIS Storage Commitment application sends the commit request (N-ACTION-RQ) message and waits for acceptance of this request (N-ACTION-RSP). After receiving this, the transaction is marked as “waiting”.

Depending on a configuration value, the association will then be closed or kept open. In the first case, there is another configurable timeout giving the number of hours (h) and minutes (m) (by default 1h:0m) to wait for the corresponding commit response (N-EVENT-REPORT). In the second case, this time is the (also configurable) time-out for the association. For both cases, if the commit response (N-EVENT-REPORT) does not arrive during the configured time, the transaction will be marked as failed. The ARCADIS does not re-send objects from a failed Storage Commitment result in any case.

If the commit response (N-EVENT-REPORT) received has the status of “complete - failure exists”, the transaction is marked as failed, else the transaction is marked as “completed”; In both cases, a message is shown to the user.

**7.1.2.1.2 Proposed Presentation Contexts - Job Completed**

The ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Storage Commitment Push Model	1.2.840.10008.1.20.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None

**7.1.2.1.3 SOP Specific Conformance Statement- Job Completed**

Storage Commitment is supported for all the SOP class UIDs as mentioned in 'Acceptable presentation contexts - Storage' in the Storage SCP section of this document.

The Referenced Study Component Sequence is not supported.

Storage Media File-Set ID and UID Attributes will not be supported in the commitment request (N-ACTION primitive) invoked by the Storage Commitment SCU.

**7.1.2.1.4 Associated Real-World Activity - Send Commit Response**

Acting as a Storage Commitment Provider, the ARCADIS Storage Commitment AE received a Storage Commitment request, carried out the request, and is ready to send back the response, but the association is not open anymore. In this case it will by itself initiate an association to send the storage commitment response (N-EVENT-REPORT) to the SCU.

**7.1.2.1.5 Proposed Presentation Contexts - Send Commitment Response**

The Siemens ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Storage Commitment Push Model	1.2.840.10008.1.20.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None

**7.1.2.1.6 SOP Specific Conformance Statement - Send Commitment Response**

Storage Media File-Set ID and UID Attributes will not be supported in the N-EVENT-REPORT primitive invoked by the Storage Commitment SCP.

**7.1.3 Association Acceptance Policy**

The ARCADIS DICOM Application Entity acts as a Service Class Provider (SCP) for the Storage Commitment Push Model Service Class (Give a commitment to store previously received instances).

To do so, the ARCADIS attempts to accept a

- N-ACTION DIMSE to receive a commitment request for the instance included or a
- N-EVENT-REPORT DIMSE to receive a storage commitment response from a previous request.

**7.1.3.1 Associated Real-World Activity - Commit SCP**

**7.1.3.1.1 Associated Real-World Activity - Receive Commit Request**

When receiving a Storage Commitment request the ARCADIS DICOM application will perform the necessary steps to check the received list of Instances against the local database or, if configured, check the Instances with the attached archive system.

**7.1.3.1.2 Accepted Presentation Contexts - Receive Commit Request**

The Siemens ARCADIS DICOM application will accept Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Storage Commitment Push Model	1.2.840.10008.1.20.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None

**7.1.3.1.3 SOP-specific Conformance Statement - Receive Commit Request**

The ARCADIS Storage Commitment AE will return success for images that are stored in the local database and failure for images that are not.

Remark: Sending data with Storage Commitment via network is a safe data transfer but does not fulfill the regulatory requirements of long-term archiving; objects with the "committed" flag may be deleted by the user.

**7.1.3.2 Associated Real-World Activity - Commit SCU**

**7.1.3.2.1 Associated Real-World Activity - Update Flags**

The ARCADIS Storage Commitment DICOM Application has sent a Storage Commitment Request and, being configured to receive response on a separate association, has closed the association, and now it gets an association request from the Storage Commitment SCP that want to send the results. The ARCADIS DICOM application will await Storage commitment Notification triggers. Any incoming Notification will be checked for validity, that is, if the related Transaction UID is still part of the Pending Request Queue.

If the Notification is valid, the Notification Identifier is evaluated and the related Instances marked with the related status. The over-all Commit Status of the higher Information Entities is derived from propagation of the States of all Image entities included in a study.

The Status Flags directly affected by Storage Commitment results and indicated in the different entities of the Patient Browser list can be one of

- “AC” or “SC” - Successful Commitment, A means archived to configured Archive destination, whereas S means sent to any other destination
- “Af” of “Sf” - Commitment failed.
- “A?” or “S?” - Commitment request is sent, response is pending.

In case of failure the user has to repeat the transfer of images to the Archive destination. Another Storage Commitment will be performed after sending is completed successfully.

**7.1.3.2.2 Accepted Presentation Contexts - Update Flags**

The Siemens ARCADIS DICOM application will accept Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Storage Commitment Push Model	1.2.840.10008.1.20.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None

**7.1.3.2.3 SOP-specific Conformance Statement - Update Flags**

If the Commitment response (N-EVENT-REPORT) received has the status of “complete - failure exists”, the transaction is marked as failed, else the transaction is marked as “completed”; in both cases a message is shown to the user.

The related status flags are set for the committed images in the local database.

The ARCADIS DICOM application will NOT support the Storage Media File Set ID attributes.



## 8 Implementation Model Query / Retrieve

The query/retrieve service class defines an application-level class of services, which facilitates the management of images and patient data against the well-defined information model of DICOM and allows a DICOM AE to retrieve images from a remote DICOM node or to request a remote DICOM AE to initiate a transfer of images to another DICOM AE. The ARCADIS DICOM query/retrieve application supports the query/retrieve services to act as SCU and SCP.

### 8.1 Application Data Flow Diagram

The ARCADIS DICOM network implementation acts as SCU and SCP for the query/retrieve network service. The product target Operating System is Windows XP.




Figure 4: ARCADIS Application Data Flow Diagram – Query/Retrieve SCU

### 8.2 Functional Definitions of Application Entities

The ARCADIS DICOM query/retrieve SCU requests the remote query/retrieve SCP to perform a search and match to the keys specified in the request in order to display the results in the ARCADIS user interface. Depending on user action (Import) the ARCADIS DICOM SCU sends a C-MOVE DIMSE to initiate a C-STORE sub-operation on the SCP to start an image transfer from remote Storage SCU (running on Query/Retrieve SCP) to the ARCADIS Storage SCP.

The ARCADIS DICOM query/retrieve SCP responds to C-FIND DIMSE from remote SCU applications. Depending on further remote request, a C-GET or a C-MOVE involves the ARCADIS DICOM query/retrieve SCP application to initiate a C-STORE association (by triggering the own Storage SCU) to send image objects to a remote Storage SCP.

All components of the DICOM query/retrieve SCP application are operating as background server processes. They exist when the machine is powered on and then respond to queries based on the records stored in its database.

### **8.3 Sequencing of Real-World Activities**

Retrieve of images is only possible if results from a previous “Search...” operation exist and those entities can be selected for “Import”.

## 9 Application Entity Specification Query/Retrieve

### 9.1 Query/Retrieve Service AEs Specification

The Query/Retrieve SCU requests that the remote SCP performs a match of all keys specified in the request, against the information in its database and the identified images will be moved over a different (C-MOVE) storage association.

The Query/Retrieve SCP responds to queries based on the records based on its database and images will be sent to the requesting SCU or to a different storage destination.

SIEMENS ARCADIS DICOM products provide Standard Conformance to the following DICOM V3.0 SOP Classes as SCU and SCP:

SOP Class Name	SOP Class UID
Patient Root Query/Retrieve Information Model - FIND	1.2.840.10008.5.1.4.1.2.1.1
Patient Root Query/Retrieve Information Model - MOVE	1.2.840.10008.5.1.4.1.2.1.2
Study Root Query/Retrieve Information Model - FIND	1.2.840.10008.5.1.4.1.2.2.1
Study Root Query/Retrieve Information Model - MOVE	1.2.840.10008.5.1.4.1.2.2.2
Patient/Study Only Query/Retrieve Information Model - FIND	1.2.840.10008.5.1.4.1.2.3.1
Patient/Study Only Query/Retrieve Information Model – MOVE	1.2.840.10008.5.1.4.1.2.3.2

**Note:** See also the Storage DICOM Conformance Statement of the ARCADIS DICOM application to compare for conformance of the C-STORE sub-operation generated by the C-MOVE DIMSE service. Furthermore compare the supported Storage Service SOP classes described in the Storage DICOM Conformance Statement of the Modality to which the images shall be transferred.

#### 9.1.1 Association Establishment Policies

##### 9.1.1.1 General

With the “Search...” function the query data are input and the DICOM query/retrieve application is started. A query request will be sent out to one remote node that can be selected from a list of configured Query Providers and the response data will be displayed for the user. Upon request (Import), the retrieval of selected items is initiated.

The default PDU size used will be 28 KB.

##### 9.1.1.2 Number of Associations

The ARCADIS DICOM application initiates several associations at a time, one for each destination to which a transfer request is being processed in the active job queue list.

The ARCADIS DICOM application is able to accept multiple associations at a time. It can handle up to 10 associations in parallel.

**9.1.1.3 Asynchronous Nature**

The ARCADIS DICOM software does not support asynchronous communication (multiple outstanding transactions over a single association).

**9.1.1.4 Implementation Identifying Information**

Implementation Class UID	1.3.12.2.1107.5.12
Implementation Version Name	SIEMENS_ ASPVC10A

**9.1.2 Association Initiation Policy**

The query user interface will request the query-data from user and triggers one C-FIND request to the selected remote node. The response data will be displayed in the query UI for further data navigation.

When requesting Import of related items the browser requests the retrieve application to send a C-MOVE request to the related remote node. The Storage SCP as described in the related section will then receive images.

**9.1.2.1 Real World Activity - Find SCU**

**9.1.2.1.1 Associated Real-World Activity - Find SCU “Search”**

The associated Real-World activity is to fill out a query form with search data and pass it as query to the network application, which issues a C-FIND over a previously built association. The remote SCP will respond with related data-entries that will be passed to a browser application. When data transfer is finished the association is closed.

**9.1.2.1.2 Proposed Presentation Contexts - Find SCU**

The ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Patient Root Query/Retrieve Model – FIND	1.2.840.10008.5.1.4.1.2.1.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Study Root Query/Retrieve Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Patient/Study Only Query/Retrieve Model – FIND	1.2.840.10008.5.1.4.1.2.3.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None

It is configurable which of the query models (or all) are to be used by the ARCADIS DICOM Query SCU application. If all Abstract Syntaxes are configured, the C-FIND SCU will use the Patient Root Model only for C-FIND requests on PATIENT level. For all other levels it will use the STUDY root model. If the remote nodes only support the Patient/Study Only Query/Retrieve Model, the C-FIND SCU will use this one.

## 9.1.2.1.3 Conformance Statement - Find SCU

The ARCADIS DICOM Query/Retrieve SCU supports hierarchical queries with all mandatory search keys. The Query SCU does not support the interactive querying of attributes on IMAGE level. The following table describes the search keys for the different query models that the SCU supports. Matching is either wildcard, which means that the user can supply a string containing wildcards, or universal, which means that the attribute is requested as return value.

Attribute name	Tag	Type	Matching	User input	return value display
<b>Patient Level <sup>a</sup></b>					
Patient Name	(0010,0010)	R	Wildcard <sup>b</sup>	enter value	yes
Patient ID	(0010,0020)	U	Wildcard <sup>b</sup>	enter value	yes
Patient's Birth date	(0010,0030)	O	universal (Null)	enter value	yes
Patient's Sex	(0010,0040)	O	universal (Null)	enter value	yes
Number of Patient related Studies	(0020,1200)	O	universal (Null)	-	yes <sup>c</sup>
Number of Patient related Series	(0020,1202)	O	universal (Null)	-	no
Number of Patient related Instances	(0020,1204)	O	universal (Null)	-	no
<b>Study Level</b>					
Patient Named	(0010,0010)	R	Wildcard <sup>b</sup>	enter value	yes
Patient ID	(0010,0020)	R	Wildcard <sup>b</sup>	enter value	yes
Patient's Birth date <sup>d</sup>	(0010,0030)	O	universal (Null)	enter value	yes
Patient's Sex <sup>d</sup>	(0010,0040)	O	universal (Null)	enter value	yes
Study Instance UID	(0020,000D)	U	single value	-	yes
Study ID	(0020,0010)	R	universal (Null)	enter value	yes
Study Date	(0008,0020)	R	universal (Null)	enter value	yes
Study Time	(0008,0030)	R	universal (Null)	-	yes
Accession Number	(0008,0050)	R	universal (Null)	-	yes
Study Description	(0008,1030)	O	universal (Null)	-	yes
Referring Physician's Name	(0008,0090)	O	universal (Null)	-	yes
Name of Physician Reading Study	(0008,1060)	O	universal (Null)	-	yes
Modalities in Study	(0008,0061)	O	universal (Null)	-	yes
Storage Media File-Set ID	(0008,0130)	O	universal (Null)	-	no
Retrieve AE Title	(0008,0054)	O	universal (Null)	-	no
Number of Study related Series	(0020,1206)	O	universal (Null)	-	yes <sup>e</sup>

a Patient Root Information Model only

b Always a '\*' is appended to the user-supplied string

c Implicitly visualized in the UI if no study and series search attributes have been entered

d Study Root Information Model only

e Implicitly if no series search attributes have been entered

Attribute name	Tag	Type	Matching	User input	return value display
Number of Study related Instances	(0020,1208)	O	universal (Null)	-	no
<b>Series Level</b>					
Series Instance UID	(0020,000E)	U	single value	-	yes
Series Number	(0020,0011)	R	universal (Null)	-	yes
Modality	(0008,0060)	R	universal (Null)	enter value	yes
Series Date	(0008,0021)	O	universal (Null)	-	yes
Series Time	(0008,0031)	O	universal (Null)	-	yes
Series Description	(0008,103E)	O	universal (Null)	enter value	yes
Body Part Examined	(0018,0015)	O	universal (Null)	enter value	yes
Performing Physician	(0008,1050)	O	universal (Null)	enter value	yes
Storage Media File-Set ID	(0008,0130)	O	universal (Null)	-	yes
Retrieve AE Title	(0008,0054)	O	universal (Null)	-	yes
Protocol Name	(0018,1030)	O	universal (Null)	-	no
Perf. Procedure Step Start Date	(0040,0244)	O	universal (Null)	-	yes
Perf. Procedure Step Start Time	(0040,0245)	O	universal (Null)	-	yes
Requested Attribute Sequence	(0040,0275)	O	universal (Null)	-	yes
> Requested Procedure ID	(0040,1001)	O	universal (Null)	-	yes
> Scheduled Procedure ID	(0040,0009)	O	universal (Null)	-	yes
Number of Series related Instances	(0020,1209)	O	universal (Null)	-	yes
<b>Image Level</b>					
SOP Instance UID	(0008,0018)	U	single value	-	no
Image Number	(0020,0013)	R	universal (Null)	-	yes
Storage Media File-Set ID	(0008,0130)	O	universal (Null)	-	no
Retrieve AE Title	(0008,0054)	O	universal (Null)	-	no
Instance Date	(0008,0023)	O	universal (Null)	-	no
Instance Time	(0008,0033)	O	universal (Null)	-	no
Number of Frames	(0028,0008)	O	universal (Null)	-	yes
Image Comments	(0020,4000)	O	universal (Null)	-	no
Referenced Request Sequence	(0040,A370)	O	sequence matching	-	Yes
>Accession Number	0008,0050)	O	single value, universal	-	Yes
>Requested Procedure ID	(0040,1000)	O	single value, universal	-	Yes

Attribute name	Tag	Type	Matching	User input	return value display
Concept Name Code Sequence	(0040,A043)	O	sequence matching	enter value	Yes
>Code Value	(0008,0100)	O	single value, universal, wildcard	-	Yes
>Coding Scheme Designator	(0008,0102)	O	single value, universal, wildcard	-	Yes
>Coding Scheme Version	(0008,0103)	O	single value, universal, wildcard	-	Yes
>Code Meaning	(0008,0104)	O	single value, universal, wildcard	enter value	Yes
Template Identifier	(0040,DB00)	O	single value, universal, wildcard	-	Yes
Completion Flag	(0040,A491)	O	single value, universal, wildcard	enter value	Yes
Verification Flag	(0040,A493)	O	single value, universal, wildcard	enter value	Yes
Verifying Observer Sequence	(0040,A073)	O	sequence matching	enter value	Yes
>Verifying Organization	(0040,A027)	O	single value, universal, wildcard	-	Yes
>Verifying DateTime	(0040,A030)	O	single value, range matching, universal	enter value	Yes
>Verifying Observer Name	(0040,A075)	O	single value, universal, wildcard	enter value	Yes
>Verifying Observer Identification Code Sequence	(0040,A088)	O	sequence matching	-	Yes
>>Code Value	(0008,0100)	O	single value, universal, wildcard	-	Yes
>>Coding Scheme Designator	(0008,0102)	O	single value, universal, wildcard	-	Yes
>>Coding Scheme Version	(0008,0103)	O	single value, universal, wildcard	-	Yes
>>Code Meaning	(0008,0104)	O	single value, universal, wildcard	-	yes

The Find SCU interprets following status codes:

Service Status	Meaning	Protocol Codes	Related Fields
----------------	---------	----------------	----------------

Refused	Out of Resources	A700	(0000,0902)
Failed	Identifier does not match SOP Class	A900	(0000,0901) (0000,0902)
	Unable to process	Cxxx	(0000,0901) (0000,0902)
Cancel	Matching terminated due to Cancel request	FE00	None
Success	Matching is complete - No final Identifier is supplied	0000	None
Pending	Matches are continuing - Current Match is supplied and any Optional Keys were supported in the same manner as Required Keys	FF00	Identifier
	Matches are continuing - Warning that one or more Optional Keys were not supported for existence and/or matching for this identifier	FF01	Identifier

**9.1.2.2 Real-World Activity – Move SCU**

**9.1.2.2.1 Associated Real-World Activity – Move SCU “Import”**

When selecting a data entry in the Query UI and activate the “Import” function, a retrieval request is passed to the archival application which issues a C-MOVE service according to the Patient Root or Study Root query model. (The Storage Service Class Conformance Statement describes the C-STORE service, which is generated by processing the C-MOVE service.)

The transferred image data are processed as described in the storage class SCP descriptions.

The possibility to request the remote C-MOVE provider (remote application that responded to the C-FIND) to move data to an application entity other than the C-MOVE SCU (the ARCADIS DICOM application) is NOT USED.

The Query UI does not support C-MOVE operation on Patient Level.



**9.1.2.2.2 Proposed Presentation Contexts - Move SCU “Import”**

The ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Patient Root Query/Retrieve Model – MOVE	1.2.840.10008.5.1.4.1.2.1.2	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Study Root Query/Retrieve Model – MOVE	1.2.840.10008.5.1.4.1.2.2.2	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Patient/Study Only Query/Retrieve Model – MOVE	1.2.840.10008.5.1.4.1.2.3.2	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None

**Note:** C-MOVE extended negotiation will not be supported by the SCU

**9.1.2.2.3 SOP Specific Conformance Statement - Move SCU “Import”**

At association establishment time the C-MOVE presentation context shall be negotiated. The C-STORE sub-operations must be done on a different association to transfer images to the own Storage Service Class SCP.

The Move SCU interprets following status codes:

Service Status	Meaning	Error Codes	Related Fields
Refused	Out of Resources - Unable to calculate number of matches	A701	(0000,0902)
	Out of Resources - Unable to perform sub operations	A702	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Failed	Identifier does not match SOP Class	A900	(0000,0901) (0000,0902)
	Unable to process	CXXX	(0000,0901) (0000,0902)
Cancel	Sub-operations terminated due to Cancel Indication	FE00	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Warning	Sub-operations Complete - One or more Failures or Warnings	B000	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Success	Sub-operations Complete - No Failures or Warning	0000	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Pending	Sub-operations are continuing	FF00	(0000,1020) (0000,1021) (0000,1022) (0000,1023)

### 9.1.3 Association Acceptance Policy

The ARCADIS DICOM application will accept associations for the following DIMSE-C operations as SCP:

- C-FIND
- C-MOVE
- C-FIND-CANCEL
- C-MOVE-CANCEL

#### 9.1.3.1 Real-World Activity - Find SCP

##### 9.1.3.1.1 Associated Real-World Activity - Find SCP

The associated Real-World activity is to respond query requests to an SCU with the query model Patient Root, Study Root and Patient/Study Only. Relational retrieve operation is NOT supported. With a C-FIND-CANCEL request the running query can be canceled at any time.

Multiple C-FIND requests over the same association are supported.

##### 9.1.3.1.2 Accepted Presentation Contexts - Find SCP

The ARCADIS DICOM application will accept Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Patient Root Query/Retrieve Model – FIND	1.2.840.10008.5.1.4.1.2.1.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None
Study Root Query/Retrieve Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None
Patient/Study Only Query/Retrieve Model – FIND	1.2.840.10008.5.1.4.1.2.3.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None

**Note:** C-FIND Extended Negotiation will NOT be supported.  
The order of preference for accepting Transfer Syntaxes is: 1. Explicit VR Little Endian, 2. Explicit VR Big Endian, 3. Implicit VR Little Endian

##### 9.1.3.1.3 SOP Specific Conformance Statement - Find SCP

The ARCADIS DICOM Query/Retrieve SCP supports hierarchical queries with all mandatory and optional search keys.

The query attribute contents will be treated case-sensitive.

With wildcard queries the symbol “?” is also treated as “\*” by the C-FIND SCP application. As a consequence the query string of “?abc\*” will be processed as “\*abc\*”.

If the value for the patient-level unique key “Patient ID” is not known, it may be returned with zero length. The attribute “Image Comments” will not be included in the C-FIND-RSP, if it is not set in the DB, even if it was requested as return key in the related C-FIND-RQ.

Usage of Storage Media File-Set ID, Retrieve AE Title with C-FIND-RSP message:

- The Storage Media File-Set ID - if existent - can be returned at Study/Series/Image Level. Only on Image Level, the values of ONLINE, NEARLINE or OFFLINE are returned to indicate the Storage Location of the related Instance.
- The Retrieve AE Title - if existent - can only be returned at Image Level (for Patient Root and Study Root models) or Study Level (for Patient/Study Only model).

Relational Queries are **not** supported.

A remote DICOM AE can cancel the running query by sending a C-FIND-CANCEL. Matches are possibly continuing (more C-FIND response with status PENDING) until the cancel operation has completed.

The supported attributes on the various query levels of the three supported information models are listed in the tables of the following sections.

**9.1.3.1.3.1 Patient Root Information Model**

Attribute Name	Tag	Usage SCU	Matching
<b>Patient Level</b>			
Patient Name	(0010,0010)	R	single value, wildcard, universal
Patient ID	(0010,0020)	U	single value, wildcard, universal
Patient’s Birth Date	(0010,0030)	O	single value, range, universal
Patient’s Birth Time	(0010,0032)	O	single value, range, universal
Patient’s Sex	(0010,0040)	O	single value, wildcard, universal
Ethnic Group	(0010,2160)	O	single value, wildcard, universal
Patient Comments	(0010,4000)	O	wildcard, universal
Number of Patient related Studies	(0020,1200)	O	Universal
Number of Patient related Series	(0020,1202)	O	Universal
Number of Patient related Instances	(0020,1204)	O	Universal
<b>Study Level</b>			
Study Instance UID	(0020,000D)	U	single value, list of UIDs
Study ID	(0020,0010)	R	single value, wildcard, universal
Study Date	(0008,0020)	R	single value, range, universal
Study Time	(0008,0030)	R	single value, range, universal
Accession Number	(0008,0050)	R	single value, wildcard, universal
Referring Physician’s Name	(0008,0090)	O	single value, wildcard, universal
Study Description	(0008,1030)	O	single value, wildcard, universal
Admitting Diagnoses Description	(0008,1080)	O	single value, wildcard, universal
Patient’s Age	(0010,1010)	O	single value, wildcard, universal
Patient’s Size	(0010,1020)	O	single value, universal
Patient’s Weight	(0010,1030)	O	single value, universal
Occupation	(0010,2180)	O	single value, wildcard, universal
Additional Patient History	(0010,21B0)	O	wildcard, universal
Name of Physician reading Study	(0008,1060)	O	single value, wildcard, universal

Attribute Name	Tag	Usage SCU	Matching
Modalities in Study	(0008,0061)	O	multiple values, universal
Number of Study related Series	(0020,1206)	O	Universal
Number of Study related Instances	(0020,1208)	O	Universal
<b>Series Level</b>			
Series Instance UID	(0020,000E)	U	single value, list of UID
Series Number	(0020,0011)	R	single value, universal
Modality	(0008,0060)	R	single value, wildcard, universal
Laterality	(0020,0060)	O	single value, wildcard, universal
Body Part Examined	(0018,0015)	O	single value, wildcard, universal
Patient Position	(0018,5100)	O	single value, wildcard, universal
Smallest Pixel Value in Series	(0028,0108)	O	single value, universal
Largest Pixel Value in Series	(0028,0109)	O	single value, universal
Protocol Name	(0018,1030)	O	single value, wildcard, universal
Series Date	(0008,0021)	O	single value, range, universal
Series Time	(0008,0031)	O	single value, range, universal
Series Description	(0008,103E)	O	single value, wildcard, universal
Operators Name	(0008,1070)	O	single value, wildcard, universal
Performing Physician's Name	(0008,1050)	O	single value, wildcard, universal
Perf. Procedure Step Start Date	(0040,0244)	O	Universal
Perf. Procedure Step Start Time	(0040,0245)	O	Universal
Number of Series related Instances	(0020,1209)	O	Universal
<b>Image Level</b>			
SOP Instance UID	(0008,0018)	U	single value, list of UID
Image Number	(0020,0013)	R	single value, universal
Image Date	(0008,0023)	O	single value, range, universal
Image Time	(0008,0033)	O	single value, range, universal
Modality	(0008,0060)	O	single value, wildcard, universal
Image Comments	(0020,4000)	O	Universal

Supported Query attributes sorted by Query Level – Patient Root Information Model

### 9.1.3.1.3.2 Study Root Information Model

Attribute Name	Tag	Usage SCU	Matching
<b>Study Level</b>			
Patient Name	(0010,0010)	R	Single value, Wildcard, universal
Patient ID	(0010,0020)	R	Single Value, Wildcard, universal
Patient's Birth Date	(0010,0030)	O	Single Value, Range, universal
Patient's Birth Time	(0010,0032)	O	Single Value, Range, universal
Patient's Sex	(0010,0040)	O	Single Value, Wildcard, universal
Patient Comments	(0010,4000)	O	Wildcard, universal
Number of Patient related Studies	(0020,1200)	O	Universal
Number of Patient related Series	(0020,1202)	O	Universal
Number of Patient related Instances	(0020,1204)	O	Universal
Study Instance UID	(0020,000D)	U	Single Value, List of UIDs
Study ID	(0020,0010)	R	Single Value, Wildcard, universal
Study Date	(0008,0020)	R	Single Value, Range, universal
Study Time	(0008,0030)	R	Single Value, Range, universal
Accession Number	(0008,0050)	R	Single Value, Wildcard, universal
Referring Physician's Name	(0008,0090)	O	Single Value, Wildcard, universal

Attribute Name	Tag	Usage SCU	Matching
Study Description	(0008,1030)	O	Single Value, Wildcard, universal
Admitting Diagnosis Description	(0008,1080)	O	Single Value, Wildcard, universal
Patient's Age	(0010,1010)	O	Single Value, Wildcard, universal
Patient's Size	(0010,1020)	O	Single Value, universal
Patient's Weight	(0010,1030)	O	Single Value, universal
Occupation	(0010,2180)	O	Single Value, Wildcard, universal
Additional Patient History	(0010,21B0)	O	Wildcard, universal
Name of Physician reading the Study	(0008,1060)	O	Single Value, Wildcard, universal
Modalities in Study	(0008,0061)	O	Multiple values, universal
Number of Study Related Series	(0020,1206)	O	Universal
Number of Study Related Instances	(0020,1208)	O	Universal
<b>Series Level</b>			
Series Instance UID	(0020,000E)	U	Single Value, List of UIDs
Series Number	(0020,0011)	R	Single Value, universal
Modality	(0008,0060)	R	Single Value, Wildcard, universal
Laterality	(0020,0060)	O	Single Value, Wildcard, universal
Body Part Examined	(0018,0015)	O	Single Value, Wildcard, universal
Patient Position	(0018,5100)	O	Single Value, Wildcard, universal
Smallest Pixel Value in Series	(0028,0108)	O	Single Value, universal
Largest Pixel Value in Series	(0028,0109)	O	Single Value, universal
Protocol Name	(0018,1030)	O	Single Value, Wildcard, universal
Series Date	(0008,0021)	O	Single Value, Range, universal
Series Time	(0008,0031)	O	Single Value, Range, universal
Series Description	(0008,103E)	O	Single Value, Wildcard, universal
Operator's Name	(0008,1070)	O	Single Value, Wildcard, universal
Performing Physician's Name	(0008,1050)	O	Single Value, Wildcard, universal
Performed Procedure Step Start Date	(0040,0244)	O	universal
Performed Procedure Step Start Time	(0040,0245)	O	universal
Number of Series related Instances	(0020,1209)	O	universal
<b>Image Level</b>			
SOP Instance UID	(0008,0018)	U	Single Value, List of UIDs
Image Number	(0020,0013)	R	Single Value, universal
Image Date	(0008,0023)	O	Single Value, Range, universal
Image Time	(0008,0033)	O	Single Value, Range, universal
Modality	(0008,0060)	O	Single Value, Wildcard, universal
Image Comments	(0020,4000)	O	universal
Concept Name Code Sequence	(0040,A043)	O	sequence matching
>Code Value	(0008,0100)	O	single value, universal, wildcard
>Coding Scheme Designator	(0008,0102)	O	single value, universal, wildcard
>Coding Scheme Version	(0008,0103)	O	single value, universal, wildcard
>Code Meaning	(0008,0104)	O	single value, universal, wildcard
Template Identifier	(0040,DB00)	O	single value, universal, wildcard
Completion Flag	(0040,A491)	O	single value, universal, wildcard
Verification Flag	(0040,A493)	O	single value, universal, wildcard
Verifying Observer Sequence	(0040,A073)	O	sequence matching
>Verifying Organization	(0040,A027)	O	single value, universal, wildcard
>Verifying DateTime	(0040,A030)	O	single value, range matching, universal
>Verifying Observer Name	(0040,A075)	O	single value, universal, wildcard
>Verifying Observer Identification Code Sequence	(0040,A088)	O	sequence matching
>>Code Value	(0008,0100)	O	single value, universal, wildcard
>>Coding Scheme Designator	(0008,0102)	O	single value, universal, wildcard
>>Coding Scheme Version	(0008,0103)	O	single value, universal, wildcard

Attribute Name	Tag	Usage SCU	Matching
>>Code Meaning	(0008,0104)	O	single value, universal, wildcard

Supported Query attributes sorted by Query Level – Study Root Information Model

### 9.1.3.1.3.3 Patient/Study Only Information Model

Attribute Name	Tag	Usage SCU	Matching
<b>Patient Level</b>			
Patient Name	(0010,0010)	R	Single value, Wildcard, universal
Patient ID	(0010,0020)	U	Single Value, Wildcard, universal
Patient's Birth Date	(0010,0030)	O	Single Value, Range, universal
Patient's Birth Time	(0010,0032)	O	Single Value, Range, universal
Patient's Sex	(0010,0040)	O	Single Value, Wildcard, universal
Ethnic Group	(0010,2160)	O	Single Value, Wildcard, universal
Patient Comments	(0010,4000)	O	Wildcard, universal
Number of Patient related Studies	(0020,1200)	O	universal
Number of Patient related Series	(0020,1202)	O	universal
Number of Patient related Instances	(0020,1204)	O	universal
<b>Study Level</b>			
Study Instance UID	(0020,000D)	U	Single Value, List of UIDs
Study ID	(0020,0010)	R	Single Value, Wildcard, universal
Study Date	(0008,0020)	R	Single Value, Range, universal
Study Time	(0008,0030)	R	Single Value, Range, universal
Accession Number	(0008,0050)	R	Single Value, Wildcard, universal
Referring Physician's Name	(0008,0090)	O	Single Value, Wildcard, universal
Study Description	(0008,1030)	O	Single Value, Wildcard, universal
Admitting Diagnosis Description	(0008,1080)	O	Single Value, Wildcard, universal
Patient's Age	(0010,1010)	O	Single Value, Wildcard, universal
Patient's Size	(0010,1020)	O	Single Value, universal
Patient's Weight	(0010,1030)	O	Single Value, universal
Occupation	(0010,2180)	O	Single Value, Wildcard, universal
Additional Patient History	(0010,21B0)	O	Wildcard, universal
Name of Physician reading the Study	(0008,1060)	O	Single Value, Wildcard, universal
Modalities in Study	(0008,0061)	O	Multiple values, universal
Number of Study Related Series	(0020,1206)	O	universal
Number of Study Related Instances	(0020,1208)	O	universal

Supported Query attributes sorted by Query Level – Patient/Study Only Information Model

The Find SCP returns following status codes:

Service Status	Meaning	Error Codes	Related Fields
Refused	Out of Resources	A700	(0000,0902)
Failed	Identifier does not match SOP Class	A900	(0000,0901) (0000,0902)
	Unable to process	C001	(0000,0901) (0000,0902)
Cancel	Matching terminated due to Cancel request	FE00	None
Success	Matching is complete - No final Identifier is supplied	0000	None
Pending	Matches are continuing - Current Match is supplied and any Optional Keys were supported in the same manner as Required Keys	FF00	Identifier
	Matches are continuing - Warning that one or more Optional Keys were not supported for existence and/or matching for this identifier	FF01	Identifier

**9.1.3.2 Real-World Activity - Move SCP**

**9.1.3.2.1 Associated Real-World Activity - Move SCP**

The associated Real-World activity is to respond to retrieve requests to an SCU. The SCP supports the query model Patient Root, Study Root and Patient/Study Only. The Storage Service Class Conformance Statement describes the C-STORE service, which is generated by the C-MOVE service. Relational retrieve operation is NOT supported.

Multiple C-MOVE requests over the same association are NOT supported.

**9.1.3.2.2 Accepted Presentation Contexts - Move SCP**

The ARCADIS DICOM application will accept Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Patient Root Query/Retrieve Model – MOVE	1.2.840.10008.5.1.4.1.2.1.2	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None
Study Root Query/Retrieve Model – MOVE	1.2.840.10008.5.1.4.1.2.2.2	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None
Patient/Study Only Query/Retrieve Model – MOVE	1.2.840.10008.5.1.4.1.2.3.2	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCP	None

**Note:** C-MOVE Extended negotiation will NOT be supported.  
The order of preference for accepting Transfer Syntaxes is: 1. Explicit VR Little Endian, 2. Explicit VR Big Endian, 3. Implicit VR Little Endian.

**9.1.3.2.3 SOP Specific Conformance Statement - Move SCP**

At association establishment time the C-MOVE presentation context shall be negotiated. The C-STORE sub-operations are done on a different association, specified in the C-MOVE request, to transfer images to a remote SCP of the Storage Service Class. Relational retrieve operation is NOT supported.

All unique keys have to be supplied according to the selected Query/Retrieve Level. The related tables in the C-FIND SCP section will give information about “U” marked key attributes.

The Move SCP returns following status codes:

Service Status	Meaning	Error Codes	Related Fields
----------------	---------	-------------	----------------

Service Status	Meaning	Error Codes	Related Fields
Refused	Out of Resources - Unable to calculate number of matches	A701	(0000,0902)
	Out of Resources - Unable to perform sub operations	A702	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Failed	Identifier does not match SOP Class	A900	(0000,0901) (0000,0902)
	Unable to process	C001	(0000,0901) (0000,0902)
Cancel	Sub-operations terminated due to Cancel Indication	FE00	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Warning	Sub-operations Complete - One or more Failures of Warnings	B000	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Success	Sub-operations Complete - No Failures or Warning	0000	(0000,1020) (0000,1021) (0000,1022) (0000,1023)
Pending	Sub-operations are continuing	FF00	(0000,1020) (0000,1021) (0000,1022) (0000,1023)



## 10 Implementation Model Print

The Print Management Service Classes define an application-level class of services, which facilitate the printing of images on a hardcopy medium. The print management SCU and print management SCP are peer DICOM print management application entities. The *ARCADIS* DICOM print application supports the print management DIMSE services to act as SCU.

### 10.1 Application Data Flow Diagram

The *ARCADIS* DICOM network implementation acts as SCU for the print management network service. The product target Operating System is Windows XP.




Figure 5: DICOM Application Data Flow Diagram – Print SCU

### 10.2 Functional Definition of Application Entities

The Print SCU is invoked by the user interface to setup film-sheet layout and whenever an image is ready to be printed on film. The Print SCU will hold and maintain all data needed to compile a complete film-sheet from the data (images, layout, configuration) received. Whenever a film-sheet is ready to print the related data is used to supply the Information to the SOP Classes of the Print Management Service Class. A queue is maintained, in order to intermediately store several film-sheets in case of resource problems on printer. The SCU will only supply and require the mandatory SOP Classes of the Print Management Service Class.

### 10.3 Sequencing of Real-World Activities

Not applicable

## 11 Application Entity Specification Print

### 11.1 Print Management AE Specification

The *ARCADIS* print management SCU (HCS) invokes print management DIMSE services to transfer images from the local AE to the remote SCP AE to print images with defined layout on a selected network-based DICOM hardcopy printer. This is done in a “full-page” print mode.

SIEMENS *ARCADIS* DICOM products provide Standard Conformance to the following DICOM V3.0 Print Management Meta SOP Classes as an SCU:

SOP Class Name	SOP Class UID
Basic Grayscale Print Management Meta SOP Class	1.2.840.10008.5.1.1.9
- Basic Film Session SOP Class	1.2.840.10008.5.1.1.1
- Basic Film Box SOP Class	1.2.840.10008.5.1.1.2
- Basic Grayscale Image Box SOP Class	1.2.840.10008.5.1.1.4
- Printer SOP Class	1.2.840.10008.5.1.1.16
Print Job SOP Class	1.2.840.10008.5.1.1.14
Presentation LUT SOP Class	1.2.840.10008.5.1.1.23

SOP Class Name	SOP Class UID
Basic Color Print Management Meta SOP Class	1.2.840.10008.5.1.1.18
- Basic Film Session SOP Class	1.2.840.10008.5.1.1.1
- Basic Film Box SOP Class	1.2.840.10008.5.1.1.2
- Basic Color Image Box SOP Class	1.2.840.10008.5.1.1.4.1
- Printer SOP Class	1.2.840.10008.5.1.1.16
Print Job SOP Class	1.2.840.10008.5.1.1.14

#### 11.1.1 Association Establishment Policies

##### 11.1.1.1 General

Whenever a film is completely set up and printed by command or automatism, the job is prepared for processing. As soon as the queue is ready to process the job is activated and worked according the processing data. The related Print application will initiate an association to the print destination and process the printing of the related information.

The default PDU size used will be 28 KB.

**11.1.1.2 Number of Associations**

The *ARCADIS* DICOM Print application initiates one association at a time for each different print device configured.

**11.1.1.3 Asynchronous Nature**

The *ARCADIS* DICOM software does not support asynchronous communication (multiple outstanding transactions over a single association).

**11.1.1.4 Implementation Identifying Information**

Implementation Class UID	1.2.12.2.1107.5.12
Implementation Version Name	SIEMENS_ ASPVC10A

**11.1.2 Association Initiation Policy**

Triggered by the Print job queue the Print Management SCU establishes an association by using the DICOM association services. With the help of the N-GET request for the Printer SOP Class the Status is determined before printing.

With no problem encountered with the N-CREATE/N-SET Services for the related Basic Print SOP Classes the film sheet is set up for printing and the image(s) is (are) transferred to the printer device.

After the last film is printed from queue, the Print application will leave open the association for another 60 seconds. If a new film job is ready for printing within this time limit, the job will be immediately processed over the still open association. If there is no new job, the association is closed if the time-out elapsed. This is done to optimize automated printing.

During the "idle-time" (no open association to printer) the Print application will issue a cyclic camera status request (using N-GET of Printer SOP Class) every 5 minutes.

**11.1.2.1 Associated Real-World Activity**

**11.1.2.1.1 Associated Real-World Activity – Printing a Printer Job Queue Entry**

Whenever the user prepares a film-sheet, it is forwarded to the Printer Job queue. As soon as the associated Printer device is available the job is activated and association is set up.

The film sheet is internally processed, converted to a Standard/1-1 page and then the page image is sent. Status is controlled by awaiting any N-EVENT message all through the transfer until the last image or film-sheet is sent.

If the response from the remote application contains a status other than Success or Warning the association is aborted.

**11.1.2.1.2 Proposed Presentation Context (Presentation Context Table)**

The Siemens *ARCADIS* DICOM Print application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Basic Grayscale Print Management Meta SOP class	1.2.840.10008.5.1.1.9	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Basic Color Print Management Meta SOP class	1.2.840.10008.5.1.1.18	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Basic film session SOP class	1.2.840.10008.5.1.1.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Basic Film Box SOP class	1.2.840.10008.5.1.1.2	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Basic Grayscale Image Box SOP class	1.2.840.10008.5.1.1.4	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Basic Color Image Box SOP class	1.2.840.10008.5.1.1.4.1	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Printer SOP class	1.2.840.10008.5.1.1.16	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Print Job SOP class	1.2.840.10008.5.1.1.14	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None
Presentation LUT SOP class	1.2.840.10008.5.1.1.23	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None

**11.1.2.1.3 SOP specific Conformance Statement – Meta SOP Classes**

The *ARCADIS* DICOM print management SCU conforms to the DICOM Basic Grayscale Print Management Meta SOP Class and the Basic Color Print Management Meta SOP Class.

The application uses a setting platform to define the properties of the connected SCP, e.g.:

- maximum number of print jobs in the queue
- maximum number of print copies
- supported film sizes of the connected DICOM SCP
- supported film formats of the DICOM SCP
- lookup table definition.

The printing is only suspended in the case of a failure return status of the SCP.

**11.1.2.1.3.1 Basic Film Session SOP class**

The Basic Film Session information object definition describes all the user-defined parameters, which are common for all the films of a film session. The Basic Film Session refers to one or more Basic Film Boxes and that are printed on one hardcopy printer.

The ARCADIS DICOM print management SCU supports the following DIMSE Service operations for the Basic Film Session SOP Class as SCU:

- N-CREATE
- N-DELETE

Following attributes are used in the N-CREATE-RQ message:

Attribute Name	Tag	Usage SCU	Supported Values
Number of Copies	(2000,0010)	U	1
Medium Type	(2000,0030)	U	BLUE FILM CLEAR FILM PAPER
Film Destination	(2000,0040)	U	MAGAZINE PROCESSOR

The number of Copies sent to the DICOM Printer is always 1, the job is sent n times for n copies.

The affected SOP Instance UID received with the N-CREATE-RSP message will be kept internally and used for later requests (e.g. N-DELETE-RQ) on the Basic Film Session – see below:

Attribute Name	Tag	Source of Information
Requested SOP Instance UID	(0000,1000) → (0000,1001)	Affected SOP Instance UID of N-CREATE-RSP on Basic Film Session

The N-DELETE-RQ on the Basic Film Session SOP Class is used to remove the complete Basic Film Session SOP Instance hierarchy.

The Basic Film Session SOP class interprets the following status codes (from N-CREATE-RSP, N-DELETE-RSP messages):

Service Status	Meaning	Error Codes
Failed	Film session SOP instances hierarchy does not contain film box SOP instances	C600
	Unable to create print job, print queue is full	C601
	Image size is larger than images box size	C603
Warning	Memory allocation not supported	B600
	Film session printing is not supported	B601
	Film box does not contain image box (empty page)	B602
Success	Film belonging to the film session are accepted for printing	0000

**11.1.2.1.3.2 Basic Film Box SOP class**

The Basic Film Box information object definition describes all the user-defined parameter of one film of the film session. The Basic Film Box information description defines the presentation parameters, which are common for all images on a given sheet of film.

The Basic Film Box refers to one or more Image Boxes.

Supported Service Elements as SCU are:

- N-CREATE, N-ACTION and N-DELETE

The Basic Film Box SOP class N-CREATE-RQ message uses the following attributes (the actual values for each attribute depend on DICOM printer configuration within the *ARCADIS* DICOM print management SCU):

Attribute Name	Tag	Usage SCU	Supported Values
Image Display Format	(2010,0010)	M	STANDARD\1,1
Referenced Film Session Sequence	(2010,0500)	M	
> Referenced SOP Class UID	(0008,1150)	M	1.2.840.10008.5.1.1.1
> Referenced SOP Instance UID	(0008,1155)	M	
Film Orientation	(2010,0040)	M	PORTRAIT
Film Size ID	(2010,0050)	M	8INX10IN, 10INX12IN, 10INX14IN, 11INX14IN,, 14INX14IN, 14INX17IN, 24CMX24CM, 24CMX30CM
Magnification Type	(2010,0060)	M	BILINEAR, CUBIC, NONE, REPLICATE
Border Density	(2010,0100)	U	BLACK, WHITE
Max Density	(2010,0130)	U	0 < Value
Min Density	(2010,0120)	U	0 < Value < 50
Illumination	(2010,015E)	U	0 < Value Required if Presentation LUT is present.
Reflective Ambient Light	(2010,0160)	U	0 < Value Required if Presentation LUT is present.
Referenced Presentation LUT Sequence	(2050,0500)	U	

The N-CREATE-RSP message from the Print SCP includes the Referenced Image Box Sequence with SOP Class/Instance UID pairs, which will be kept internally to be further used for the subsequent Basic Image Box SOP Class N-SET-RQ messages.

When all Image Boxes (including parameters) for the film-sheet have been set, the *ARCADIS* DICOM print manager will issue an N-ACTION-RQ message with the SOP Instance UID of the Basic Film Box and the Action Type ID of 1.

The affected SOP Instance UID received with N-CREATE-RSP message will be kept internally and used for later requests (e.g. N-DELETE-RQ) on the Basic Film Box - see below:

Attribute Name	Tag	Source of Information
Requested SOP Instance UID	(0000,1000) → (0000,1001)	Affected SOP Instance UID of N-CREATE-RSP on Basic Film Box

The Basic Film Box SOP class interprets the following status codes:

Service Status	Meaning	Error Codes
Failure	Unable to create print job, print queue is full	C602
	Image size is larger than images box size	C603
Warning	Film box does not contain image box (empty page)	B603
	Requested MinDensity or MaxDensity outside of Printer's operating range	B605
Success	Film accepted for printing	0000

### 11.1.2.1.3.3 Basic Grayscale Image Box SOP Class

The Basic Grayscale Image Box information object definition is the presentation of an image and image related data in the image area of a film. The Basic Image Box information describes the presentation parameters and image pixel data, which apply to a single image of a sheet of film.

The Grayscale Image Box SOP Class uses only the N-SET-RQ with the following attributes:

Attribute Name	Tag	Usage SCU	Supported Values
Image Position	(2020,0010)	M	1
BASIC Grayscale Image Sequence	(2020,0110)	M	
> Samples per Pixel	(0028,0002)	M	1
> Photometric Interpretation	(0028,0004)	M	MONOCHROME2
> Rows	(0028,0010)	M	
> Columns	(0028,0011)	M	
> Pixel Aspect Ratio	(0028,0034)	M	
> Bits Allocated	(0028,0100)	M	8/16
> Bits Stored	(0028,0101)	M	8/12
> High Bit	(0028,0102)	M	7/11
> Pixel Representation	(0028,0103)	M	0
> Pixel Data	(7FE0,0010)	M	

The Grayscale Image Box SOP class interprets the following status codes:

Service Status	Meaning	Error Codes
Failure	Image contains more pixel than printer can print in Image Box	C603
	Insufficient memory in printer to store the image	C605
Warning	Requested MinDensity or MaxDensity outside of Printer's operating range	B605
Success		0000

### 11.1.2.1.3.4 Presentation LUT SOP Class

The objective of the Presentation LUT is to realize image hardcopy printing tailored for specific modalities, applications and user preferences.

The output of the Presentation LUT is Presentation Values (P-Values). P-Values are approximately related to human perceptual response. They are intended to facilitate common input for hardcopy. P-Values are intended to be independent of the specific class or characteristics of the hardcopy device.

The Presentation LUT SOP Class uses only the N-CREATE-RQ with the following attributes:

Attribute Name	Tag	Usage SCU	Supported Values
Presentation LUT Shape	(2050,0020)	U	IDENTITY

The affected SOP Instance UID received with N-CREATE-RSP message will be kept internally and is used for later requests on the Basic Film Box (N-CREATE-RQ) and on the Presentation LUT (N-DELETE-RQ) - see below:

Attribute Name	Tag	Source of Information
Requested SOP Instance UID	(0000,1000) → (0000,1001)	Affected SOP Instance UID of N-CREATE-RSP on Presentation LUT

The Presentation LUT SOP class interprets the following status codes:

Service Status	Meaning	Error Codes
Warning	Requested MinDensity or MaxDensity outside of HCD's operating range. HCD will use its respective minimum or maximum density value instead.	B605
Success	Presentation LUT successfully created	0000

### 11.1.2.1.3.5 Printer SOP Class

The Printer SOP Class provides the possibility to monitor the status of the hardcopy printer in a synchronous and an asynchronous way.

The SCU uses the mandatory N-EVENT Report DIMSE operation to monitor the changes of the printer status in an asynchronous way.

It can directly ask the Printer (SCP) for its status or receive events from the Printer asynchronously:

- N-GET as SCU
- N-EVENT-REPORT as SCU

In both cases the following information is supported:

#### *Used Printer N-EVENT Report attributes*

Event-type Name	Event	Attributes	Tag	Usage SCU
Normal	1			
Warning	2	Printer Status Info	(2110,0020)	U
Failure	3	Printer Status Info	(2110,0020)	U

#### *Mandatory Printer N-GET-RSP, N-EVENT-REPORT-RQ attributes*

Attribute Name	Tag	Usage SCP	Supported Values
Printer Status	(2110,0010)	M	NORMAL, FAILURE, WARNING
Printer Status Info	(2110,0020)	M	See tables in Annex for details.

**Note:** For a detailed description on how *ARCADIS* reacts on different printer status messages, please refer to the Annex section "DICOM Print SCU – detailed status displays".

### 11.1.2.1.3.6 Print Job SOP Class

The Print Job SOP Class is the possibility to monitor the execution of the print process.

The *ARCADIS* DICOM Print Management application supports the optional N-EVENT-REPORT DIMSE operation to receive the changes of the Print Job Status in an asynchronous way.

It can receive events from the Print SCP asynchronously.

Note: *ARCADIS* does not support receiving N-EVENT messages from camera during print sessions; normally this is configurable in the camera.





Used Print Job N-EVENT Report attributes

Event-type Name	Event	Attributes	Tag	Usage SCU
Normal	1	Execution Status Info	(2100,0030)	U
		Print Job ID	(2100,0010)	-- (Print Queue Management SOP Class not supported)
		Film Session Label	(2000,0050)	U
		Printer Name	(2110,0030)	U
Printing	2	Execution Status Info	(2100,0030)	U
		Print Job ID	(2100,0010)	-- (Print Queue Management SOP Class not supported)
		Film Session Label	(2000,0050)	U
		Printer Name	(2110,0030)	U
Done	3	Execution Status Info	(2100,0030)	U
		Print Job ID	(2100,0010)	-- (Print Queue Management SOP Class not supported)
		Film Session Label	(2000,0050)	U
		Printer Name	(2110,0030)	U
Failure	4	Execution Status Info	(2100,0030)	U
		Print Job ID	(2100,0010)	-- (Print Queue Management SOP Class not supported)
		Film Session Label	(2000,0050)	U
		Printer Name	(2110,0030)	U

**Note:** For a detailed description on how *ARCADIS* reacts on different printer status messages, please refer to the Annex section "DICOM Print SCU – detailed status displays".

### 11.1.3 Association Acceptance Policy

Not applicable

## 12 Implementation Model Worklist

The Basic Worklist Management Service class defines an application-level class of service, which facilitates the transfer of worklists from the information system to the imaging modality. The worklist is queried by the AE and supplies the SCU with the scheduled tasks, which have to be performed on the modality. The ARCADIS DICOM worklist application supports the worklist service as SCU.

### 12.1 Application Data Flow Diagram

The ARCADIS DICOM network implementation acts as SCU for the Basic Worklist Service using the Modality Worklist SOP Class. The product target Operating System is Windows XP.




Figure 6: ARCADIS Application Flow Diagram – Basic Worklist SCU

### 12.2 Functional Definitions of Application Entities

The worklist SCU ("broad query") is invoked from the patient browser user interface or by timer to request the worklist from a remote Information System (Modality Worklist Class SCP). This is done to perform a match to the internal worklist query keys specified in the C-Find DIMSE service issued for the Modality Worklist Model.

The worklist SCP responds to the C-FIND query and scheduled imaging service requests (scheduled procedure steps) and patient demographic information will be downloaded from the information system to the ARCADIS modality. All information retrieved will be hold in the scheduling database for usage during Patient registration procedure.

Furthermore the patient based Query dialog from the patient browser allows entering specific matching criteria ("narrow query") for the issue worklist query. With the response data the Patient Registration dialog can be populated according availability within the worklist response identifier.

### **12.3 Sequencing of Real-World Activities**

The “narrow” (interactive) Worklist Query requires that sufficient matching keys or a unique matching key are/is entered before the query is issued. Only then a single response can be expected to complete the registration dialog.

## 13 Application Entity Specification Worklist

### 13.1 Modality Worklist Service AE Specification

The Modality worklist SCU (patient registration in conjunction with the network application) requests that the remote SCP performs a match of all keys specified in the query against the information in its worklist database.

The ARCADIS DICOM network implementation acts as SCU for the Basic Worklist Service using the Modality Worklist SOP Class:

SOP Class Name	SOP Class UID
Modality Worklist Information Model - FIND	1.2.840.10008.5.1.4.31

#### 13.1.1 Association Establishment Policies

##### 13.1.1.1 General

It is possible to configure a cyclic update of the modality scheduler database through a background worklist request with date/time and modality information.

In addition the user can request worklist update with "Update Worklist". No duplicate entries will be added in the Scheduler DB. Entries are uniquely identified by the Study Instance UID (0020, 000D) for the Requested Procedure and the SPS ID (0040, 009) in the SPS Sequence (0040, 0100).

An interactive worklist query can be issued with search criteria entered in the patient based Query dialog from the patient browser.

The default PDU size used will be 28 KB.

##### 13.1.1.2 Number of Associations

The ARCADIS DICOM application initiates one association at a time to query worklist entry data.

##### 13.1.1.3 Asynchronous Nature

The ARCADIS DICOM software does not support asynchronous communication (multiple outstanding transactions over a single association).

##### 13.1.1.4 Implementation Identifying Information

Implementation Class UID	1.3.12.2.1107.5.12
Implementation Version Name	SIEMENS_ ASPVC10A

**13.1.2 Association Initiation Policy**

The network application will cyclically query the worklist and by request of patient registration interface. Ever then it establishes an association by using the DICOM association services.

The following DIMSE-C operation is supported as SCU:

- C-FIND

**13.1.2.1 Real-World Activity**

**13.1.2.1.1 Associated Real-World Activity - Query (Update) Worklist**

A network application will perform worklist queries with the C-FIND request at regular intervals. In addition it can be triggered by immediate request. The received worklist items will be compared with the contents of the local scheduler database. New items will be inserted into scheduler database.

After each broad-query, all RP/SPS that were canceled or rescheduled to another modality at the RIS will be automatically removed from the Scheduler DB if:

1. the Examination of this procedure has not been started or finished yet, and
2. the corresponding configuration item “Automatic removal of canceled/rescheduled Request” was checked in the Service UI under DICOM/HIS-RIS Node.

No automatic clean up of the scheduler DB is performed after a Patient base Query since the worklist received does not give the complete list of all currently scheduled procedures for the modality.

**13.1.2.1.2 Proposed Presentation Contexts**

The ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Modality Worklist Information Model- FIND	1.2.840.10008.5.1.4.31	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None

13.1.2.1.3 SOP Specific Conformance Statement

- Search Key Attributes of the Worklist C-FIND

The ARCADIS DICOM worklist SCU supports “broad worklist queries” with all required search keys. The following tables describe the “broad query” search keys that the SCU supports.

Attribute Name	Tag	Matching Key Type	Query Value
Scheduled Procedure Step			
Scheduled Procedure Step Sequence	(0040,0100)	R	
>Scheduled Station AE Title	(0040,0001)	R	<own AET> or “**” <sup>a</sup>
>Scheduled Procedure Step Start Date	(0040,0002)	R	<act. Date>-<act. Date> or range from UI <sup>b</sup>
>Scheduled Procedure Step Start Time	(0040,0003)	R	<zero length> or range from UI <sup>b</sup>
>Modality	(0008,0060)	R	“**” or <own Modality> <sup>a</sup>

- Return Key Attributes of the Worklist C-FIND

The ARCADIS DICOM worklist SCU supports worklist queries with return key attributes of all types. The following tables describe the return keys that the SCU supports.

An “x” in the **UI** column will indicate the attribute is visualized when browsing the Worklist results with Patient Browser and/or during Patient Registration. The Patient Browser display is additionally influenced by the related Browser configuration.

A tag in the **IOD** column will indicate that the related attribute is included into the SOP Instances of the IOD’s created during processing of this worklist request.

A tag in the **MPPS** column will indicate that the related attribute is included into the SOP Instances of the MPPS objects created during processing of this worklist request. (See also the tables “Attributes used for the Performed Procedure Step N-CREATE” and “Attributes used for the Performed Procedure Step N-SET”.)

Attribute Name	Tag	Return Key Type	UI	IOD	MPPS
<b>SOP Common</b>					
Specific Character Set	(0008,0005)	1C	-	(0008,0005)	(0008,0005)
<b>Scheduled Procedure Step</b>					
Scheduled Procedure Step Sequence	(0040,0100)	1			
>Modality	(0008,0060)	1	x	(0008,0060)	(0008,0060)
>Requested Contrast Agent	(0032,1070)	2C	x	(0032,1070)	
>Scheduled Station AE Title	(0040,0001)	1	x		(0040,0241) <sup>c</sup>
>Scheduled Procedure Step Start Date	(0040,0002)	1	x		
>Scheduled Procedure Step Start Time	(0040,0003)	1	x		
>Scheduled Procedure Step End Date	(0040,0004)	3	-		
>Scheduled Procedure Step End Time	(0040,0005)	3	-		
>Scheduled Performing Physician’s Name	(0040,0006)	1	x	(0008,1050) <sup>d</sup>	(0008,1050) <sup>d</sup>

<sup>a</sup> This depends on user configuration (Options->Configuration->Patient Registration) if the "own AET" is provided or not. Use the "HIS/RIS" tabcard for configuration.

<sup>b</sup> It depends on user configuration (Options->Configuration->Patient Registration) if the actual Date with a full time range or an interactive input dialog for date/time specification is used.

<sup>c</sup> “Scheduled Station AE Title” is taken as default for “Performed Station AE Title”

<sup>d</sup> “Scheduled Performing Physician’s Name” is taken as default for “Performing Physician’s Name”

Attribute Name	Tag	Return Key Type	UI	IOD	MPPS
>Scheduled Procedure Step Description	(0040,0007)	1C	x	(0040,0007) (0040,0254) <sup>a</sup>	(0040,0007) (0040,0254) <sup>e</sup>
>Scheduled Protocol Code Sequence <sup>b</sup>	(0040,0008)	1C	-	(0040,0008) (0040,0260) <sup>c</sup>	(0040,0008) (0040,0260) <sup>d</sup>
>>Code Value	(0008,0100)	1C	x		
>>Coding Scheme Designator	(0008,0102)	1C	x		
>>Coding Scheme Version	(0008,0103)	3	x		
>>Code Meaning	(0008,0104)	3	x		
>Scheduled Procedure Step ID	(0040,0009)	1	x	(0040,0009) (0040,0253) <sup>d</sup>	(0040,0009) (0040,0253) <sup>a</sup>
>Scheduled Station Name	(0040,0010)	2	x		
>Scheduled Procedure Step Location	(0040,0011)	2	x		(0040,0242) <sup>e</sup>
>Pre-Medication	(0040,0012)	2C	x		
>Scheduled Procedure Step Status	(0040,0020)	3	x		
>Comments on the Scheduled Procedure Step	(0040,0400)	3	-		
<b>Requested Procedure</b>					
Referenced Study Sequence <sup>f</sup>	(0008,1110)	2	-	(0008,1110)	(0008,1110)
>Referenced SOP Class UID	(0008,1150)	1C	-		
>Referenced SOP Instance UID	(0008,1155)	1C	-		
Study Instance UID	(0020,000D)	1	-	(0020,000D)	(0020,000D)
Requested Procedure Description	(0032,1060)	1C	x	(0032,1060)	(0032,1060)
Requested Procedure Code Sequence <sup>f</sup>	(0032,1064)	1C	-	(0008,1032) <sup>g</sup> (0032,1064)	(0008,1032) <sup>d</sup>
>Code Value	(0008,0100)	1C	x		
>Coding Scheme Designator	(0008,0102)	1C	x		
>Coding Scheme Version	(0008,0103)	3	x		
>Code Meaning	(0008,0104)	3	x		
Requested Procedure ID	(0040,1001)	1	x	(0040,1001) (0020,0010) <sup>h</sup>	(0040,1001) (0020,0010) <sup>e</sup>
Reason for the Requested Procedure	(0040,1002)	3	-		
Requested Procedure Priority	(0040,1003)	2	x		
Patient Transport Arrangements	(0040,1004)	2	-		
Requested Procedure Location	(0040,1005)	3	-		
Confidentiality Code	(0040,1008)	3	-		
Reporting Priority	(0040,1009)	3	-		
Names of intended Recipients of Results	(0040,1010)	3	-	(0008,1048)	
Requested Procedure Comments	(0040,1400)	3	x		
<b>Imaging Service Request</b>					
Accession Number	(0008,0050)	2	x	(0008,0050)	(0008,0050)
Referring Physician's Name	(0008,0090)	2	x	(0008,0090)	
Requesting Physician	(0032,1032)	2	x	(0032,1032)	(0032,1032)
Requesting Service	(0032,1033)	3	x	(0032,1033)	
Reason for the Imaging Service Request	(0040,2001)	3	-		
Issuing Date of Imaging Service Request	(0040,2004)	3	-		
Issuing Time of Imaging Service Request	(0040,2005)	3	-		

<sup>a</sup> "Scheduled Procedure Step Description" is taken as default for "Performed Procedure Step Description"

<sup>b</sup> Uses universal sequence match

<sup>c</sup> "Scheduled Protocol Code Sequence" is taken as default for "Performed Protocol Code Sequence"

<sup>d</sup> "Scheduled Procedure Step ID" is taken as default for "Performed Procedure Step ID"

<sup>e</sup> "Scheduled Procedure Step Location" is taken as default for "Performed Location"

<sup>f</sup> Uses universal sequence match

<sup>g</sup> "Requested Procedure Code Sequence" is taken as default for "Procedure Code Sequence"

<sup>h</sup> "Requested Procedure ID" is taken as default for "Study ID"



Attribute Name	Tag	Return Key Type	UI	IOD	MPPS
Placer Order Number / Imaging Service Request <sup>a</sup>	(0040,2016)	3	-		(0040,2016)
Filler Order Number / Imaging Service Request <sup>b</sup>	(0040,2017)	3	-		(0040,2017)
Order entered by ...	(0040,2008)	3	-		
Order Enterer's location	(0040,2009)	3	-		
Order Callback Phone Number	(0040,2010)	3	-		
Imaging Service Request Comments	(0040,2400)	3	x		
<b>Visit Identification</b>					
Institution Name	(0008,0080)	3	x	(0008,0080)	
Institution Address	(0008,0081)	3	-	(0008,0081)	
Institution Code Sequence <sup>c</sup>	(0008,0082)	3	-		
>Code Value	(0008,0100)	1C	-		
>Coding Scheme Designator	(0008,0102)	1C	-		
>Coding Scheme Version	(0008,0103)	3	-		
>Code Meaning	(0008,0104)	3	-		
Admission ID	(0038,0010)	2	x		
Issuer of Admission ID	(0038,0011)	3	-		
<b>Visit Status</b>					
Visit Status ID	(0038,0008)	3	-		
Current Patient Location	(0038,0300)	2	x		
Patient's Institution Residence	(0038,0400)	3	-		
Visit Comments	(0038,4000)	3	-		
<b>Visit Relationship</b>					
Referenced Study Sequence <sup>c</sup>	(0008,1110)	3	-		
>Referenced SOP Class UID	(0008,1150)	1C	-		
>Referenced SOP Instance UID	(0008,1155)	1C	-		
Referenced Patient Sequence <sup>c</sup>	(0008,1120)	2	-		(0008,1120)
>Referenced SOP Class UID	(0008,1150)	1C	-		
>Referenced SOP Instance UID	(0008,1155)	1C	-		
<b>Visit Admission</b>					
Referring Physician's Name	(0008,0090)	2	x	(0008,0090)	
Referring Physician's Address	(0008,0092)	3	-		
Referring Physician's Phone Numbers	(0008,0094)	3	-		
Admitting Diagnosis Description	(0008,1080)	3	x	(0008,1080)	
Admitting Diagnosis Code Sequence <sup>c</sup>	(0008,1084)	3	-		
>Code Value	(0008,0100)	1C	-		
>Coding Scheme Designator	(0008,0102)	1C	-		
>Coding Scheme Version	(0008,0103)	3	-		
>Code Meaning	(0008,0104)	3	-		
Route of Admissions	(0038,0016)	3	-		
Admitting Date	(0038,0020)	3	-		
Admitting Time	(0038,0021)	3	-		
<b>Patient Identification</b>					
Patient's Name	(0010,0010)	1	x	(0010,0010)	(0010,0010)
Patient ID	(0010,0020)	1	x	(0010,0020)	(0010,0020)
Issuer of Patient ID	(0010,0021)	3	-	(0010,0021)	
Other Patient Ids	(0010,1000)	3	x	(0010,1000)	
Other Patient Names	(0010,1001)	3	x	(0010,1001)	
Patient's Birth Name	(0010,1005)	3	-	(0010,1005)	
Patient's Mother's Birth Name	(0010,1060)	3	-	(0010,1060)	

<sup>a</sup> Old tag (0040,2006) is retired and not used.

<sup>b</sup> Old tag (0040,2007) is retired and not used.

<sup>c</sup> Uses universal sequence match

Attribute Name	Tag	Return Key Type	UI	IOD	MPPS
Medical Record Locator	(0010,1090)	3	-	(0010,1090)	
<b>Patient Demographic</b>					
Patient's Birth Date	(0010,0030)	2	x	(0010,0030)	(0010,0030)
Patient's Birth Time	(0010,0032)	3	-	(0010,0032)	
Patient's Sex	(0010,0040)	2	x	(0010,0040)	(0010,0040)
Patient's Insurance Plan Code Sequence <sup>c</sup>	(0010,0050)	3	-	(0010,0050)	
>Code Value	(0008,0100)	1C	-		
>Coding Scheme Designator	(0008,0102)	1C	-		
>Coding Scheme Version	(0008,0103)	3	-		
>Code Meaning	(0008,0104)	3	-		
Patient's Age	(0010,1010)	3	x	(0010,1010)	
Patient's Size	(0010,1020)	3	x	(0010,1020)	
Patient's Weight	(0010,1030)	2	x	(0010,1030)	
Patient's Address	(0010,1040)	3	x	(0010,1040)	
Military Rank	(0010,1080)	3	x	(0010,1080)	
Branch of Service	(0010,1081)	3	-	(0010,1081)	
Country of Residence	(0010,2150)	3	-	(0010,2150)	
Region of Residence	(0010,2152)	3	-	(0010,2152)	
Patient's Telephone Numbers	(0010,2154)	3	-	(0010,2154)	
Ethnic Group	(0010,2160)	3	x	(0010,2160)	
Occupation	(0010,2180)	3	-	(0010,2180)	
Patient's Religious Preference	(0010,21F0)	3	-	(0010,21F0)	
Patient Comments	(0010,4000)	3	x	(0010,4000)	
Patient Data Confidentiality Constraint Description	(0040,3001)	2	x	(0040,3001)	
<b>Patient Medical</b>					
Medical Alerts	(0010,2000)	2	x	(0010,2000)	
Contrast Allergies	(0010,2110)	2	x	(0010,2110)	
Pregnancy Status	(0010,21C0)	2	x	(0010,21C0)	
Smoking Status	(0010,21A0)	3	x	(0010,21A0)	
Last Menstrual Date	(0010,21D0)	3	x	(0010,21D0)	
Additional Patient History	(0010,21B0)	3	x	(0010,21B0)	
Special Needs	(0038,0050)	2	x	(0038,0050)	
Patient State	(0038,0500)	2	x	(0038,0500)	
<b>Patient Relationship</b>					
Referenced Study Sequence <sup>a</sup>	(0008,1110)	3	-		
>Referenced SOP Class UID	(0008,1150)	1C	-		
>Referenced SOP Instance UID	(0008,1155)	1C	-		
Referenced Visit Sequence <sup>a</sup>	(0008,1125)	3	-		
>Referenced SOP Class UID	(0008,1150)	1C	-		
>Referenced SOP Instance UID	(0008,1155)	1C	-		
Referenced Patient Alias Sequence <sup>a</sup>	(0038,0004)	3	-		
>Referenced SOP Class UID	(0008,1150)	1C	-		
>Referenced SOP Instance UID	(0008,1155)	1C	-		

### 13.1.2.1.4 Associated Real-World Activity – Get Worklist

With "Get Worklist" in the patient based Worklist Query dialog, the entered attributes are used to form a worklist request identifier. With the response data the Patient Registration dialog can be updated to perform examination in advance. The response data are additionally placed in the scheduler database

<sup>a</sup> Uses universal sequence match

**13.1.2.1.5 Proposed Presentation Contexts – Get Worklist**

This RWA will propose the same Presentation Contexts as with “Update Worklist”. Please see table in section 13.1.2.1.2.

**13.1.2.1.6 SOP Specific Conformance – Get Worklist**

**13.1.2.1.6.1.1.1 Search Key Attributes of the Worklist C-FIND**

The ARCADIS DICOM worklist SCU supports “narrow worklist queries” with all required search keys. The following tables describe the “narrow query” search keys that the SCU supports.

Attribute Name	Tag	Matching Key Type	Query Value
<b>Scheduled Procedure Step</b>			
Scheduled Procedure Step Sequence	(0040,0100)	R	
>Scheduled Performing Physician’s Name	(0040,0006)	R	input from UI or <zero length>
<b>Requested Procedure</b>			
Requested Procedure ID	(0040,1001)	O	input from UI or <zero length>
<b>Imaging Service Request</b>			
Accession Number	(0008,0050)	O	input from UI or <zero length>
Referring Physician’s Name	(0008,0090)	O	input from UI or <zero length>
<b>Visit Status</b>			
Current Patient Location	(0038,0300)	O	input from UI or <zero length>
<b>Patient Identification</b>			
Patient’s Name	(0010,0010)	R	input from UI or <zero length>
Patient ID	(0010,0020)	R	input from UI or <zero length>

**13.1.2.1.6.1.1.2 Return Key Attributes of the Worklist C-FIND**

Please see list for “Update Worklist” RWA.

**13.1.2.1.6.1.1.3 Status Codes of the Worklist C-FIND**

The worklist SCU interprets following status codes:

Service Status	Meaning	Error Codes	Related Fields
Refused	Out of Resources	A700	(0000,0902)
Failed	Identifier does not match SOP Class	A900	(0000,0901) (0000,0902)
	Unable to process	Cxxx	(0000,0901) (0000,0902)
Cancel	Matching terminated due to Cancel request	FE00	None
Success	Matching is complete - No final Identifier is supplied	0000	None
Pending	Matches are continuing - Current Match is supplied and any Optional Keys were supported in the same manner as Required Keys	FF00	Identifier
	Matches are continuing - Warning that one or more Optional Keys were not supported for existence and/or matching for this identifier	FF01	Identifier

## 14 Implementation Model MPPS

The Modality Performed Procedure Step Service class defines an application-level class of service, which facilitates the transfer of procedure, billing and radiation dose information from the imaging modality to the information system. The Performed Procedure Step is created and set by the AE and supplies the SCP with the information about a real-world procedure, which is performed, on the modality. The ARCADIS DICOM Modality Performed Procedure Step application supports the MPPS service as SCU.

### 14.1 Application Data Flow Diagram

The ARCADIS DICOM network implementation acts as SCU for the Modality Performed Procedure Step SOP Class. The product target Operating System is Windows XP.




Figure 7: ARCADIS Application Flow Diagram – MPPS SCU

### 14.2 Functional Definitions of Application Entities

With registering a Patient (i.e. a Scheduled Procedure Step from Worklist), the ARCADIS DICOM application will create a MPPS Instance and communicate it to the MPPS SCP.

Furthermore a manual update can be performed with the ARCADIS MPPS user interface. Only there it is possible to set the state of the MPPS to “Completed” or “Discontinued”. If done so, the DICOM application will no longer allow updates on the related MPPS Instance.

The ARCADIS will not only allow a "1:1 -relationship" of Scheduled Procedure Steps and Performed Procedure Steps, but also supports the "simple group-case" (grouping several SPS of the same Requested Procedure) , "complex group-case" (grouping several SPS from different Requested Procedures) and "append case" from the respective IHE-scenarios.

The ARCADIS will support creation of “unscheduled cases” by allowing MPPS Instances to be communicated for locally registered Patients.

## 15 AE Specification MPPS

### 15.1 Modality Performed Procedure Step AE Specification

The Modality Performed Procedure Step SCU (Patient Registration and MPPS UI) provides information about a performed real-world Procedure to a remote SCP (Information System).

SIEMENS ARCADIS DICOM products provide Standard Conformance to the following DICOM V3.0 SOP Class as an SCU:

SOP Class Name	SOP Class UID
Modality Performed Procedure Step	1.2.840.10008.3.1.2.3.3

#### 15.1.1.1 General

ARCADIS does the creation of MPPS Instance automatically whenever a patient is registered for image acquisition through the Patient Registration dialog.

Further updates on the MPPS data can be done interactively from the related MPPS user interface. The MPPS “Complete” or “Discontinued” states can only be set from user interface.

The default PDU size used will be 28 KB.

#### 15.1.1.2 Number of Associations

The ARCADIS DICOM application initiates one association at a time to create or set MPPS instance.

#### 15.1.1.3 Asynchronous Nature

The ARCADIS DICOM software does not support asynchronous communication (multiple outstanding transactions over a single association).

#### 15.1.1.4 Implementation Identifying Information

Implementation Class UID	1.3.12.2.1107.5.12
Implementation Version Name	SIEMENS_ ASPVC10A

### 15.1.2 Association Initiation Policy

The ARCADIS DICOM Application Entity acts as a Service Class User (SCU) for the Modality Performed Procedure Step Service Class (to notify a RIS about status of a procedure while it is performed).

To do so, the ARCADIS will issue a

- N-CREATE DIMSE according to the CREATE Modality Performed Procedure Step SOP Instance operation or a

- N-SET DIMSE to update the contents and state of the MPPS according to the SET Modality Performed Procedure Step Information operation.

**15.1.2.1 Real World Activity**

**15.1.2.1.1 Associated Real-World Activity - Patient registered**

A patient is registered by the Patient Registration “Exam” action. From this event the trigger to create a MPPS Instance is derived. The related Instance is then immediately communicated to the configured RIS system. An association is established and the MPPS Instance is sent.

**15.1.2.1.2 Proposed Presentation Contexts - Patient registered**

The ARCADIS DICOM application will propose Presentation Contexts as shown in the following table:

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Modality Performed Procedure Step	1.2.840.10008.3.1.2.3.3	Implicit VR Little Endian Explicit VR Little Endian Explicit VR Big Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1 1.2.840.10008.1.2.2	SCU	None

**15.1.2.1.3 SOP Specific Conformance Statement- Patient registered**

**15.1.2.1.3.1 Attributes used for the Performed Procedure Step N-CREATE**

The Siemens ARCADIS DICOM Modality Performed Procedure Step SCU informs the remote SCP when the examination of a scheduled procedure step will be performed (i.e. the patient is registered). The N-CREATE message is sent when the examination is started with successful registration of the patient data. The following table describes the supported attributes of an N-CREATE message.

Attribute Name	Tag	Type	Value
SOP Common			
Specific Character Set	(0008,0005)	1C	from MWL or created
Performed Procedure Step Relationship			
Scheduled Step Attribute Sequence	(0040,0270)	1	
>Study Instance UID	(0020,000D)	1	from MWL or created
>Referenced Study Sequence	(0008,1110)	2	from MWL or <zero length>
>>Referenced SOP Class UID	(0008,1150)	1C	
>>Referenced SOP Instance UID	(0008,1155)	1C	
>Accession Number	(0008,0050)	2	from MWL or user input
>Placer Order Number/Imaging Service Request	(0040,2016)	3	from MWL or <zero length>
>Filler Order Number/Imaging Service Request	(0040,2017)	3	from MWL or <zero length>
>Requested Procedure ID	(0040,1001)	2	from MWL or user input
>Requested Procedure Description	(0032,1060)	2	from MWL or <zero length>
>Scheduled Procedure Step ID	(0040,0009)	2	from MWL or <zero length>
>Scheduled Procedure Step Description	(0040,0007)	2	from MWL or <zero length>
>Scheduled Action Item Sequence	(0040,0008)	2	from MWL or <zero length>
>>Code Value	(0008,0100)	1C	

>>Coding Scheme Designator	(0008,0102)	1C	
>>Coding Scheme Version	(0008,0103)	3	
>>Code Meaning	(0008,0104)	3	
Patient's Name	(0010,0010)	2	from MWL or user input
Patient ID	(0010,0020)	2	from MWL or user input
Patient's Birth Date	(0010,0030)	2	from MWL or user input
Patient's Sex	(0010,0040)	2	from MWL or user input
Referenced Patient Sequence	(0008,1120)	2	from MWL or <zero length>
>Referenced SOP Class UID	(0008,1150)	1C	
>Referenced SOP Instance UID	(0008,1155)	1C	
Performed Procedure Step Information			
Performed Procedure Step ID	(0040,0253)	1	From SPS ID or created
Performed Station AE Title	(0040,0241)	1	own AE Title
Performed Station Name	(0040,0242)	2	own hostname
Performed Location	(0040,0243)	2	from SPS location or <zero length>
Performed Procedure Step Start Date	(0040,0244)	1	created
Performed Procedure Step Start Time	(0040,0245)	1	created
Performed Procedure Step Status	(0040,0252)	1	"IN PROGRESS"
Performed Procedure Step Description	(0040,0254)	2	from SPS Description or <zero length>
Performed Procedure Type Description	(0040,0255)	2	<zero length>
Procedure Code Sequence	(0008,1032)	2	from Requested Procedure Code or <zero length>
>Code Value	(0008,0100)	1C	
>Coding Scheme Designator	(0008,0102)	1C	
>Coding Scheme Version	(0008,0103)	3	
>Code Meaning	(0008,0104)	3	
Performed Procedure Step End Date	(0040,0250)	2	<zero length>
Performed Procedure Step End Time	(0040,0251)	2	<zero length>
Image Acquisition Results			
Modality	(0008,0060)	1	XA or CT
Study ID	(0020,0010)	2	from Requested Procedure ID or created
Performed Protocol Code Sequence	(0040,0260)	2	from Scheduled Action Item Sequence or <zero length>
>Code Value	(0008,0100)	1C	
>Coding Scheme Designator	(0008,0102)	1C	
>Coding Scheme Version	(0008,0103)	3	
>Code Meaning	(0008,0104)	3	
Performed Series Sequence	(0040,0340)	2	
>Performing Physician's Name	(0008,1050)	2C	from MWL or user input
>Operator's Name	(0008,1070)	2C	User input
>Series Instance UID	(0020,000E)	1C	created
>Series Description	(0008,103E)	2C	<zero length>
>Retrieve AE Title	(0008,0054)	2C	<zero length>
>Protocol Name	(0018,1030)	1C	
>Referenced Image Sequence	(0008,1140)	2C	<zero length>
>Referenced Standalone SOP Instance Sequence	(0040,0220)	2C	<zero length>

**15.1.2.1.3.2 Status Codes of the Performed Procedure Step N-CREATE**

The Performed Procedure Step SCU interprets following status codes:

Service Status	Meaning	Error Codes (0000.0900)
Failure	Processing Failure	0110
	No such attribute	0105
	Invalid attribute value	0106
	Duplicate SOP Instance	0111
	No such SOP Instance	0112
	No such SOP Class	0118
	Class Instance conflict	0119
	Missing attribute	0120
	Missing attribute value	0121
	Resource limitation	0213
Success	MPPS Instance created	0000

**15.1.2.1.4 Associated Real-World Activity – MPPS UI-Update**

With the MPPS UI the status of the MPPS Instance can be set to “COMPLETED” or “DISCONTINUED”. There is no cyclic update during performance of the procedure.

**15.1.2.1.5 Proposed Presentation Contexts – MPPS UI-Update**

This RWA will propose the same Presentation Contexts as with “Patient registered”. Please see table in section 15.1.2.1.2.

**15.1.2.1.6 SOP Specific Conformance Statement – MPPS UI-Update**

**15.1.2.1.6.1 Attributes used for the Performed Procedure Step N-SET**

The Siemens ARCADIS DICOM Modality Performed Procedure Step SCU informs the remote SCP about the performed examination and its status. The N-SET message is sent only per ended examination (finished status “COMPLETED” or incomplete status “DISCONTINUED”). The following table describes the supported attributes of an N-SET message.

Attribute Name	Tag	Type	Value
Performed Procedure Step Information			
Performed Procedure Step Status	(0040,0252)	3	“COMPLETED” or “DISCONTINUED”
Performed Procedure Step Description	(0040,0254)	3	from SPS Description or user input
Performed Procedure Type Description	(0040,0255)	3	User input
Procedure Code Sequence	(0008,1032)	3	from Requested Procedure Code
>Code Value	(0008,0100)	1C	
>Coding Scheme Designator	(0008,0102)	1C	
>Coding Scheme Version	(0008,0103)	3	
>Code Meaning	(0008,0104)	3	
Performed Procedure Step End Date	(0040,0250)	1	created
Performed Procedure Step End Time	(0040,0251)	1	created
Image Acquisition Results			
Performed Protocol Code Sequence	(0040,0260)	3	from Scheduled Action Item Sequence or user input
>Code Value	(0008,0100)	1C	



>Coding Scheme Designator	(0008,0102)	1C	
>Coding Scheme Version	(0008,0103)	3	
>Code Meaning	(0008,0104)	3	
Performed Series Sequence	(0040,0340)	1	
>Performing Physician's Name	(0008,1050)	2C	from MWL or user input
>Protocol Name	(0018,1030)	1C	
>Operator's Name	(0008,1070)	2C	user input
>Series Instance UID	(0020,000E)	1C	from related SOP Instance
>Series Description	(0008,103E)	2C	from related SOP Instance
>Retrieve AE Title	(0008,0054)	2C	from Storage Commitment response or <zero length>
>Referenced Image Sequence	(0008,1140)	2C	<zero length>
>>Referenced SOP Class UID	(0008,1150)	1C	
>>Referenced SOP Instance UID	(0008,1155)	1C	
>Referenced Standalone SOP Instance Sequence	(0040,0220)	2C	<zero length>
Image Area Dose Product	(0018,115E)	3	
Total Time of Fluoroscopy	(0040,0300)	3	
Total Number of Exposures	(0040,0301)	3	
Entrance Dose in mGy	(0040,8302)	3	
Comments on Radiation Dose	(0040,0310)	3	user input
Billing Procedure Step Sequence	(0040,0320)	3	user input
Film Consumption Sequence	(0040,0321)	3	
>Medium Type	(2000,0030)	3	user input
>Film Size ID	(2010,0050)	3	user input
>Number of Films	(2100,0170)	3	user input
Billing Supplies and Devices Sequence	(0040,0324)	3	
>Quantity Sequence	(0040,0293)	3	
>>Quantity	(0040,0294)	3	user input
>>Measuring Units Sequence	(0040,0295)	3	user input
>Billing Item Sequence	(0040,0296)	3	user input

15.1.2.1.6.2 Status Codes of the Performed Procedure Step N-SET

The Performed Procedure Step SCU interprets following status codes:

Service Status	Meaning	Error Codes (0000.0900)
Failure	Processing Failure: Performed Procedure Step Object may no longer be updated.	0110
	No such attribute	0105
	Invalid attribute value	0106
	No such SOP Instance	0112
	Invalid Object instance	0117
	No such SOP Class	0118
	Class Instance conflict	0119
	Missing attribute value	0121
	Resource limitation	0213
Success	MPPS Instance set	0000

## 16 Communication Profiles

### 16.1 Supported Communication Stacks

The Siemens ARCADIS DICOM application provides DICOM V3.0 TCP/IP Network Communication Support as defined in Part 8 of the DICOM Standard. The product target Operating System is Windows XP.

#### 16.1.1 TCP/IP Stack

The ARCADIS DICOM application uses the TCP/IP stack from the target operating system upon which it executes. It uses the MergeCOM-3 subroutine library from Merge Technologies Inc. that is based on a Berkeley socket interface.

##### 16.1.1.1 API

The ARCADIS DICOM application uses the MergeCOM library that is based on a TCP/IP socket interface.

##### 16.1.1.2 Physical Media Support

The ARCADIS DICOM application is indifferent to the physical medium over which TCP/IP executes; it inherits this from the target operating system upon which it executes.

## 17 Extensions / Specializations / Privatizations

### 17.1.1 Standard Extended / Specialized / Private SOPs

Please refer to annex for further information on these topics. A detailed overview is given there.

### 17.1.2 Private Transfer Syntaxes

Not applicable

## 18 Configuration

### 18.1 AE Title/Presentation Address Mapping

To ensure unique identification within the network the hostname should be used as part of the AE Titles (see examples below, hostname = name1). The string can be up to 16 characters long and must not contain any extended characters, only 7-bit ASCII characters (excluding Control Characters) are allowed according to DICOM Standard.

**Note:** the current implementation of syngo does not support the full DICOM Standard. Spaces and special characters (like &<> ") in the AE title string are not supported.

#### 18.1.1 DICOM Verification

The Verification Service uses the AE configuration of the DICOM Service that is checked with the C-ECHO message. E.g. Verification will use the Storage AE, if initiated to check the configuration of a remote DICOM node.

#### 18.1.2 DICOM AE Title

Within ARCADIS there are local application entity titles for HIS/RIS, Study Transfer - which are used for Storage and Query/Retrieve - and Print. They can be configured via Service-UI in Configuration / DICOM / General (e.g. STU\_NAME1).

The port number is set to the fixed value of 104.

### 18.2 Configurable Parameters

The Application Entity Titles, host names and port numbers for remote AE are configured using the ARCADIS Service/Installation Tool. For each AET the list of services supported can be configured.

#### 18.2.1 Storage, Storage Commitment and Query/Retrieve

The ARCADIS Service/Installation Tool can be used to set the AET's, port-numbers, host-names, IP-addresses and capabilities for the remote nodes (SCP's). The user can select transfer syntaxes, compression modes and query models for each SCP separately.

- A quality factor, which determines the proposed transfer syntax in case that a user has initiated the C-STORE. By convention, 0 means: Only Uncompressed Transfer Syntax(es) are proposed, 100 means: Lossless Transfer Syntax is proposed, and any other value between 1 and 99 means that JPEG Lossy Transfer Syntax is proposed. One Uncompressed Transfer Syntax will be proposed in any case. This parameter is general for all destination nodes.

- a “compression type supported” which determines the proposed transfer syntax in case that the C-STORE was initiated as a sub-operation of an incoming C-MOVE-RQ. By convention, 0 means: Only Uncompressed Transfer Syntax(es) are proposed, 1 means: Lossless Transfer Syntax is proposed, and 2 means that a JPEG Lossy Transfer Syntax is proposed. One uncompressed transfer syntax will be proposed in any case. This parameter can be set for each configured destination node.

**Additional configurable parameters for Storage Commitment are:**

**When acting as SCU:**

- flag to indicate whether the association will be kept open to receive the response or to close the association and be prepared to receive the response on another association.
- time-out, which defines how long the association of N-ACTION, is kept to receive an N-EVENT-REPORT on the same association. The same value is used to wait for an N-EVENT-REPORT on another association. (default 1 h)

**When acting as SCP:**

- flag to indicate if an archive system is installed

### 18.2.2 Print

The ARCADIS Service/Installation Tool can be used to configure the SCP (DICOM-Printer).

These parameters are mandatory to set:

- AET,
- host-name,
- IP-address and
- Port-number.

These parameters have defaults as per configuration file and can be changed:

- default camera (yes/no),
- pixel size,
- additional or changed film sheet formats (e.g. inch 14x14, inch 14x17, ...),
- list with mapping pixel size to each film sheet format,
- minimal density,
- stored printed film jobs,
- media type,
- film destination.

### 18.2.3 Modality Worklist

The Service application can be used to set the AETs, port numbers, host names, IP addresses, capabilities and time-outs for the remote nodes (SCPs)

**Additional configurable parameters for Modality Worklist Query are:**

- Query Waiting time - the time to wait for the C-FIND-RSP after sending the C-FIND-RQ (default 20 sec.)

- Max Query Match Number - the maximum number of entries accepted in one worklist (default is 200)
- Query Interval: the time between two C-FIND-RQs to the Hospital Information system (default is 60 min.)
- Broad Worklist Query behavior (two values are defined):
  - Set the AE Title search attribute to the own AE Title, and the Modality search attribute to “\*”.
  - Set the Modality search attribute to the own modality and the AE Title search attribute to “\*”.

### 18.3 Default Parameters

This installation tool also uses some default parameters:

- max PDU size set to 262144 Bytes (256 kB)
- time-out for accepting/rejecting an association request: 60 s
- time-out for responding to an association open/close request: 60 s
- time-out for accepting a message over network: 60 s
- time-out for waiting for data between TCP/IP-packets: 60 s

The time-outs for waiting for a Service Request/Response message from the remote node are as follows:

- for Storage SCP/SCU: 600 s
- for Storage Commitment SCU:  
time-out for Response to N-ACTION: 600 s
- for Query/Retrieve SCP/SCU: 600 s
- for Print Management SCU:
  - time-out for Response to N-SET-RQ: 240 s
  - time-out for Response to other Requests: 60 s

## 19 Support of Extended Character Sets

The ARCADIS DICOM application supports the ISO 8859 Latin 1 (ISO-IR 100) character set.

Also the Japanese language character set JIS X 0201 – here ISO-IR 13 Japanese katakana is supported.

When there is a mismatch between the SCS tags (0008, 0005) and the characters in an IOD received by the system, then the following measures are taken to make the characters DICOM con-form:

- Try to import with ISO\_IR 100. If ISO\_IR 100 fails, convert each illegal character to a '?’.

## Media Storage Conformance Statement

This chapter will contain the Conformance Statement to all “Offline Media Application Profiles (incl. private extensions)” supported by the ARCADIS archive options.

Those application profiles supported shall be:

- Standard Application Profiles
- Augmented Application Profiles

# 1 Introduction

## 1.1 Purpose

This DICOM Conformance Statement is written according to part PS 3.2 of [1].

The applications described in this conformance statement are the SIEMENS ARCADIS based on *syngo*® software<sup>a</sup>. The ARCADIS DICOM offline media storage service implementation acts as FSC, FSU and/or FSR for the specified application profiles and the related SOP Class instances.

## 1.2 Scope

This DICOM Conformance Statement refers to SIEMENS ARCADIS products. The following table relates ARCADIS software names to SIEMENS products:

Software Name	SIEMENS <i>syngo</i> -based Product
VB13	ARCADIS

## 1.3 Definitions, Abbreviations

### 1.3.1 Definitions

DICOM            Digital Imaging and Communications in Medicine  
 DIMSE            DICOM Message Service Element  
 DIMSE-C        DICOM Message Service Element with Composite information objects

### 1.3.2 Abbreviations

ACR                American College of Radiology  
 AE                DICOM Application Entity  
 ASCII            American Standard Code for Information Interchange  
 DB                Database  
 DCS                DICOM Conformance Statement  
 FSC                File Set Creator  
 FSR                File Set Reader  
 FSU                File Set Updater  
 IOD                DICOM Information Object Definition  
 ISO                International Standard Organization  
 LEONARDO        AX-Workstation (for Angiographic/Radiographic viewing)  
 MOD                Magneto-optical Disk  
 NEMA              National Electrical Manufacturers Association  
 O                  Optional Key Attribute  
 PDU                DICOM Protocol Data Unit  
 R                  Required Key Attribute  
 RWA                Real-World Activity  
 U                  Unique Key Attribute

## 1.4 References

[1]                Digital Imaging and Communications in Medicine (DICOM), NEMA PS 3.1-3.18, 2004

<sup>a</sup> *syngo* is a registered trademark of Siemens AG.

## 1.5 Remarks

DICOM, by itself, does not guarantee interoperability. However, the Conformance Statement facilitates a first-level validation for interoperability between different applications supporting the same DICOM functionality as SCU and SCP, respectively.

This Conformance Statement is not intended to replace validation with other DICOM equipment to ensure proper exchange of information intended.

The scope of this Conformance Statement is to facilitate communication with Siemens and other vendors' Medical equipment. The Conformance Statement should be read and understood in conjunction with the DICOM 3.0 Standard [DICOM]. However, by itself it is not guaranteed to ensure the desired interoperability and a successful interconnectivity.

The user should be aware of the following important issues:

- The comparison of different conformance statements is the first step towards assessing interconnectivity between Siemens and non-Siemens equipment.
- Test procedures should be defined and tests should be performed by the user to validate the connectivity desired. DICOM itself and the conformance parts do not specify this.
- The standard will evolve to meet the users' future requirements. Siemens is actively involved in developing the standard further and therefore reserves the right to make changes to its products or to discontinue its delivery.
- Siemens reserves the right to modify the design and specifications contained herein without prior notice. Please contact your local Siemens representative for the most recent product information.



## 2 Implementation Model

### 2.1 Application Data Flow Diagram




Figure 8: ARCADIS Application Data Flow Diagram – Archive

The DICOM archive application will serve as an interface to the CD-R offline medium device. It serves interfaces to include the offline media directory into the browser and to copy SOP instances to a medium or retrieve SOP Instances from medium into local storage.

The DICOM Archive application will support the 120mm CD-R medium

The FSU role will update new SOP Instances only to media with pre-existing File-sets conforming to the Application Profiles supported.

The contents of the DICOMDIR will be temporarily stored in Archive-Database.

### 2.2 Functional Definitions of AEs

The ARCADIS DICOM offline media storage application consists of the DICOM Archive application entity serving all interfaces to access offline media. The DICOM Archive application is capable of

1. creating a new File-set onto an unwritten medium (Export to...).
2. updating an existing File-set by writing new SOP Instances onto the medium (Export to...).
3. importing SOP Instances from the medium onto local storage
4. reading the File-sets DICOMDIR information into temporary database and pass it to display applications.

### 2.3 Sequencing of Real-World Activities

The DICOM Archive application will not perform updates before the Directory information of the DICOMDIR is completely read.

When performing updates, the SOP instances are checked for existence before updating. Duplicate instances will be avoided.

**2.4 File Meta Information Options**

Implementation Class UID	1.3.12.2.1107.5.12
Implementation Version Name	SIEMENS_ ASPVC10A

**3 AE Specifications**

**3.1 DICOM Archive Specification**

The DICOM Archive provides Standard conformance to Media Storage Service Class (Interchange Option). In addition augmented conformance is provided to store extra data attributes important for the full feature support of the ARCADIS product SW. Details are listed in following Table:

Application Profiles Supported	Real-World Activity	Role	SC Option
AUG-STD-GEN-CD AUG-STD-XA1K-CD AUG-STD-GEN-DVD AUG-STD-XA1K-DVD	Browse Directory Information	FSR	Interchange
	Import into local Storage	FSR	Interchange
	Export to local Archive Media	FSC, FSU	Interchange
STD-GEN-CD STD-GEN-DVD STD-CTMR-CD STD-XA1K-CD STD-XA1K-DVD  STD-US-zz-yF-xxxxxx *2	Browse Directory Information	FSR	Interchange
	Import into local Storage	FSR	Interchange

\*2 - All combinations of the following values for zz, yF and xxxxxx are supported: yF={SF|MF}, zz={ID|SC|CC},

On syngo-based products the augmented profiles will be preferably used by the system. The General Purpose Interchange Profile (STD-GEN-CD), Ultrasound Profile (STD-US-xxx), CT and MR Image Profile (STD-CTMR-xxx) and 1024 X-Ray Angiographic Profile (STD-XA1K-CD) will be supported with read capability of the related media.

**3.1.1 File Meta Information for the Application Entity**

The Source Application Entity Title is set by configuration. See Chapter "Configuration" for details.

**3.1.2 Real-World Activities for this Application Entity**

**3.1.2.1 Real-World Activity: Browse Directory Information**

The DICOM Archive application acts as FSR using the interchange option when requested to read the media directory.

The DICOM archive application will read the DICOMDIR and insert those directory entries that are valid for the application profiles supported, into a local database. The database can then be used for browsing media contents.

- Note:

IconImageSQ is also supported in DICOMDIR. But only those Icon Images with BitsAllocated (0028, 0100) equal to 8 and size of 64x64 or 128x128 pixels are imported into database and are visible in the Browser.

### 3.1.2.1.1 Application Profiles for the RWA: Browse Directory Information

See Table in section 3.1 for the Application Profiles listed that invoke this Application Entity for the Browse Directory Information RWA.

### 3.1.2.2 Real-World Activity: Import into local Storage

The DICOM Archive application acts as FSR using the interchange option when requested to read SOP Instances from the medium into the local storage.

The SOP Instance selected from the media directory will be copied into the local storage. Only SOP Instances, that are valid for the application profile supported and are listed as supported by the Storage SCP Conformance section (Network DCS, 5.1.3), can be retrieved from media storage. This is due to the fact that the Browse Directory Information will filter all SOP Instances not matching the Application profiles supported.

For media conforming to the STD-GEN-CD Profile the following SOP classes will be supported as an FSR:

Information Object Definition	SOP Class UID	Transfer Syntax UID
CT image	1.2.840.10008.5.1.4.1.1.2	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
MR Image	1.2.840.10008.5.1.4.1.1.4	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
NM Image	1.2.840.10008.5.1.4.1.1.20	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
Secondary Capture Image	1.2.840.10008.5.1.4.1.1.7	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
Ultrasound Image (retired)	1.2.840.10008.5.1.4.1.1.6	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
Ultrasound Image	1.2.840.10008.5.1.4.1.1.6.1	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
Ultrasound Multi-frame Image (retired)	1.2.840.10008.5.1.4.1.1.3	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
Ultrasound Multi-frame Image	1.2.840.10008.5.1.4.1.1.3.1	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1
<b>X-Ray Angiographic Image</b>	<b>1.2.840.10008.5.1.4.1.1.12.1</b>	<b>Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1</b>
X-Ray Radiofluoroscopic Image	1.2.840.10008.5.1.4.1.1.12.2	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1

### 3.1.2.2.1 Application Profiles for the RWA: Import into local Storage

See Table in section 3.1 for the Application Profiles listed that invoke this Application Entity for the Import into Local Storage RWA.

### 3.1.2.3 Real-World Activity: Export to local Archive Media

The DICOM Archive application acts as FSU (for media with existing DICOM file-set) or FSC (media not initialized) using the interchange option when requested to copy SOP Instances from the local storage to local Archive Medium.

The DICOM Archive application will receive a list of SOP Instances to be copied to the local archive medium. According to the state of the medium inserted (new medium, Medium with DICOM file-set) the validity of the SOP Instances according to the applicable profile is checked. Only valid SOP Instances are accepted.

When the DICOM archive application is requested to copy SOP Instances the preferred application profile-according configuration will be used to validate and copy the referred SOP Instances. When creating a new file-set no Descriptor File will be allocated and the related ID is not used.

The DICOM archive application will not close the medium.

#### 3.1.2.3.1 Application Profiles for the RWA: Export to local Archive Media

See Table in section 3.1 for the Application Profiles listed that invoke this Application Entity for the Export to local Archive Media RWA.

## 4 Augmented and Private Profiles

### 4.1 Augmented Application Profiles

### 4.2 AUG-STD-GEN-CD and STD-XA1K-CD

#### 4.2.1 Transfer Syntax Augmentation

Additionally to the standard profiles following transfer syntaxes are supported:

SOP Class UID	Transfer Syntax UID	FSC	FSR	FSU
1.2.840.10008.5.1.4.1.1.2	Explicit VR Little Endian Uncompressed 1.2.840.10008.1.2.1	M	M	O
1.2.840.10008.5.1.4.1.1.2	JPEG Lossless Process 14 (selection value 1) 1.2.840.10008.1.2.4.70	O	M	O
1.2.840.10008.5.1.4.1.1.2	Explicit VR Big Endian Uncompressed 1.2.840.10008.1.2.2	O	M	O
1.2.840.10008.5.1.4.1.1.2	JPEG Lossy (baseline or extended) 1.2.840.10008.1.2.4.50 1.2.840.10008.1.2.4.51	O	O	O
1.2.840.10008.5.1.4.1.1.2	RLE Lossless 1.2.840.10008.1.2.5	O	O	O

**FSC, FSR, FSU – denote the requirements for those roles**

O – Optional

M - Mandatory

#### 4.2.2 Directory Augmentation

Conforming Application Entities shall include in the DICOMDIR File the Basic Directory IOD containing Directory Records at the Patient and subsidiary levels appropriate to the SOP Classes in the File-set. Directory Records shall reference all DICOM files in the File-set incorporating SOP Instances defined for the specific Application profile.

##### 4.2.2.1 Basic Directory IOD Specialization

This Application Profile makes use of optional attributes of the Basic Directory IOD to support recognition of Patient’s Storage Service request results in spanning multiple volumes (file sets). Therefore the File Set Descriptor File can be used and is then referenced by optional Basic Directory IOD attributes. If existent, the specified Descriptor File may be used by FSR applications. Any FSU, FSC shall make a clear Statement if the Descriptor File mechanism is used according to the specialization defined in this Application Profile.

The Descriptor Files shall have the following contents:

One single Line without any control-characters and according to the Basic Character-Set having the following defined text:

“MULTIVOLUME: xx of yy”

xx, yy are replaced by the actual Number of the volume (xx) and the Total Number of Volumes in the set (yy).

If used, the Descriptor File shall have the File ID “README” and reside in same directory level as the DICOMDIR. It is referenced by the attribute [0004, 1141] File-set Descriptor File ID having the defined content of “README”.

**4.2.2.2 Additional Keys**

File-set Creators and Updaters are required to generate the mandatory elements specified in PS 3.3, Annex F of the DICOM Standard. Table below: Additional DICOMDIR Keys specifies the additional associated keys. At each directory record level other additional data elements can be added, but it is not required that File Set Readers be able to use them as keys. Refer to the Basic Directory IOD in PS 3.3.

Key Attribute	Tag	Directory Record Level	Type	Notes
Date of Birth	(0010,0030)	PATIENT	2C	required, if present in SOP Instance
Patient's Sex	(0010,0040)	PATIENT	2C	required, if present in SOP Instance
Series Date	(0008,0021)	SERIES	3	
Series Time	(0008,0031)	SERIES	3	
Institute Name	(0008,0080)	SERIES	2C	required, if present in SOP Instance
Institution Address	(0008,0081)	SERIES	2C	required, if present in SOP Instance
Series Description	(0008,103E)	SERIES	3	
Performing Physician's Name	(0008,1050)	SERIES	2C	required, if present in SOP Instance
Image Type	(0008,0008)	IMAGE	1C	required, if present in SOP Instance
SOP Class UID	(0008,0016)	IMAGE	3	
SOP Instance UID	(0008,0018)	IMAGE	3	
Content Date	(0008,0023)	IMAGE	3	
Content Time	(0008,0033)	IMAGE	3	
Referenced Image Sequence	(0008,1140)	IMAGE	1C	required, if present in SOP Instance
> Referenced SOP Class UID	(0008,1150)			
> Referenced SOP Instance UID	(0008,1155)			
Image Position (Patient)	(0020,0032)	IMAGE	2C	required, if present in SOP Instance
Image Orientation (Patient)	(0020,0037)	IMAGE	2C	required, if present in SOP Instance
Frame of Reference UID	(0020,0052)	IMAGE	2C	required, if present in SOP Instance
Rows	(0028,0010)	IMAGE	3	
Columns	(0028,0011)	IMAGE	3	
Pixel Spacing	(0028,0030)	IMAGE	1C	required, if present in SOP Instance
Calibration Image	(0050,0004)	IMAGE	2C	required, if present in SOP Instance
Icon Image Sequence	(0088,0200)	IMAGE	3	required for Image SOP Classes
> Samples per Pixel	(0028,0002)			1
> Photometric Interpretation	(0028,0004)			MONOCHROME2
> Rows	(0028,0010)			64 , 128
> Columns	(0028,0011)			64 , 128
> Bits Allocated	(0028,0100)			8
> Bits Stored	(0028,0101)			8
> High Bit	(0028,0102)			7
> Pixel Representation	(0028,0103)			0
> Pixel Data	(7FE0,0010)			Icon Image
Curve Number	(0020,0024)	CURVE	1C	required, if present in SOP Instance

### **4.2.2.3 Icon Images**

Directory Records of type SERIES or IMAGE may include Icon Images. The Icon Image pixel data shall be as specified in PS 3.3 "Icon Image Key Definition", and restricted such, that Bits Allocated (0028,0100) and Bits Stored (0028,0101) shall be equal 8, and Rows (0028,0010) and Columns (0028,0011) shall be equal to 128 for XA Images and 64 for all other Images. The Photometric Interpretation (0028, 0004) shall always be restricted to "MONOCHROME2".

## 5 Configuration

### 5.1 AE Title Mapping

#### 5.1.1 DICOM Media Storage AE Title

The DICOM Storage application provides the application entity title:

CsalmageManager

## 6 Support of Extended Character Sets

The Siemens ARCADIS DICOM archive application supports the ISO 8859 Latin 1 (ISO-IR 100) character set.

Also the Japanese language character sets JIS X 0201 (ISO-IR 13 Japanese katakana) is supported.

When there is a mismatch between the SCS tags (0008, 0005) and the characters in an IOD received by the system, then the following measures are taken to make the characters DICOM conform:

- Try to import with ISO\_IR 100. If ISO\_IR 100 fails, convert each illegal character to a '?'.



**A ANNEX****A.1 ARCADIS Settings of Extended XA IOD**

Attribute Name	Tag	Notes
Specific Character Set	(0008,0005)	from configuration
Image Type	(0008,0008)	See A.4.1
SOP Class UID	(0008,0016)	1.2.840.10008.5.1.4.1.1.12.1
SOP Instance UID	(0008,0018)	Generated
Study Date	(0008,0020)	<yyyymmdd>
Series Date	(0008,0021)	<yyyymmdd>
Acquisition Date	(0008,0022)	<yyyymmdd>
Content Date	(0008,0023)	<yyyymmdd>
Study Time	(0008,0030)	<hhmmss.frac>
Series Time	(0008,0031)	<hhmmss.frac>
Acquisition Time	(0008,0032)	<hhmmss.frac>
Content Time	(0008,0033)	<hhmmss.frac>
Accession Number	(0008,0050)	RIS or Input or null
Modality	(0008,0060)	XA
Manufacturer	(0008,0070)	SIEMENS
Institution Name	(0008,0080)	from configuration
Referring Physician's Name	(0008,0090)	RIS or Input or null
Station Name	(0008,1010)	from configuration
Study Description	(0008,1030)	RIS requested Procedure Description or Input or not available
Series Description	(0008,103E)	Operating Mode
Performing Physician's Name	(0008,1050)	Input
Operator's Name	(0008,1070)	Input
Manufacturer's Model Name	(0008,1090)	ARCADIS
Referenced Image Sequence	(0008,1140)	Reference to what modify
> Referenced SOP Class UID	(0008,1150)	Referenced UID
> Referenced SOP Instance UID	(0008,1155)	Referenced UID
Patient's Name	(0010,0010)	RIS or Input
Patient ID	(0010,0020)	RIS or Input
Patient's Birth Date	(0010,0030)	RIS or Input
Patient's Sex	(0010,0040)	RIS or Input
Patient's Age	(0010,1010)	Calculated
Patient Comments	(0010,4000)	Input
KVP	(0018,0060)	Current value
Device Serial Number	(0018,1000)	Service parameter
Software Version	(0018,1020)	SIEMENS ASPIAVE31G * VC10A *
Protocol Name	(0018,1030)	Application field – body region – exam set
Frame Time	(0018,1063)	for multiframe
Field of View Shape	(0018,1147)	ROUND
Field of View Dimension	(0018,1149)	Current value
Exposure Time	(0018,1150)	Current value
X-Ray Tube Current	(0018,1151)	(mA) see also (0018,8151)
Average Pulse Width	(0018,1154)	Current value
Radiation Setting	(0018,1155)	Current value
Radiation Mode	(0018,115A)	CONTINUOUS   PULSED
Preferred Playback Sequencing	(0018,1244)	Current value
Recommended Display Frame Rate	(0018,2144)	Current value
Positioner Motion	(0018,1500)	STATIC
Shutter Shape	(0018,1600)	CIRCULAR

Attribute Name	Tag	Notes
Center of Circular Shutter	(0018,1610)	<row number>,<column number>
Radius of Circular Shutter	(0018,1612)	<number of pixel in row direction>
X-Ray Tube Current in $\mu$ A	(0018,8151)	$\mu$ A
Study Instance UID	(0020,000D)	From RIS   Generated
Series Instance UID	(0020,000E)	Generated UID
Study ID	(0020,0010)	Generated UID
Series Number	(0020,0011)	Generated UID
Image Number	(0020,0013)	Generated number
Patient Orientation	(0020,0020)	no values
Series in Study	(0020,1000)	Current value
Samples per Pixel	(0028,0002)	1
Photometric Interpretation	(0028,0004)	MONOCHROME2
Number of Frames	(0028,0008)	Current value
Frame Increment Pointer	(0028,0009)	(0018,1063)
Rows	(0028,0010)	1024
Columns	(0028,0011)	1024
Bits Allocated	(0028,0100)	16
Bits Stored	(0028,0101)	12
High Bit	(0028,0102)	11
Pixel Representation	(0028,0103)	0
Pixel Intensity Relationship	(0028,1040)	LIN   LOG if subtraction
Window Center	(0028,1050)	Multiple values
Window Width	(0028,1051)	Multiple values
Modality LUT Sequence	(0028,3000)	If subtraction
>> LUT Descriptor	(0028,3002)	4096/0/16
>> LUT Explanation	(0028,3003)	explanation
>> Modality LUT Type	(0028,3004)	US
>> LUT Data	(0028,3006)	LUT data
VOI LUT Sequence	(0028,3010)	Up to 4 LUTS possible
> Window Center	(0028,1050)	Corresponding linear value
> Window Width	(0028,1051)	Corresponding linear value
> VOI LUT Sequence	(0028,3010)	Up to 4
>> LUT Descriptor	(0028,3002)	4096/0/16 or 4096/0/12 (configurable)
>> LUT Explanation	(0028,3003)	Explanation
>> LUT Data	(0028,3006)	LUT Data
Recommended Viewing Mode	(0028,1090)	SUB, if subtraction
Mask Subtraction Sequence	(0028,6100)	If subtraction
>> Mask Operation	(0028,6101)	AVG_SUB
>> Applicable Frame Range	(0028,6102)	1\Number of Frames -1
>> Mask Frame Numbers	(0028,6110)	Current value
>> Mask Sub-pixel Shift	(0028,6114)	0.0\0.0
Private Creator	(0029,00xx)	CARDIO-D.R.1.0
Attributes according to A.3.5		
Performed Procedure Step Start Date	(0040,0244)	Supplied, even if MPPS SOP class is not supported
Performed Procedure Step Start Time	(0040,0245)	Supplied, even if MPPS SOP class is not supported
Performed Procedure Step End Date	(0040,0250)	Current value
Performed Procedure Step EndTime	(0040,0251)	Current value
Performed Procedure Step ID	(0040,0253)	Supplied, even if MPPS SOP class is not supported
Icon Image Sequence	(0088,0200)	Image Stamp

## A.2 ARCADIS Settings of Extended CT IOD

Attribute Name	Tag	Notes
Specific Character Set	(0008,0005)	from configuration
Image Type	(0008,0008)	See A.4.1
SOP Class UID	(0008,0016)	1.2.840.10008.5.1.4.1.1.2
SOP Instance UID	(0008,0018)	UID
Study Date	(0008,0020)	<yyyymmdd>
Series Date	(0008,0021)	<yyyymmdd>
Acquisition Date	(0008,0022)	<yyyymmdd>
Content Date	(0008,0023)	<yyyymmdd>
Study Time	(0008,0030)	<hhmmss.frac>
Series Time	(0008,0031)	<hhmmss.frac>
Acquisition Time	(0008,0032)	<hhmmss.frac>
Content Time	(0008,0033)	<hhmmss.frac>
Accession Number	(0008,0050)	RIS or Input or null
Modality	(0008,0060)	CT
Manufacturer	(0008,0070)	SIEMENS
Institute Name	(0008,0080)	from configuration
Referring Physician's Name	(0008,0090)	RIS or Input or null
Station Name	(0008,1010)	from configuration
Study Description	(0008,1030)	RIS requested Procedure Description or Input or not available
Series Description	(0008,103E)	Operating Mode
Performing Physician's Name	(0008,1050)	Input
Operator's Name	(0008,1070)	Input
Manufacturer's Model Name	(0008,1090)	ARCADIS
Referenced Image Sequence	(0008,1140)	Reference to what modify
> Referenced SOP Class UID	(0008,1150)	Referenced UID
> Referenced SOP Instance UID	(0008,1155)	Referenced UID
Patient's Name	(0010,0010)	RIS or Input
Patient ID	(0010,0020)	RIS or Input
Patient's Birth Date	(0010,0030)	RIS or Input
Patient's Sex	(0010,0040)	RIS or Input
Patient's Age	(0010,1010)	Calculated
Patient Comments	(0010,4000)	Input
Slice Thickness	(0018,0050)	Current value
KVP	(0018,0060)	Current value
Device Serial Number	(0018,1000)	From Service
Software Version	(0018,1020)	SIEMENS ASPIA VE31G * VC10A *
Study Instance UID	(0020,000D)	Generated UID
Series Instance UID	(0020,000E)	Generated UID
Study ID	(0020,0010)	Current value
Series Number	(0020,0011)	Current value
Acquisition Number	(0020,0012)	Current value
Instance Number	(0020,0013)	Current value
Image Position Patient	(0020,0032)	Current value
Image Orientation Patient	(0020,0037)	Current value
Frame of Reference UID	(0020,0052)	Generated UID
Series in Study	(0020,1000)	Current value
Position Reference Indicator	(0020,1040)	n.a
Samples per Pixel	(0028,0002)	1
Photometric Interpretation	(0028,0004)	MONOCHROME2
Number of Frames	(0028,0008)	1
Frame Increment Pointer	(0028,0009)	(0018,1063)

Attribute Name	Tag	Notes
Rows	(0028,0010)	256
Columns	(0028,0011)	256
Pixel Spacing	(0028,0030)	Current value
Bits Allocated	(0028,0100)	1024
Bits Stored	(0028,0101)	12
High Bit	(0028,0102)	11
Pixel Representation	(0028,0103)	0
Window Center	(0028,0050)	Current value
Window Width	(0028,0051)	Current value
Rescale Intercept	(0028,1052)	0
Rescale Slope	(0028,1053)	1
Private Creator	(0029,00xx)	ISOC 3D NAVIGATIONMATRIX.R. 1.0
Attributes according to A.3.6		
Performed Procedure Step Start Date	(0040,0244)	Supplied, even if MPPS SOP class is not supported
Performed Procedure Step Start Time	(0040,0245)	Supplied, even if MPPS SOP class is not supported
Performed Procedure Step End Date	(0040,0250)	Current value
Performed Procedure Step EndTime	(0040,0251)	Current value
Performed Procedure Step ID	(0040,0253)	Supplied, even if MPPS SOP class is not supported
Icon Image Sequence	(0088,0200)	Stamp

### A.3 Siemens Standard Extended Modules

IE	Module	Reference	Usage	Note
Image	CSA Image Header	A.3.1	U	private GG information
	CSA Series Header	A.3.2	U	
	MEDCOM Header	A.3.3	U	private <i>syngo</i> information
	MEDCOM OOG	A.3.4	U	if object graphics is attached to image
	Edge Enhancement	A.3.5	U	private Filter information
	3D Scan	A.3.6	U	3D information
	Acquisition Parameter	A.3.7	U	Private GG acquisition information

#### A.3.1 CSA Image Header Module

The table in this section contains private IOD Attributes that describe the CSA Image Header:

Attribute Name	Tag	Owner	Type	Notes
CSA Image Header Type	(0029,xx08)	SIEMENS CSA HEADER	1	CSA Image Header identification characteristics.

				Defined Terms: NUM 4 = NUMARIS/4 SOM 5 = SOMARIS/5
CSA Image Header Version	(0029,xx09)	SIEMENS CSA HEADER	3	Version of CSA Image Header Info (0029, xx10) format.
CSA Image Header Info	(0029,xx10)	SIEMENS CSA HEADER	3	Manufacturer model dependent information.

**A.3.2 CSA Series Header Module**

The table in this section contains private IOD Attributes that describe the CSA Series Header:

Attribute Name	Tag	Owner	Type	Notes
CSA Series Header Type	(0029,xx18)	SIEMENS CSA HEADER	1	CSA Series Header identification characteristics. Defined Terms: NUM 4 = NUMARIS/4
CSA Series Header Version	(0029,xx19)	SIEMENS CSA HEADER	3	Version of CSA Series Header Info (0029, xx20) format.
CSA Series Header Info	(0029,xx20)	SIEMENS CSA HEADER	3	Manufacturer model dependent information.

**A.3.3 MEDCOM Header Module**

The table in this section contains private IOD Attributes that describe MEDCOM Header:

Attribute Name	Tag	Owner	Type	Notes
MedCom Header Type	(0029,xx08)	SIEMENS MEDCOM HEADER	1C	MedCom Header identification characteristics. Defined Terms: MEDCOM 1 (Required if MedCom Header Info (0029, xx10) present.)
MedCom Header Version	(0029,xx09)	SIEMENS MEDCOM HEADER	2C	Version of MedCom Header Info (0029, xx10) format. (Required if MEDCOM Header Info (0029, xx10) present.)
MedCom Header Info	(0029,xx10)	SIEMENS MEDCOM HEADER	3	Manufacturer model dependent information. The value of the attribute MedCom Header Info (0029, xx10) can be build up in each user-defined format.
MedCom History Information	(0029,xx20)	SIEMENS MEDCOM HEADER	3	MedCom defined Patient Registration history information. See A.2.3.1.
PMTF Information 1	(0029,xx31)	SIEMENS MEDCOM HEADER	3	Transformation Information
PMTF Information 2	(0029,xx32)	SIEMENS MEDCOM HEADER	3	Transformation Information
PMTF Information 3	(0029,xx33)	SIEMENS MEDCOM HEADER	3	Transformation Information
PMTF Information 4	(0029,xx34)	SIEMENS MEDCOM HEADER	3	Transformation Information
PMTF Information 5	(0029,xx35)	SIEMENS MEDCOM HEADER	3	Transformation Information
Application Header Sequence	(0029,xx40)	SIEMENS MEDCOM HEADER	3	Sequence of Application Header items. Zero or more items are possible.
>Application Header Type	(0029,xx41)	SIEMENS MEDCOM HEADER	1C	Application Header identification characteristics. Required, if Sequence is sent.
>Application Header ID	(0029,xx42)	SIEMENS MEDCOM HEADER	3	Identification of an application header
>Application Header Version	(0029,xx43)	SIEMENS MEDCOM HEADER	3	Version of CSA Series Header Info (0029, xx44) format.
>Application Header Info	(0029,xx44)	SIEMENS MEDCOM	3	Application dependent information.

		HEADER		
Workflow Control Flags	(0029,xx50)	SIEMENS MEDCOM HEADER	3	Eight free definable flags.
Archive Management Flag Keep Online	(0029,xx51)	SIEMENS MEDCOM HEADER	3	Flag to control remote archive management system to keep the image always online (also when already archived). Enumerated Values: 00 = remote control not required 01 = keep image online
Archive Management Flag Do Not Archive	(0029,xx52)	SIEMENS MEDCOM HEADER	3	Flag to control remote archive management system not to archive the related image. Enumerated Values: 00 = remote control not required 01 = don't archive image
Image Location Status	(0029,xx53)	SIEMENS MEDCOM HEADER	3	Image location status to control retrieving. Defined Terms: ONLINE = retrieving has to be done as usual, NEARLINE = move request to SCP and delay according to value of Estimated Retrieve Time (0029,xx54), OFFLINE = invoking a retrieve operation initiates an operator request, INVALID = invoking a retrieve operation would always result in an error.
Estimated Retrieve Time	(0029,xx54)	SIEMENS MEDCOM HEADER	3	Estimated retrieve time in seconds. A value less than zero (< 0) indicates location is OFFLINE or INVALID.
Data Size of Retrieved Images	(0029,xx55)	SIEMENS MEDCOM HEADER	3	Data size of images in MByte.

**A.3.3.1 MEDCOM History Information**

The value of the attribute MEDCOM History Information (0029, xx20) is defined in the following way:

Part	Name	Type	Bytes	Notes
Header	Identifier	string	32	Always "CSA HISTORY"
	Version	string	32	e.g. "V1.10"
n Items	Class Name	string	64	
	Modification String	string	1024	

**A.3.4 MEDCOM OOG Module**

The table in this section contains private IOD Attributes that describe MEDCOM Object Oriented Graphics (OOG). This module is used whenever object graphics is drawn on the image and need to be stored as graphic object properties. Given the condition that the module contents was not removed by other modalities, the graphic objects remain re-animatable if such an image was transferred and is then retrieved back

Attribute Name	Tag	Owner	Type	Notes
MedCom OOG Type	(0029,xx08)	SIEMENS MEDCOM OOG	1	MEDCOM Object Oriented Graphics (OOG) identification characteristics. Defined Terms: MEDCOM OOG 1 MEDCOM OOG 2
MedCom OOG Version	(0029,xx09)	SIEMENS MEDCOM OOG	3	Version of MEDCOM OOG Info (0029, xx10) format.
MedCom OOG Info	(0029,xx10)	SIEMENS MEDCOM OOG	3	MEDCOM Object Oriented Graphics (OOG) data.

The graphics objects are also fully drawn in the Image Overlay Plane for compatibility with other products, which do not support the MedCom OOG module. Any system not supporting the MedCom OOG module shall remove the OOG module and its contents when modifying the image overlay plane content.

**A.3.5 Edge Enhancement Module private Viewing**

The table in this section contains private IOD Attributes that describe additional Attributes for advanced Viewing features modify

Attribute Name	Tag	Owner	VR	Type	Notes
Edge filter sequence	(0029,xx00)	CARDIO-D.R. 1.0	SQ	3	Standard formula according Dynaview Extensions (one or more items possible)
Size of the convolution kernel	(0029,xx01)	CARDIO-D.R. 1.0	US	1C	x-/y-size value pair Each value shall be greater or equal to 3.
Convolution kernel matrix	(0029,xx02)	CARDIO-D.R. 1.0	US	1C	Row-by-row list of the kernel Coefficients.
Edge enhancement gain factor	(0029,xx03)	CARDIO-D.R. 1.0	FL	1C	Applied Filter gain Factor. Range is from 0 to 100 Percent.

**A.3.6 3D Navigation Private Attributes**

The table in this section shows the private attributes that have been introduced to support 3D navigation systems. This set of attributes is appended to CT type images.

Attribute Name	Tag	Owner	VR	Type	Notes
Filename	(0029,xx00)	ISOC 3D NAVIGATIONMATRIX.R. 1.0	LO	1	Transformation Matrix File
Isoc ID	(0029,xx01)	ISOC 3D NAVIGATIONMATRIX.R. 1.0	LO	1	C-Arm Identifier
Scan ID	(0029,xx02)	ISOC 3D NAVIGATIONMATRIX.R. 1.0		1	Scan Identifier
Calibration ID	(0029,xx03)	ISOC 3D NAVIGATIONMATRIX.R. 1.0	LO	1	Calibration Identifier
Transformation Matrix	(0029,xx04)	ISOC 3D NAVIGATIONMATRIX.R. 1.0	DS	1	A homogenous 4x4 transformation matrix
Second Filename	(0029,xx05)	ISOC 3D NAVIGATIONMATRIX.R. 1.0	LO	1	Full Pathname of the Transformation Matrix File
Projections Requested	(0029,xx06)	ISOC 3D NAVIGATIONMATRIX.R. 1.0	US	1	Number of Requested Projections
Projections Done	(0029,xx07)	ISOC 3D NAVIGATIONMATRIX.R. 1.0	US	1	Number of Actually Performed Projections

### A.3.7 Private Acquisition Parameter

The next table in this section shows the private attributes that have been introduced to support image quality control.

Attribute Name	DICOM Tag	Owner	VR	Description
Characteristic Curve	0021,xx01	SIEMENS SP ACQ 2.0	LO	Name of the characteristic curve
Noise Reduction	0021,xx02	SIEMENS SP ACQ 2.0	LO	Name of the noise reduction mode
Pulse Rate	0021,xx03	SIEMENS SP ACQ 2.0	FD	Pulse rate in pulses/sec. Only set if the operating mode was PFC or DCM.
Dose Level	0021,xx04	SIEMENS SP ACQ 2.0	CS	Dose level specifier (LOW, MEDIUM or HIGH)
Storage Rate	0021,xx05	SIEMENS SP ACQ 2.0	FD	Storage rate in frames/sec. Only set for multi-frame images.
Examination Set	0021,xx10	SIEMENS SP ACQ 2.0	LO	Name of the examination set used for the acquisition. Surgery: application field/body region/examination set.
Operating Mode	0021,xx11	SIEMENS SP ACQ 2.0	LO	Name of the operating mode used for acquisition
Bone Display	0021,xx12	SIEMENS SP ACQ 2.0	CS	BLACK or WHITE
Camera Gain	0021,xx13	SIEMENS SP ACQ 2.0	US	Camera gain mode (1 or 2)
Nominal Brightness	0021,xx14	SIEMENS SP ACQ 2.0	US	Nominal brightness value used by the host software
Nominal Brightness Image Chain	0021,xx15	SIEMENS SP ACQ 2.0	US	Nominal brightness value used by the ASPIA image chain component
AGA Gamma Min	0021,xx16	SIEMENS SP ACQ 2.0	US	Minimum gamma value used by the ASPIA image chain for gamma correction (AGA).
AGA Lambda	0021,xx17	SIEMENS SP ACQ 2.0	US	Lambda value (gradient of the gamma curve) used by the ASPIA image chain for gamma correction (AGA).
AGA Offset	0021,xx18	SIEMENS SP ACQ 2.0	US	Offset value of the gamma curve used by the ASPIA image chain for gamma correction (AGA).
Peak Opacification Mode	0021,xx20	SIEMENS SP ACQ 2.0	CS	Peak Op mode (MIN or MAX). Only set if the operating mode was Subtraction or Roadmap
Landmark Factor	0021,xx21	SIEMENS SP ACQ 2.0	US	Landmark factor in percent. Only set if the operating mode was Subtraction or Roadmap
KV Stop	0021,xx30	SIEMENS SP ACQ 2.0	CS	Specifies whether the KV Stop mode was active on



				the host (ON) or not (OFF)
Power Mode	0021,xx31	SIEMENS SP ACQ 2.0	CS	Specifies whether the Power mode was active on the host (ON) or not (OFF)
Image Intensifier Zoom Level	0021,xx32	SIEMENS SP ACQ 2.0	US	Zoom level of the image intensifier zoom (0 = no zoom)
Power Reduction Level	0021,xx33	SIEMENS SP ACQ 2.0	US	Specifies whether a special power reduction curve has been applied during acquisition (0 = not applied, 1 = curve 1, 2 = curve 2)
Dose Reduction	0021,xx34	SIEMENS SP ACQ 2.0	CS	Specifies whether the dose reduction mode was active during acquisition (ON/OFF)
Automatic Iris Control	0021,xx35	SIEMENS SP ACQ 2.0	CS	Specifies whether the automatic iris control was active during acquisition (ON/OFF)
Metering Section Name	0021,xx36	SIEMENS SP ACQ 2.0	LO	Specifies the name of the metering section that was active during acquisition.
Viewing Mode ETC	0021,xx37	SIEMENS SP ACQ 2.0	CS	<u>Denotes the viewing mode of the ETC (e.g. Simple/ Advanced) activated at acquisition time.</u>
Facility Type	0021,xx40	SIEMENS SP ACQ 2.0	US	0 Unknown 1 Varic 2 Orbic 3 Orbic Cube 10 Carvic (size unknown) 11 Carvic (9 inch) 12 Carvic (13 inch)
Vignetting	0021,xx41	SIEMENS SP ACQ 2.0	FD	Vignetting of the image in percent.
Focal Distance	0021,xx42	SIEMENS SP ACQ 2.0	US	Distance between detector plane and focal spot in mm.
Default native SUB LUT	0021,xx50	SIEMENS SP ACQ 2.0	US	Specifies which of the center/width values stated in DICOM attribute (0021,xx51) and (0021,xx52) should be used by default for native image representation of subtraction images. The first element in the mentioned DICOM attributes is indicated by position value 1.
Window Center Native	0021,xx51	SIEMENS SP ACQ 2.0	DS	Specifies the center values of the center/width value pairs representing the native

				SUB LUTs approximation.
Window Width Native	0021,xx52	SIEMENS SP ACQ 2.0	DS	Specifies the width values of the center/width value pairs representing the native SUB LUTs approximation.
Vessel Image Frame Position	0021,xx55	SIEMENS SP ACQ 2.0	UL	Specifies the frame position of the vessel image frame within a SUB multiframe.
Vessel Image Frame Transparency	0021,xx56	SIEMENS SP ACQ 2.0	FL	Specifies the transparency level that should be applied to the vessel image during image blending.
Center Windowing for Catheter Image	0021,xx57	SIEMENS SP ACQ 2.0	UL	Specifies the center value of the C/W value pair used to approximate a LUT for the windowing of the catheter image.
Width Windowing for Catheter Image	0021,xx58	SIEMENS SP ACQ 2.0	UL	Specifies the width value of the C/W value pair used to approximate a LUT for the windowing of the catheter image.
k-factor LIH	0021,xx59	SIEMENS SP ACQ 2.0	US	The number of integration images used for the noise reduction of a LIH.
PeakOp Frame Position	0021,xx60	SIEMENS SP ACQ 2.0	UL	Specifies the position of the PeakOp Frame within the Multiframe.
Show Vessel RoadAB	0021,xx61	SIEMENS SP ACQ 2.0	US	Specifies whether the vessel frame was displayed over the PeakOp of phase B. 0 – was not displayed 1 – was displayed
Additional attributes for a 3D scan				
Direction Forward	0029,xx01	SIEMENS SP 3D	US	Scan direction (0 = forwards, 1 = backwards)
Scan Protocol	0029,xx02	SIEMENS SP 3D	US	Requested number of images to be taken by the scan
Angulation	0029,xx03	SIEMENS SP 3D	FD	Angulation in degrees
C-Arm Left	0029,xx04	SIEMENS SP 3D	US	Horizontal C-arm position (0 or 1)
Projection Matrix File	0029,xx05	SIEMENS SP 3D	LO	Full path name to the file containing the projection matrices.

BV Mask	0029,xx06	SIEMENS SP 3D	US	BV mask used
Looking Direction	0029,xx07	SIEMENS SP 3D	US	Looking direction

**A.4 Standard Extensions of all SOP Classes**

The following tables list the data dictionary of all DICOM IOD attributes where the DICOM standard definitions are extended:

Attribute Name	Tag	Private Creator	Type	Notes
Image Type	(0008,0008)	-	1	see A.4.1 additional Defined Terms:  Defined Terms for value 3: OTHER  Defined Terms for value 4: CSA 3D EDITOR CSA 3D FLY PATH CSA 3D FLY VRT CSA 3D FUSION CSA AVERAGE CSA BLACK IMAGE CSA RESAMPLED CSA MIP CSA MPR CSA MPR CURVED CSA MPR THICK CSA SSD CSA SUBTRACT CT_SOM4 * ECAT ACF ECAT NORMAL ECAT 3D SINO ECAT 3D SINO FLT SHS *
Patient Position	(0018,5100)	-	2C	see A.4.2 additional Defined Terms for the Magnetom Open: HLS HLP FLS FLP HLDL HLDR FLDL FLDR

All SOP classes may contain additional type 3 attributes which DICOM standard defines in a different DICOM IOD or DICOM SOP class (attributes from Normalized SOP classes).

This is the case for example for

- Rescale Slope (0028,1053)
- Rescale Intercept (0028,1052)

which are also used in the MR IOD.

**A.4.1 Image Type**

The Image Type (0008, 0008) attribute identifies important image identification characteristics. These characteristics are:

1. Pixel Data Characteristics:

- is the image an ORIGINAL Image; an image whose pixel values are based on original or source data, or

- is the image a DERIVED Image; an image whose pixel values have been derived in some manner from the pixel value of one or more other images.
  2. Patient Examination Characteristics:
- is the image a PRIMARY Image; an image created as a direct result of the Patient examination, or
- is the image a SECONDARY Image; an image created after the initial Patient examination.
  3. Modality Specific Characteristics (SOP Specific Characteristics).
  4. Implementation specific identifiers; other implementation specific identifiers shall be documented in an implementation's conformance claim.

The Image Type attribute is multi-valued and shall be provided in the following manner:

- Value 1 shall identify the Pixel Data Characteristics; Enumerated Values for the Pixel Data Characteristics are:
  - ORIGINAL = identifies an Original Image
  - DERIVED = identifies a Derived Image
- Value 2 shall identify the Patient Examination Characteristics; Enumerated Values for the Patient Examination Characteristics are:
  - PRIMARY = identifies a Primary Image
  - SECONDARY = identifies a Secondary Image
- Value 3 shall identify any Image IOD specific specialization, the following terms are defined in addition to the DICOM standard definitions:
  - OTHER = is also used for converted non-Axial and non-Localizer CT images
  - MPR = for 3D MPR images
  - PROJECTION IMAGE = for 3D MIP and SSD images
- Value 4 which are implementation specific, the following terms are defined in addition to the DICOM standard definitions:
  - original *syngo* generated data set types:
    - CSA 3D EDITOR = object created by 3D Editor
    - CSA 3D FLY PATH = object created by Fly Through Path
    - CSA 3D FLY VRT = object created by Fly Through Volume Rendering Technique
    - CSA 3D FUSION = object created by Fusion
    - CSA AVERAGE = image was created by Average
    - CSA BLACK IMAGE = SC Image with black pixels, only graphics information is of interest
    - CSA RESAMPLED = derived image created by zooming or panning original image
    - CSA REPORT = *syngo* reporting (documentation of diagnosis)
    - CSA RESULT = *syngo* reporting (postprocessing results)
    - CSA MIP = image created by Maximum Intensity Projection
    - CSA MIP THIN = image created by Maximum Intensity Projection
    - CSA MPR = image created by Multi Planar Reconstruction
    - CSA MPR CURVED = image created by Multi Planar Reconstruction
    - CSA MPR THICK = image created by Multi Planar Reconstruction
    - CSA MPR THIN = image created by Multi Planar Reconstruction
    - CSA SSD = SC Image as Shaded Surface Display
    - CSA SUBTRACT = image was created by Subtraction
    - ECAT ACF = CTI PET Attenuation Correction
    - ECAT NORMAL = CTI PET Normalization
    - ECAT 3D SINO = CTI PET 3D Sinogram Short
    - ECAT 3D SINO FLT = CTI PET 3D Sinogram Float

- Converted images
  - CT\_SOM4 NONE = converted SOMARIS image
  - CT\_SOM4 CONV = converted SOMARIS Convolution Kernel file
  - CT\_SOM4 DART = converted SOMARIS Dental Artificial image
  - CT\_SOM4 DEVA = converted SOMARIS Dental Evaluation image
  - CT\_SOM4 DGRA = converted SOMARIS Dental Graphics image
  - CT\_SOM4 DMEA = converted SOMARIS Dynamic Measurement image
  - CT\_SOM4 DPAN = converted SOMARIS Dental Panorama image
  - CT\_SOM4 DPAR = converted SOMARIS Dental Paraxial image
  - CT\_SOM4 EBT = converted SOMARIS Evolution image
  - CT\_SOM4 HIS = converted SOMARIS Histogram Graphics image
  - CT\_SOM4 HISC = converted SOMARIS Histogram Graphics image
  - CT\_SOM4 MUL = converted SOMARIS Multiscan image
  - CT\_SOM4 OEVA = converted SOMARIS Osteo Evaluation image
  - CT\_SOM4 OTOM = converted SOMARIS Osteo Tomogram image
  - CT\_SOM4 OTOP = converted SOMARIS Osteo Topogram image
  - CT\_SOM4 PLOT = converted SOMARIS Plot image
  - CT\_SOM4 QUAL = converted SOMARIS Quality image
  - CT\_SOM4 R2D = converted SOMARIS 2D Rebuild image
  - CT\_SOM4 R3D = converted SOMARIS 3D Rebuild image
  - CT\_SOM4 R3DE = converted SOMARIS 3D Rebuild image
  - CT\_SOM4 RMAX = converted SOMARIS Maximum Intensity Projection image
  - CT\_SOM4 RMIN = converted SOMARIS Minimum Intensity Projection image
  - CT\_SOM4 ROT = converted SOMARIS Rotation Mode image
  - CT\_SOM4 RRAD = converted SOMARIS Radiographic Projection image
  - CT\_SOM4 RVIT = converted SOMARIS Vessel Image Tool image
  - CT\_SOM4 RVRT = converted SOMARIS Volumetric Rendering image
  - CT\_SOM4 SAVE = converted SOMARIS Evolution Screen Save image
  - CT\_SOM4 SCAN = converted SOMARIS Standard Mode image
  - CT\_SOM4 SEQ = converted SOMARIS Sequence Mode image
  - CT\_SOM4 SER = converted SOMARIS Serial Mode image
  - CT\_SOM4 SIN = converted SOMARIS Sinogram image
  - CT\_SOM4 SINC = converted SOMARIS Sinogram image
  - CT\_SOM4 SPI = converted SOMARIS Spiral Mode image
  - CT\_SOM4 STA = converted SOMARIS Static Mode image
  - CT\_SOM4 TAB = converted SOMARIS Correction Table image
  - CT\_SOM4 TOP = converted SOMARIS Topogram image
  - CT\_SOM4 GTOP = converted SOMARIS Topo Graphics image
  - CT\_SOM4 PEVG = converted SOMARIS Pulmo Evaluation image
  - CT\_SOM4 PEVI = converted SOMARIS Pulmo Evaluation image
  - CT\_SOM4 PUL = converted SOMARIS Pulmo Respiration image
  - CT\_SOM4 PROT = converted SOMARIS Protocol image
  - CT\_SOM4 TEXT = converted SOMARIS Text image
  - CT\_SOM4 ICD = converted SOMARIS Interventional Cine image
  - SHS DENT = converted MagicView Dental Tomogram image
  - SHS DPAN = converted MagicView Dental Panorama image
  - SHS DPAR = converted MagicView Dental Paraxial image
  - SHS 3D\_CURVED = converted MagicView image
  - SHS 3D\_MIP = converted MagicView Maximum Intensity Projection image
  - SHS 3D\_MPR = converted MagicView Multi Planar Reconstruction image
  - SHS 3D\_SSD = converted MagicView Shaded Surface Display image
  - SHS 3D\_VRT = converted MagicView Volumetric Rendering image

**A.5 Structured Report Templates**

**A.5.1 ESWL SR Templates**

**A.5.2 TID Tnx001 ESWL Report**

Root Template of the ESWL Report.  
This template contains the top level structure and includes sub-templates for the various procedures. This template together with its subordinate templates, describe both the findings and summary of therapy.

Type: Extensible

**TID Tnx001  
ESWL Report**

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CONTAINER	EV (C-10, 99SMS_ESWL, "Litho-Report")	1	M		Root node
2	>	HAS CONCEPT MOD	INCLUDE	DTID (1204) Language of Content Item and Descendants	1	U		Language
3	>	CONTAINS	TEXT	EV(C-202, 99SMS_ESWL, "Amublant / ward")	1	M		
4	>	CONTAINS	PNAME	EV(C-202, 99SMS_ESWL, "Performing Physician")	1	M		
5	>	CONTAINS	PNAME	EV(C-204, 99SMS_ESWL, "Assistance")	1	U		
6	>	CONTAINS	PNAME	EV(C-206, 99SMS_ESWL, "Anesthetist")	1	U		
7	>	CONTAINS	INCLUDE	DTID (ESWL Findings)	1	U		
8	>	CONTAINS	INCLUDE	DTID (ESWL Therapy Data)	1	U		
9	>	CONTAINS	INCLUDE	DTID (ESWL Treatment Summary)	1	U		
10	>	CONTAINS	INCLUDE	DTID (ESWL Follow-up Examination)	1	U		

**A.5.2.1 TID Tnx002 ESWL Findings**

Contains the diagnosis data.

Type: Extensible

**TID Tnx002  
ESWL Findings**

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CONTAINER	EV(C-20,99SMS_ESWL, "Findings")	1	U		

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
2	>	CONTAINS	CODE	EV(C-300, 99SMS_ESWL, "Diagnosis with")	1-n	M		DCID(Tnx-200)
3	>	CONTAINS	CODE	EV(C-306, 99SMS_ESWL, "Localisation with")	1-n	M		DCID(Tnx-200)
4	>	CONTAINS	CODE	EV(C-310, 99SMS_ESWL, "Urinary Obstruction")	1	M		DCID(Tnx-201)
5	>	CONTAINS	TEXT	EV(C-314, 99SMS_ESWL, "First Treatment / Retreatment")	1	M		User defined list
6	>	CONTAINS	TEXT	EV(C-316, 99SMS_ESWL, "Primary Stone Location")	1	M		15 user defined values. Default : Upper Calyx right Middle Calyx right Renal Pelvis right Lower Calyx right Upper Ureter right Middle Ureter right Lower Ureter right Upper Calyx left Renal Pelvis left Middle Calyx left Lower Calyx left Upper Ureter left Middle Ureter left Lower Ureter left Bladder Stone
7	>	CONTAINS	TEXT	EV(C-406, 99SMS_ESWL, "Primary Stone Location (id)")	1	M		Internal Id = corresponds to row 6
8	>	CONTAINS	TEXT	EV(C-318, 99SMS_ESWL, "Primary Stone Size")	1	M		User defined list
9	>	CONTAINS	TEXT	EV(C-320, 99SMS_ESWL, "Multiple Stones")	1	M		Yes, no

**A.5.2.2 TID Tnx003 ESWL Therapy Data**

Contains the therapy characteristics.

Type: Extensible

**TID Tnx003  
ESWL Therapy Data**

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CONTAINER	EV(C-30,99SMS_ESWL, "Therapy Data")	1	U		
2	>	CONTAINS	TEXT	EV(C-400, 99SMS_ESWL, "Auxiliary Procedures before ESWL")	1-n	U		
3	>	CONTAINS	TEXT	EV(C-402, 99SMS_ESWL, "Other Auxiliary Procedures before ESWL")	1	U		
5	>	CONTAINS	TEXT	EV(C-408, 99SMS_ESWL, "Patient Position")	1	U		User defined list.
6	>	CONTAINS	NUM	EV(C-410, 99SMS_ESWL, "Number of Shockwaves")	1	U		UnitsEV(1,UCUM,"Unity") Value= Integer >= 1



	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
7	>	CONTAINS	NUM	EV(C-412, 99SMS_ESWL, "Cumulative Energy")	1	U		UnitsEV(J,UCUM,"joule")
8	>	CONTAINS	NUM	EV(C-414, 99SMS_ESWL, "Medium Energy ( Level)")	1	U		UnitsEV(mJ,UCUM,"millijoule")
9	>	CONTAINS	NUM	EV(C-416, 99SMS_ESWL, "Maximum Energy ( Level)")	1	U		UnitsEV(mJ,UCUM,"millijoule")
10	>	CONTAINS	NUM	EV(C-418, 99SMS_ESWL, "Shock Wave Frequency")	1	U		UnitsEV(1,UCUM,"Unity") Value= Integer >= 1
11	>	CONTAINS	TEXT	EV(C-420, 99SMS_ESWL, "ECG Trigger")	1	U		0,1
12	>	CONTAINS	TIME	EV(C-422, 99SMS_ESWL, "Start of the Therapy")	1	U		
13	>	CONTAINS	TIME	EV(C-424, 99SMS_ESWL, "End of the Therapy")	1	U		
14	>	CONTAINS	NUM	EV(C-426, 99SMS_ESWL, "Therapy Duration")	1	U		Unit = DT(min,UCUM,minute)

**A.5.2.3 TID Tnx004 ESWL Treatment Summary**

Contains summaries related to the performed treatments.

Type: Extensible

**TID Tnx004  
ESWL Treatment Summary**

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CONTAINER	EV(C-40, 99SMS_ESWL, "Treatment Summary")	1	M		
2	>	CONTAINS	TIME	EV(C-500, 99SMS_ESWL, "Start of the Treatment")	1	U		
3	>	CONTAINS	TIME	EV(C-502, 99SMS_ESWL, "End of the Treatment")	1	U		
4	>	CONTAINS	NUM	EV(C-504, 99SMS_ESWL, "Treatment Duration")	1	U		Unit = DT ("min", UCUM, "minute")
5	>	CONTAINS	TEXT	EV(C-506, 99SMS_ESWL, "Anesthesia")	1	U		The values are user defined at this moment.

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
6	>	CONTAINS	NUM	EV(C-508, 99SMS_ESWL, "Anesthesia Dose/Amount")	1	U		UnitsEV(1,UCUM,"Unity")
7	>	CONTAINS	NUM	EV(C-510, 99SMS_ESWL, "Pain")	1	U		UnitsEV(1,UCUM,"Unity") 1-10
8	>	CONTAINS	NUM	EV(C-512, 99SMS_ESWL, "X-Ray Radiation Duration")	1	U		Unit = DT(min,UCUM,minute)
9	>	CONTAINS	NUM	EV(C-514, 99SMS_ESWL, "X-Ray Area Dose Product")	1	M		UnitsEV(Gy.cm^2,UCUM,"Gray-squarecentimeter")
10	>	CONTAINS	NUM	EV(C-526, 99SMS_ESWL, "X-Ray air kerma")	1	C	If configured	UnitsEV(Gy,UCUM,"Gray")
11	>	CONATINS	TEXT	EV(C-516, 99SMS_ESWL, "Localization")	1	U		User defined list.
12	>	CONTAINS	TEXT	EV(C-518, 99SMS_ESWL, "Desintegration")	1	U		User defined list.
13	>	CONTAINS	TEXT	EV(C-520, 99SMS_ESWL, "Complications")	1-n	U		User defined list.
14	>	CONTAINS	TEXT	EV(C-522, 99SMS_ESWL, "Other Complications")	1	U		
15	>	CONTAINS	TEXT	EV(C-524, 99SMS_ESWL, "Comment")	1	U		

**A.5.2.4 TID Tnx005 ESWL Follow-up Examination**

Contains information about further examinations.

Type: Extensible

**TID Tnx005  
ESWL Treatment Summary**

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CONTAINER	EV(C-50, 99SMS_ESWL, "Follow-up Examination")	1	M		
2	>	CONTAINS	CODE	EV(C-600, 99SMS_ESWL, "Stonefragmentation")	1	U		DCID(Tnx-202)
3	>	CONTAINS	CODE	EV(C-604, 99SMS_ESWL, "Control of Result with")	1-n	U		DCID(Tnx-200)
4	>	CONTAINS	TEXT	EV(C-608, 99SMS_ESWL, "Auxiliary Procedures")	1-n	U		User defined list.
5	>	CONTAINS	TEXT	EV(C-610, 99SMS_ESWL, "Other Auxiliary Procedure")	1	U		
6	>	CONTAINS	TEXT	EV(C-612, 99SMS_ESWL, "Further actions")	1-n	U		User defined list.

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
7	>	CONTAINS	TEXT	EV(C-614, 99SMS_ESWL, "Other Further actions")	1	U		
8	>	CONTAINS	TEXT	EV(C-616, 99SMS_ESWL, "Comment to the follow up examination")	1	U		

**A.5.2.5 Context Groups**

Type: Extensible, Version 1.0 and Coding Scheme Designator = 99SMS\_ESWL

**CID Tnx200  
ESWL Diagnosis with**

Code Value (0008,0100)	Code Meaning (0008,0104)
C-10000	X-Ray
C-10002	Ultrasound
C-10004	KUB
C-10006	IVP

**CID Tnx201  
ESWL Urinary obstruction**

Code Value (0008,0100)	Code Meaning (0008,0104)
C-10100	None
C-10102	Left
C-10104	Right

**TID Tnx202  
ESWL Stonefragmentation**

Code Value (0008,0100)	Code Meaning (0008,0104)
C-10200	Stonefree
C-10202	< 3 mm
C-10204	3 – 5 mm
C-10206	> 5 mm

**A.5.3 Dose Report**

**A.5.4 TID Tnx002 Dose Report**

Root Template of the Dose Report.

**TID Tnx010**

**Radiation Summary Report**

	NL	Rel with Parent	VT	Concept Name	VM	Req Type	Condition	Value Set Constraint
1			CONTAINER	EV(C-10, 99SMS_RADSUM, Radiation Summary Report")	1	M		Root node
2	>	HAS CONCEPT MOD	CODE	DTID (1204) Language of Content Item and Descendants	1	U		Language
4	>	CONTAINS	NUM	EV(C-202, 99SMS_RADSUM, "Cumulative fluoro time")	1	M		Unit = DT ("sec", UCUM, "second")
5	>	CONTAINS	NUM	EV(C-204, 99SMS_RADSUM, "Cumulative dose area product")	1	U		UnitsEV(Gy.cm^2,UCUM,"Gray-squarecentimeter")
6	>	CONTAINS	DATETIME	EV(C-206, 99SMS_RADSUM, "Start of the examination")	1	U		
7	>	CONTAINS	DATETIME	EV(C-208, 99SMS_RADSUM, "End of the examination")	1	U		
8	>	CONTAINS	NUM	EV(C-210, 99SMS_RADSUM, "Cumulative air kerma")	1	U	If configured	UnitsEV(Gy,UCUM,"Gray")

**A.6 DICOM Print SCU – detailed status displays**

The following tables document the behavior of the ARCADIS DICOM Print AE in response to messages received for the printer SOP class and the print job SOP class.

Definitions of camera symbols:

- Idle: Camera is installed and ready; idle icon is displayed.
- Interact: The user has to react in near future, but not immediately.  
Example: A camera was low in 8x10 clear sheets: LOW 8x10 CLR was sent by n-event-report.
- Queue Stopped: The user has to react immediately. Either the camera needs immediate interaction or a job has been aborted.  
Example: A camera is out of 8x10 clear sheets, or camera is down, or a film job is aborted.

Note: different camera symbols are displayed according to the Printer Status Info.

The Printer Status (Success, Warning, and Failure) is not evaluated, since the Printer Status Info is much more detailed and allows a more appropriate reaction of the system.

## A.6.1 Common Status Information

*"Common Status Info evaluation"*

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
NORMAL	Camera is ready	Camera is ready	<None>/idle
BAD RECEIVE MGZ	There is a problem with the film receive magazine. Films from the printer cannot be transported into the magazine.	Problem with receive magazine.	<None>/interact
BAD SUPPLY MGZ	There is a problem with the film supply magazine. Films from this magazine cannot be transported into the printer.	Problem with supply magazine.	<None>/interact
CALIBRATING	Printer is performing self calibration; it is expected to be available for normal operation shortly.	Self calibration. Please wait.	<None>/idle
CALIBRATION ERR	An error in the printer calibration has been detected; quality of processed films may not be optimal.	Problem in calibration. Film quality may not be optimal.	<None>/interact
CHECK CHEMISTRY	A problem with the processor chemicals has been detected; quality of processed films may not be optimal.	Problem with chemistry. Film quality may not be optimal.	<None>/interact
CHECK SORTER	There is an error in the film sorter	Error in film sorter.	<None>/interact
CHEMICALS EMPTY	There are no processing chemicals in the processor, films will not be printed and processed until the processor is back to normal.	Camera chemistry empty. Please check.	<None>/interact
CHEMICALS LOW	The chemical level in the processor is low, if not corrected, it will probably shut down soon.	Camera chemistry low. Please check.	<None>/interact
COVER OPEN	One or more printer or processor covers, drawers, doors are open.	Camera cover, drawer or door open.	<None>/interact
ELEC CONFIG ERR	Printer configured improperly for this job.	Camera configured improperly for this job. Queue stopped.	<b>Queue for this camera will be STOPPED/ Queue stopped</b>
ELEC DOWN	Printer is not operating due to some unspecified electrical hardware problem.	Camera electrical hardware Problem.	<None>/interact
ELEC SW ERROR	Printer not operating for some unspecified software error.	Camera software problem. Queue stopped.	<b>Queue for this camera will be STOPPED/ Queue stopped</b>
EMPTY 8x10	The 8x10 inch film supply magazine is empty.	8x10 film supply empty.	<None>/interact
EMPTY 8x10 BLUE	The 8x10 inch blue film supply magazine is empty.	8x10 blue film supply empty.	<None>/interact
EMPTY 8x10 CLR	The 8x10 inch clear film supply magazine is empty.	8x10 clear film supply empty.	<None>/interact
EMPTY 8x10 PAPR	The 8x10 inch paper supply magazine is empty.	8x10 paper supply empty.	<None>/interact
EMPTY 10x12	The 10x12 inch film supply magazine is empty.	10x12 film supply empty.	<None>/interact
EMPTY 10x12 BLUE	The 10x12 inch blue film supply magazine is empty.	10x12 blue film supply empty.	<None>/interact
EMPTY 10x12 CLR	The 10x12 inch clear film supply magazine is empty.	10x12 clear film supply empty.	<None>/interact
EMPTY 10x12 PAPR	The 10x12 inch paper supply magazine is empty.	10x12 paper supply empty.	<None>/interact
EMPTY 10x14	The 10x14 inch film supply magazine is empty.	10x14 film supply empty.	<None>/interact
EMPTY 10x14 BLUE	The 10x14 inch blue film supply magazine is empty.	10x14 blue film supply empty.	<None>/interact
EMPTY 10x14 CLR	The 10x14 inch clear film supply magazine is empty.	10x14 clear film supply empty.	<None>/interact
EMPTY 10x14 PAPR	The 10x14 inch paper supply magazine is empty.	10x14 paper supply empty.	<None>/interact
EMPTY 11x14	The 11x14 inch film supply magazine is empty.	11x14 film supply empty.	<None>/interact
EMPTY 11x14 BLUE	The 11x14 inch blue film supply magazine is empty.	11x14 blue film supply empty.	<None>/interact

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
EMPTY 11x14 CLR	The 11x14 inch clear film supply magazine is empty.	11x14 clear film supply empty.	<None>/interact
EMPTY 11x14 PAPER	The 11x14 inch paper supply magazine is empty.	11x14 paper supply empty.	<None>/interact
EMPTY 14x14	The 14x14 inch film supply magazine is empty.	14x14 film supply empty.	<None>/interact
EMPTY 14x14 BLUE	The 14x14 inch blue film supply magazine is empty.	14x14 blue film supply empty.	<None>/interact
EMPTY 14x14 CLR	The 14x14 inch clear film supply magazine is empty.	14x14 clear film supply empty.	<None>/interact
EMPTY 14x14 PAPER	The 14x14 inch paper supply magazine is empty.	14x14 paper supply empty.	<None>/interact
EMPTY 14x17	The 14x17 inch film supply magazine is empty.	14x17 film supply empty.	<None>/interact
EMPTY 14x17 BLUE	The 14x17 inch blue film supply magazine is empty.	14x17 blue film supply empty.	<None>/interact
EMPTY 14x17 CLR	The 14x17 inch clear film supply magazine is empty.	14x17 clear film supply empty.	<None>/interact
EMPTY 14x17 PAPER	The 14x17 inch paper supply magazine is empty.	14x17 paper supply empty.	<None>/interact
EMPTY 24x24	The 24x24 inch film supply magazine is empty.	24x24 film supply empty.	<None>/interact
EMPTY 24x24 BLUE	The 24x24 inch blue film supply magazine is empty.	24x24 blue film supply empty.	<None>/interact
EMPTY 24x24 CLR	The 24x24 inch clear film supply magazine is empty.	24x24 clear film supply empty.	<None>/interact
EMPTY 24x24 PAPER	The 24x24 inch paper supply magazine is empty.	24x24 paper supply empty.	<None>/interact
EMPTY 24x30	The 24x30 inch film supply magazine is empty.	24x30 film supply empty.	<None>/interact
EMPTY 24x30 BLUE	The 24x30 inch blue film supply magazine is empty.	24x30 blue film supply empty.	<None>/interact
EMPTY 24x30 CLR	The 24x30 inch clear film supply magazine is empty.	24x30 clear film supply empty.	<None>/interact
EMPTY 24x30 PAPER	The 24x30 inch paper supply magazine is empty.	24x30 paper supply empty.	<None>/interact
EMPTY A4 PAPER	The A4 paper supply magazine is empty.	A4 paper supply empty.	<None>/interact
EMPTY A4 TRANS	The A4 transparency supply magazine is empty.	A4 transparency supply empty.	<None>/interact
EXPOSURE FAILURE	The exposure device has failed due to some unspecified reason.	Exposure device has failed.	<None>/interact
FILM JAM	A film transport error has occurred and a film is jammed in the printer or processor.	Film jam.	<None>/interact
FILM TRANSP ERR	There is a malfunction with the film transport, there may or may not be a film jam.	Film transport problem.	<None>/interact
FINISHER EMPTY	The finisher is empty.	Finisher is empty.	<None>/interact
FINISHER ERROR	The finisher is not operating due to some unspecified reason.	Finisher problem.	<None>/interact
FINISHER LOW	The finisher is low on supplies.	Finisher low.	<None>/interact
LOW 8x10	The 8x10 inch film supply magazine is low.	8x10 film supply low.	<None>/interact
LOW 8x10 BLUE	The 8x10 inch blue film supply magazine is low.	8x10 blue film supply low.	<None>/interact
LOW 8x10 CLR	The 8x10 inch clear film supply magazine is low.	8x10 clear film supply low.	<None>/interact
LOW 8x10 PAPER	The 8x10 inch paper supply magazine is low.	8x10 paper supply low.	<None>/interact
LOW 10x12	The 10x12 inch film supply magazine is low.	10x12 film supply low.	<None>/interact
LOW 10x12 BLUE	The 10x12 inch blue film supply magazine is low.	10x12 blue film supply low.	<None>/interact
LOW 10x12 CLR	The 10x12 inch clear film supply magazine is low.	10x12 clear film supply low.	<None>/interact
LOW 10x12 PAPER	The 10x12 inch paper supply magazine is low.	10x12 paper supply low.	<None>/interact
LOW 10x14	The 10x14 inch film supply magazine is low.	10x14 film supply low.	<None>/interact

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
LOW 10x14 BLUE	The 10x14 inch blue film supply magazine is low.	10x14 blue film supply low.	<None>/interact
LOW 10x14 CLR	The 10x14 inch clear film supply magazine is low.	10x14 clear film supply low.	<None>/interact
LOW 10x14 PAPR	The 10x14 inch paper supply magazine is low.	10x14 paper supply low.	<None>/interact
LOW 11x14	The 11x14 inch film supply magazine is low.	11x14 film supply low.	<None>/interact
LOW 11x14 BLUE	The 11x14 inch blue film supply magazine is low.	11x14 blue film supply low.	<None>/interact
LOW 11x14 CLR	The 11x14 inch clear film supply magazine is low.	11x14 clear film supply low.	<None>/interact
LOW 11x14 PAPR	The 11x14 inch paper supply magazine is low.	11x14 paper supply low.	<None>/interact
LOW 14x14	The 14x14 inch film supply magazine is low.	14x14 film supply low.	<None>/interact
LOW 14x14 BLUE	The 14x14 inch blue film supply magazine is low.	14x14 blue film supply low.	<None>/interact
LOW 14x14 CLR	The 14x14 inch clear film supply magazine is low.	14x14 clear film supply low.	<None>/interact
LOW 14x14 PAPR	The 14x14 inch paper supply magazine is low.	14x14 paper supply low.	<None>/interact
LOW 14x17	The 14x17 inch film supply magazine is low.	14x17 film supply low.	<None>/interact
LOW 14x17 BLUE	The 14x17 inch blue film supply magazine is low.	14x17 blue film supply low.	<None>/interact
LOW 14x17 CLR	The 14x17 inch clear film supply magazine is low.	14x17 clear film supply low.	<None>/interact
LOW 14x17 PAPR	The 14x17 inch paper supply magazine is low.	14x17 paper supply low.	<None>/interact
LOW 24x24	The 24x24 inch film supply magazine is low.	24x24 film supply low.	<None>/interact
LOW 24x24 BLUE	The 24x24 inch blue film supply magazine is low.	24x24 blue film supply low.	<None>/interact
LOW 24x24 CLR	The 24x24 inch clear film supply magazine is low.	24x24 clear film supply low.	<None>/interact
LOW 24x24 PAPR	The 24x24 inch paper supply magazine is low.	24x24 paper supply low.	<None>/interact
LOW 24x30	The 24x30 inch film supply magazine is low.	24x30 film supply low.	<None>/interact
LOW 24x30 BLUE	The 24x30 inch blue film supply magazine is low.	24x30 blue film supply low.	<None>/interact
LOW 24x30 CLR	The 24x30 inch clear film supply magazine is low.	24x30 clear film supply low.	<None>/interact
LOW 24x30 PAPR	The 24x30 inch paper supply magazine is low.	24x30 paper supply low.	<None>/interact
LOW A4 PAPR	The A4 paper supply magazine is low.	A4 paper supply low.	<None>/interact
LOW A4 TRANS	The A4 transparency supply magazine is low...	A4 transparency supply low.	<None>/interact
NO RECEIVE MGZ	The film receive magazine is not available.	Film receiver not available.	<None>/interact
NO RIBBON	The ribbon cartridge needs to be replaced.	Replace ribbon cartridge.	<None>/interact
NO SUPPLY MGZ	The film supply magazine is not available.	Film supply not available.	<None>/interact
CHECK PRINTER	The printer is not ready at this time; operator intervention is required to make the printer available.	Check camera.	<None>/interact
CHECK PROC	The processor is not ready at this time; operator intervention is required to make the printer available.	Check processor.	<None>/interact
PRINTER DOWN	The printer is not operating due to some unspecified reason.	Camera down.	<None>/interact
PRINTER INIT	The printer is not ready at this time; it is expected to become available without intervention. For example, it may be in a normal warm-up state.	Camera initializing.	<None>/Idle
PRINTER OFFLINE	An operator or service person has disabled the printer.	Camera off-line.	<None>/interact
PROC DOWN	The processor is not operating due to	Processor down.	<None>/interact

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
	some unspecified reason.		
PROC INIT	The processor is not ready at this time; it is expected to become available without intervention. For example, it may be in a normal warm-up state.	Processor initializing.	<None>/Idle
PROC OVERFLOW FL	Processor chemicals are approaching the overflow full mark.	Processor chemicals near overflow.	<None>/interact
PROC OVERFLOW HI	Processor chemicals have reached the overflow full mark.	Processor chemicals overflow.	<None>/interact
QUEUED	Print job in Queue	--	<None>/Idle
RECEIVER FULL	The film receive magazine is full.	Receiver full.	<None>/interact
REQ MED NOT INST	The requested film, paper, or other media supply magazine is installed in the printer, but may be available with operator intervention.	Install media supply.	<None>/interact
REQ MED NOT AVAI	The requested film, paper, or other media requested is not available on this printer.	Media supply not available on this camera. Queue stopped. Change camera.	<b>Queue for this camera will be STOPPED/ Queue stopped</b>
RIBBON ERROR	There is an unspecified problem with the print ribbon.	Error with print ribbon.	<None>/interact
SUPPLY EMPTY	The printer is out of film.	Camera out of film.	<None>/interact
SUPPLY LOW	The film supply is low.	Film supply low.	<None>/interact
UNKNOWN	There is an unspecified problem.	Unspecified problem with camera.	<None>/interact



**A.6.2 Additional Status Information – AGFA printers***“Additional Agfa printer Status Info evaluation”*

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
WARMING UP	Printer is in the warm-up stage. Spooling of print jobs to disk is still possible.	Camera is warming up.	<None>/idle
OFFLINE	OFFLINE Printer is switched off-line. Spooling of print jobs to disk is still possible.	Camera is switched off-line.	<None>/interact
NONE	General printer warning, no specific information is available. Spooling of print jobs to disk is still possible.	--	<None>/idle

**A.6.3 Additional Status Information – Kodak PACS Link (formerly Imation)***“Additional Kodak PACS Link (Imation) printer Status Info evaluation”*

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
SUPPLY MGZ ERR	The supply magazine has an error.	Film supply has an error.	<None>/interact

**A.6.4 Additional Status Information – Kodak 1901***“Additional Kodak 190 printer Status Info evaluation”*

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
PRINTER STOPPED	The printer has stopped.	Camera has stopped.	<None>/interact
FATAL ERROR	Fatal Error.	Fatal Error. Queue stopped.	<b>Queue for this camera will be STOPPED/ Queue stopped</b>

**A.6.5 Additional Status Information – Kodak 2180/1120***“Additional Kodak 2180/1120 printer Status Info evaluation”*

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
PRINTER NOT RDY	Printer not ready.	Camera not ready...	<None>/interact
CHECK PROCESSOR	Check processor.	Check processor.	<None>/interact
NO TONER	No toner.	No toner.	<None>/interact
FATAL	Fatal Error.	Fatal Error. Queue stopped.	<b>Queue for this camera will be STOPPED/ Queue stopped</b>

## A.6.6 Additional Status Information – Codonics

*“Additional Codonics printer Status Info evaluation”*

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
STANDARD	Printer is ready.	Camera is ready.	<None>/Normal
LOAD A-SIZE	Load A-Size media.	Load A-Size media.	<None>/interact
LOAD A-DVPAPER	Load A-Size black and white paper.	Load A-Size black and white paper.	<None>/interact
LOAD A-CVPAPER	Load A-Size color paper.	Load A-Size color paper.	<None>/interact
LOAD A-CVTRANS	Load A-Size transparencies.	Load A-Size transparencies.	<None>/interact
LOAD A4-SIZE	Load A4-Size media.	Load A4-Size media.	<None>/interact
LOAD A4-DVPAPER	Load A4-Size black and white paper.	Load A4-Size black and white paper.	<None>/interact
LOAD A4-CVPAPER	Load A4-Size color paper.	Load A4-Size color paper.	<None>/interact
LOAD A4-CVTRANS	Load A4-Size transparencies.	Load A4-Size transparencies.	<None>/interact
LOAD LA-SIZE	Load LA-Size media.	Load LA-Size media.	<None>/interact
LOAD LA-DVPAPER	Load LA-Size black and white paper.	Load LA-Size black and white paper.	<None>/interact
LOAD LA-CVPAPER	Load LA-Size color paper.	Load LA-Size color paper.	<None>/interact
LOAD LA-CVTRANS	Load LA-Size transparencies.	Load LA-Size transparencies.	<None>/interact
LOAD LA4-SIZE	Load LA4-Size media.	Load LA4-Size media.	<None>/interact
LOAD LA4-DVPAPER	Load LA4-Size black and white paper.	Load LA4-Size black and white paper.	<None>/interact
LOAD LA4-CVPAPER	Load LA4-Size color paper.	Load LA4-Size color paper.	<None>/interact
LOAD LA4-CVTRANS	Load LA4-Size transparencies.	Load LA4-Size transparencies.	<None>/interact
LOAD XLA-SIZE	Load XLA-Size media.	Load XLA-Size media.	<None>/interact
LOAD XLA-DVPAPER	Load XLA-Size black and white paper.	Load XLA-Size black and white paper.	<None>/interact
LOAD XLA-CVPAPER	Load XLA-Size color paper.	Load XLA-Size color paper.	<None>/interact
LOAD XLA-CVTRANS	Load XLA-Size transparencies.	Load XLA-Size transparencies.	<None>/interact
LOAD XLA4-SIZE	Load XLA4-Size media.	Load XLA4-Size media.	<None>/interact
LOAD XLA4-DVPAPER	Load XLA4-Size black and white paper.	Load XLA4-Size black and white paper.	<None>/interact
LOAD XLA4-CVPAPER	Load XLA4-Size color paper.	Load XLA4-Size color paper.	<None>/interact
LOAD XLA4-CVTRANS	Load XLA4-Size transparencies.	Load XLA4-Size transparencies.	<None>/interact
LOAD XLW-SIZE	Load XLW-Size media.	Load XLW-Size media.	<None>/interact
LOAD XLW-DVPAPER	Load XLW-Size black and white paper.	Load XLW-Size black and white paper.	<None>/interact
LOAD XLW-CVPAPER	Load XLW-Size color paper.	Load XLW-Size color paper.	<None>/interact
LOAD 8X10-SIZE	Load 8x10 media.	Load 8x10 media.	<None>/interact
LOAD 8X10-DVFILM	Load XLW-Size black and white film.	Load XLW-Size black and white film.	<None>/interact
SUPPLY MISSING	The film supply magazine specified for this job is not available.	Film supply not available.	<None>/interact
RIBBON MISSING	Ribbon is missing.	Ribbon is missing.	<None>/interact
RIBBON EMPTY	Ribbon is empty.	Ribbon is empty.	<None>/interact
TOP COVER OPEN	Top cover of printer is open.	Top cover of camera is open.	<None>/interact

### A.6.7 Additional DICOM Execution Status Information

*“Additional DICOM Execution Status Info evaluation”*

Printer Status Info/ Execution Status Info	Description	Message string visible in 'Status Bar'	Other action for UI/ 'camera symbol'
INVALID PAGE DES	The specified page layout cannot be printed or other page description errors have been detected.	Film Job cannot be printed on this camera. Queue stopped. Please redirect film job.	<b>Queue for this camera will be STOPPED/ Queue stopped</b>
INSUFFICIENT MEMORY	There is not enough memory available to complete this job.	Not enough memory available in camera. Queue stopped. Please continue queue or change camera.	<b>Queue for this camera will be STOPPED/ Queue stopped</b>
NONE	General printer warning, no specific information is available. Spooling of print jobs to disk is still possible.	--	<None>/Idle